

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Buena Borough Police Department 616 Central Ave. Minotola, NJ 08341	2. Contact: Sgt. Petrilla 3. Phone Number: 856-697-2810	4. Date: 8/31/10 5. County: Atlantic
--	--	---

6. Alcotest Instrument Serial Number: ARXB-0074	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A
---	---	---

9. Reason for Service:
Alcotest instrument returned from outside maintenance and placed back in service.


10. Comments:
**Fuel cell replaced. Cuvette replaced. Barb, brass, metric w/ valve replaced.
 See Draeger packing slip.**

Note: The Alcotest data from 03/30/10 until 08/13/10 was not retrieved by Draeger.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steve Dellanoc #6027
Name & Badge Number (Print)


Signature

8/31/10
Date

Packing Slip

Customer no Order No Order date
 150065405 10425715 08/19/2010

Packing slip no. Ship date
 80540965 08/19/2010

Please reference on inquiries

ship to
 150065406
 BOROUGH OF BUENA POLICE DEPARTMENT
 ATTN: CHIEF ADAMS
 616 CENTRAL AVENUE
 ATLANTIC COUNTY
 MINOTOLA, NJ 08341
 USA

bill to

Your Purchase Order Number

7110 WARRANTY; ARXB-0074
 Any questions? Please contact:
 MARA HERRERA

Customer
 BOROUGH OF BUENA
 ATTN: FINANCE DEPARTMENT
 P.O. BOX 620
 ATLANTIC COUNTY
 MINOTOLA, NJ 08341-0620
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARXB-0074			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED			
0020	1 EA	6809500 CUVETTE, ALCOTEST PV LARGE PREWARM TRAC Ordered / Back ordered : 0 / 0 EA REPLACED			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5588

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150065405 10425715 08/19/2010

Packing slip no. Ship date
80540965 08/19/2010

Please reference on Inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6809512 BARB, BRASS, METRIC W/ VALVE Ordered / Back ordered : 0 / 0 EA REPLACED			
0040	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			
0050	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Buena Borough Police Department 616 Central Ave. Minotola, NJ 08341		2. Contact: Sgt. Petrilla		4. Date: 08/13/2010	
		3. Phone Number: 856-697-2810		5. County: Atlantic	
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXB-0074</p>		7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>		8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Complaint received from Buena Borough Police of ERROR [023] IR-SYSTEM on display screen of Alcotest.

10. Comments:
Operational condition of the instrument was assessed. Upon arrival at the police department the instrument displayed ERROR [023] IR-SYSTEM. I powered the instrument off and back on and heard an audible series of beeps then ERROR [071] HEATER SYSTEM appeared on the display screen. I was not able to clear this message from display screen.

I then observed the simulator and temperature probe. The simulator jar was extremely hot to the touch, the heating element glowed bright red, and approximately one-quarter of the jar still contained solution which was murky due to the melted plastic baffle. I was unable to pull any reports from the memory but was provided with a solution change report from 08/09/10 which revealed a failure due to a simulator temp error of 72.4 degrees C.

NOTE: I was unable to download the data due to the display screen being locked on ERROR [071] HEATER SYSTEM.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steve Dellanoce #6027 <small>Name & Badge Number (Print)</small>	 <small>Signature</small>	# 6027 8/13/10 <small>Date</small>
--	------------------------------	---

1. Department: Bucca Boro Police Dept		2. Contact: Sgt Petrillo	3. Phone Number: 609-697-2810	4. Date: 9/24/10
5. County: Atlantic		6. Alcotest Instrument Serial Number: N/A		
7. Simulator Component Serial Number: D085T-0006		8. Temperature Probe Component Serial Number: D0WTFa-058		

9. Reason for Service:
 Simulator and temperature probe components returned from outside maintenance and placed back in service.

10. Comments:
 Components rechecked. See Drawer packing slip.

11. The above instrument/component has been found to be in satisfactory working condition; no further action required.

12. The above instrument/component is placed out of service pending further evaluation.

13. The above instrument/component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the foregoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the foregoing statements made by me are willfully false, I am subject to punishment."

Name & Badge Number (Print): Tr II Steve DeLance # 6027

Signature: Tr II Steve DeLance # 6027

Date: 9/24/10


Packing Slip

Customer no 150065405 Order No 10429212
 Order date 09/08/2010

Packing slip no 80544982
 Ship date 09/08/2010

Please reference on inquiries

ship to
 150065406
 BOROUGH OF BUENA POLICE DEPARTMENT
 ATTN: CHIEF ADAMS
 616 CENTRAL AVENUE
 ATLANTIC COUNTY
 MINOTOLA, NJ 08341
 USA



Customer
 BOROUGH OF BUENA
 ATTN: FINANCE DEPARTMENT
 P.O. BOX 620
 ATLANTIC COUNTY
 MINOTOLA, NJ 08341-0620
 USA

Your Purchase Order Number
 SIM/PROBE WARRANTY
 Any questions? Please contact:
 MARA HERRERA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	Sh. Quant.	Part no. Description	Ship via: Fed Ex Saver
0010	1 EA	MPCERT CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 0 / 0 EA	
0020	1 EA	MPCAL/TP CAL CHARGE, SIM PROBE Ordered / Back ordered : 0 / 0 EA	

NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box.
 Please call Customer Service at 866-385-5900.
 SIM: DDBJ-0006 PROBE: DDMJP2-088

Dräger Safety Diagnostics Inc.
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct # 6209426615
 Acct Name: Dräger Safety Diagnostics Inc.
 Transil Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Dräger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150065405 10429212 09/08/2010

Packing slip no. Ship date
80544982 09/08/2010
Please reference on inquiries

Page 2 / 2

Item#	sh. Qunt.	Part no. Description			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 0 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: <i>Buena Borough Police Dept 616 Central Ave Minotola, NJ 08341</i>	2. Contact: <i>Sgt Petrilla</i>	4. Date: <i>8/31/10</i>
	3. Phone Number: <i>609-697-2810</i>	5. County: <i>Atlantic</i>


6. Alcotest Instrument Serial Number: <i>N/A</i>	7. Simulator Component Serial Number: <i>DOB5-0006</i>	8. Temperature Probe Component Serial Number: <i>DDWJP2-088</i>
---	---	--

9. Reason for Service:
Simulator temp at 34.34 with digital temp measuring system, too high outside of tolerance.

10. Comments:
Please recert simulator and temp probe together to keep expiration dates consistent. Thank you.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<i>Tpr II Steve Dellance # 6027</i> Name & Badge Number (Print)	 Signature	<i>8/31/10</i> Date
--	---	------------------------

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Buena Borough Police Department 616 Central Ave. Minotola, NJ 08341	2. Contact: Sgt. Petrilla	4. Date: 8/31/10
3. Phone Number: 856-697-2810		5. County: Atlantic

6. Alcotest Instrument Serial Number: <p style="text-align: center;">N/A</p>	7. Simulator Component Serial Number: DDXJ S3-0209 replaced by DDBJ-0006	8. Temperature Probe Component Serial Number: DDWAP2-148 replaced by DDWJP2-088
--	---	--

9. Reason for Service:
Above listed components placed in service.

10. Comments:

Above listed simulator and temperature probe replaced by simulator serial # DDBJ-0006 and temperature probe serial DDWJP2-0088.

See Draeger packing slip.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

12. The above Instrument/Component is placed out of service pending further evaluation.

13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Steve Dellanoce #6027

 Name & Badge Number (Print)

Tpr. II [Signature] # 6027

 Signature

8/31/10

 Date

Packing Slip

Customer no Order No Order date
 150065405 10425706 08/19/2010

Packing slip no. Ship date
 80540960 08/19/2010
 Please reference on Inquiries

ship to
 150065406
 BOROUGH OF BUENA POLICE DEPARTMENT
 ATTN: CHIEF ADAMS
 616 CENTRAL AVENUE
 ATLANTIC COUNTY
 MINOTOLA, NJ 08341
 USA

bill to

Your Purchase Order Number
 08/19/2010
 SIM/PROBE WARRANTY
 Any questions? Please contact:
 MARA HERRERA

Customer:
 BOROUGH OF BUENA
 ATTN: FINANCE DEPARTMENT
 P.O. BOX 620
 ATLANTIC COUNTY
 MINOTOLA, NJ 08341-0620
 USA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Ground
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		SIM: DDXJS3-029 FAILED REPLACED FOC SER# DDBJ-0006 PROBE: DDWAP2-148 FAILED REPLACED FOC SER# DDWJP2-088
0010	1 EA	4540083 SIMULATOR, CU34 110V ELEC CTRL W/O TMETE Ordered / Back ordered : 1 / 0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150065405 10425706 08/19/2010

Packing slip no. Ship date
80540960 08/19/2010

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0020	1 EA	12025 SIMULATOR, TEMPERATURE PROBE Ordered / Back ordered : 1 / 0 EA			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 1 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

