

ALCOTEST CHECKLIST

Municipality: EVESHAM

Alcotest Ser. #: ARDC - 0066

County: BURLINGTON

Date of Calibration: 08-20-2014

1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.
2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.
Ser. #: DDDC - 0001
3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.
Ser. #: DDLJ P1 - 0033
4. **Digital Temperature Measuring System Report of Calibration.**
Ser. #: 130 602 271
5. A. Alcotest 7110 Calibration Record
 B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
 C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
 D. **Alcotest Card** of operator/coordinator who performed tests.
 E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].
Ser. #: DDUN P2 - 231
6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:
- | | | | | |
|--|-----------------|-----------|----------|------------|
| <input checked="" type="checkbox"/> A. | 0.04% Solution. | <u>12</u> | <u>H</u> | <u>104</u> |
| <input checked="" type="checkbox"/> B. | 0.08% Solution. | <u>12</u> | <u>H</u> | <u>105</u> |
| <input checked="" type="checkbox"/> C. | 0.10% Solution. | <u>13</u> | <u>I</u> | <u>122</u> |
| <input checked="" type="checkbox"/> D. | 0.16% Solution. | <u>12</u> | <u>I</u> | <u>106</u> |
7. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:
- A. 0.04% used in Calibration/Linearity Testing. DDMK 53-0003
- B. 0.08% used in Calibration/Linearity Testing. DDXD 53-0193
- C. 0.10% used in Calibration/Linearity Testing DDDC - 0001
[Same as instrument's CU34 unit].
- D. 0.16% used in Calibration/Linearity Testing. DDUN - 0052
8. A. New Standard Solution Report following Calibration.
 B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
 C. Certificate of Analysis 0.10% solution for same.
Lot #: 13 G 115
 D. Alcotest card of operator/coordinator who completed change.

Dräger

Alcotest® 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

SERIAL NUMBER:

6-27-12

ARDC-0066

Draeger Safety Diagnostics, Inc.

_____ *PC*



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

D00C-0001

Certification Date

9-24-13

Technician

BC

Re-Certification Due Date

9-24-14

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

D00L3 P1-0033

Certification date:

9-24-13

Next Certification due:

9-24-14

Probe Value

106

Dräger Safety Diagnostics, Inc.
Technical Service Department

BC

Calibration complies with ISO/IEC 17025, ANSI/NCSL Z540-1, and 9001
 Cert No.: 4000-5427014



Calibration Certificate No. 17501

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: VWR International, LLC, Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA 19087
 Instrument Identification:
 Model: 61220-601 S/N: 130602271 Manufacturer: Control Company

Standards/Equipment	Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-179	Thermistor Module	A17118	2/13/14	1000332071
Temperature Calibration Bath TC-231	Temperature Probe	128	2/20/14	6-84829-30-1
Temperature Calibration Bath TC-231	Thermistor Module	A79341	2/13/14	1000332871
Temperature Calibration Bath TC-218	Temperature Probe	3039	2/20/14	6-84829-1-1
Temperature Calibration Bath TC-218	Thermistor Module	A73332	2/13/14	1000327261
Temperature Calibration Bath TC-218	Thermistor Module	A27129	11/09/13	15-815PW-1-1
Temperature Calibration Bath TC-275	Temperature Probe	5202	11/30/14	
Digital Thermometer	Temperature Calibration Bath TC-275	B16388	8/12/14	1000341967
PRT Temperature Probe	Digital Thermometer	B16815	8/14/15	83812004

Certificate Information:
 Technician: 68 Procedure: CAL-06 Test Conditions: 24.0°C 37.0%RH 1017 mBar
 Cal Date: 10/08/13 Cal Due: 10/08/15

Unit(s)	As Found	In Tol	Normal	As Ltr	In Tol	Min	Max	±U	TUR
°C	N/A	-0.001	0.000	-0.051	0.049	0.013	0.013	3.8:1	
°C	N/A	25.001	25.000	24.951	25.051	0.023	0.023	2.2:1	
°C	N/A	50.003	50.001	49.953	50.053	0.014	0.014	3.6:1	
°C	N/A	100.001	99.994	99.951	100.051	0.018	0.018	2.8:1	

This instrument was calibrated using instruments traceable to National Institute of Standards and Technology.
 A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise noted and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 and is calculated in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 and is calculated in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 and is calculated in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM).
 Normal-Standard's Reading: As Left-Tolerance's Reading: In-Tol-Tolerance: Min = As Left Normal(Rounded) - Tolerance: Max = As Left Normal(Rounded) + Tolerance: Date: 10/08/13
 Aaron Jackson, Technical Manager

Maintaining Accuracy:
 In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change over time and may at all but can be affected by aging, temperature, shock, and contamination.

Recalibration:
 For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Fax Road Friendswood, TX 77546 USA
 Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com
 Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (IAI) American Association for Laboratory Accreditation, Certificate No. 1750.01
 International Laboratory Accreditation Cooperation (ILAC) - Mutual Recognition Arrangement (MRA)
 Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CE17-01805-2005-A0-HOU-RVA
 Traceable is a registered trademark of Control Company
 Page 1 of 1
 © 2009 Control Company

Alcotest 7110 Calibration Record

Equipment Alcotest 7110 MKIII-C Serial No.: ARDC-0066
Location: EVESHAM TOWNSHIP P.D.
Calibration File No.: 00246 **Calib. Date:** 08/20/2014 **Calib. No.:** 00011
Certification File No.: 00184 **Cert. Date:** 02/25/2014 **Cert. No.:** 00006
Linearity File No.: 00185 **Lin. Date:** 02/25/2014 **Lin. No.:** 00005
Solution File No.: 00245 **Soln. Date:** 08/17/2014 **Soln. No.:** 00045
Sequential File No.: 00246 **File Date:** 08/20/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDDC-0001
Control Solution %: 0.100% **Expires:** 09/17/2015
Solution Control Lot: 13I122 **Bottle No.:** 0763

Coordinator

Last Name: GONCALVES **First Name:** MICHELLE **MI:** L
Signature: *Sgt. Michelle Goncalves #6040* **Badge No.:** 6040
Date: 08/20/2014

*Black Key Temperature Probe Serial.....# DDUNP2-231 (NA)
*Digital NIST Temperature Measuring System Serial.....# 130602271 (NA)

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARDC-0066
Location: EVESHAM TOWNSHIP P.D.
Calibration File No.: 00246 **Calib. Date:** 08/20/2014 **Calib. No.:** 00011
Certification File No.: 00247 **Cert. Date:** 08/20/2014 **Cert. No.:** 00007
Linearity File No.: 00185 **Lin. Date:** 02/25/2014 **Lin. No.:** 00005
Solution File No.: 00245 **Soln. Date:** 08/17/2014 **Soln. No.:** 00045
Sequential File No.: 00247 **File Date:** 08/20/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDDC-0001
Control Solution %: 0.100% **Expires:** 09/17/2015
Solution Control Lot: 131122 **Bottle No.:** 0763

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:32D		
Control 1 EC	0.100%	14:33D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:33D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:34D		
Control 2 EC	0.099%	14:34D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:34D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:35D		
Control 3 EC	0.099%	14:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:36D		

All tests within acceptable tolerance.

Coordinator

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature:

Sgt. Michelle Goulet

Badge No.: 6040

Date: 08/20/2014

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARDC-0066
Location: EVESHAM TOWNSHIP P.D.
Calibration File No.: 00246 Calib. Date: 08/20/2014 Calib. No.: 00011
Certification File No.: 00247 Cert. Date: 08/20/2014 Cert. No.: 00007
Linearity File No.: 00248 Lin. Date: 08/20/2014 Lin. No.: 00006
Solution File No.: 00245 Soln. Date: 08/17/2014 Soln. No.: 00045
Sequential File No.: 00248 File Date: 08/20/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDMK S3-0003
Control Solution %: 0.040% Expires: 08/24/2014
Solution Control Lot: 12H104 Bottle No.: 0596

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXD S3-0193
Control Solution %: 0.080% Expires: 08/27/2014
Solution Control Lot: 12H105 Bottle No.: 0752

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDCN-0052
Control Solution %: 0.160% Expires: 09/10/2014
Solution Control Lot: 12I106 Bottle No.: 0374

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:50D		
Control 1 EC	0.041%	14:51D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.039%	14:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		
Control 2 EC	0.040%	14:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.040%	14:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		
Control 3 EC	0.079%	14:55D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.078%	14:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56D		
Control 4 EC	0.079%	14:57D	33.9°C	*** TEST PASSED ***
Control 4 IR	0.079%	14:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58D		
Control 5 EC	0.159%	14:59D	33.9°C	*** TEST PASSED ***
Control 5 IR	0.159%	14:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00D		
Control 6 EC	0.160%	15:01D	33.9°C	*** TEST PASSED ***
Control 6 IR	0.159%	15:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02D		

All tests within acceptable tolerance.

Coordinator

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: 

Badge No.: 6040

Date: 08/20/2014

DEPARTMENT OF
Law and Public Safety
 State is to certify that

Michelle L. Gonçalves
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN THE STATE POLICE ACADEMY TO OBTAIN THE
 TITLE OF **7110 MKII-C**
 A MEMBER IN THE STATE POLICE ACADEMY
 GIVEN EMPLOYMENT DATES AT THE TIME OF **25th** **new** **December**

THE TRAINING AND **12 CYCLES**
 SUPERVISOR
 ALTERNATE SUPERVISOR

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 12-3-13	GC PA	Wm. Lewis
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2028 (Rev. 04/10)

DEPARTMENT OF
Law and Public Safety
 State is to certify that

Michelle Gonçalves
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN THE STATE POLICE ACADEMY TO OBTAIN THE
 TITLE OF **7110 MKII-C**
 A MEMBER IN THE STATE POLICE ACADEMY
 GIVEN EMPLOYMENT DATES AT THE TIME OF **23rd** **new** **March**

THE TRAINING AND **Twelve**
 SUPERVISOR
 ALTERNATE SUPERVISOR

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2028 (Rev. 04/10)



CERTIFICATE

This is to certify that


Michelle Goncalves

has successfully completed the two day Draeger Safety Diagnostics, Inc. Alcohol Coordinator Training Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

Operator Trainer and Maintenance Technician

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: March 14, 2012



Instructor: Hansueli Ryser

Dräger



CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lieutenant Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF CRIMINAL JUSTICE
PO BOX 085
TRENTON, NJ 08625-0085
TELEPHONE: (609) 984-6500

JEFFREY S. CHIESA
Attorney General

STEPHEN J. TAYLOR
Director

March 23, 2012

Col. Joseph R. Fuentes, Superintendent
Division of State Police
Division Headquarters
P.O. Box 7068
West Trenton, New Jersey 08628

Re: Breath Test Coordinator/Instructor, Certification -
Detective I Michelle L. Goncalves #6040

Dear Col. Fuentes:

Pursuant to the provisions of *N.J.A.C.* 13:51-2.1 (b) and (c), as adopted and promulgated under the provisions of *N.J.S.A.* 39:4-50.3, 39:3-10.25 and 12:7-56, I hereby approve Detective I Michelle L. Goncalves #6040, as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,

Jeffrey S. Chiesa
Attorney General

- c. Detective I Michelle L. Goncalves, Alcohol/Drug Test Unit, Division of State Police
DSFC Patrick Guilfooy, Unit Head, Alcohol/Drug Test Unit, Division of State Police





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

JEFFREY S. CHIESA
 Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0469 to 0.0499 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 9/26/2012

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 12H104

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0487 to 0.0491 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 24, 2014.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 5th day of October, 2012.

Joseph R. Fuentes
 Notary



"An Internationally Accredited Agency"

New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

JEFFREY S. CHIESA
 Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0939 to 0.0997 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 9/27/2012

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 12H105

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0967 to 0.0976 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 27, 2014.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 5th day of October, 2012.

Notary



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

KIM GUADAGNO
 Lt. Governor

JOHN J. HOFFMAN
 Acting Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.

ANALYSIS DATE: 10/23/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 131122

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1212 to 0.1222 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 17, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 28th day of October, 2013.

Linda L. DePantis
 Notary



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

KIM GUADAGNO
 Lt. Governor

JEFFREY S. CHIESA
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1878 to 0.1994 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 10/2/2012

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 121106

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1922 to 0.1932 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 10, 2014.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 5th day of October, 2012.

Notary



"An Internationally Accredited Agency"

New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDCN-0052

Certification Date

10-1-13

Technician

BC

Re-Certification Due Date

10-1-14

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDUN P2-231

Certification date:

1-21-14

Next Certification due:

1-21-15

Probe Value

106

Draeger Safety Diagnostics, Inc.
Technical Service Department

BC



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDX1 S3-0193

Certification Date

1-22-14

Technician

BC

Re-Certification Due Date

1-22-15



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDM K S3-0003

Certification Date

1-23-14

Technician

BC

Re-Certification Due Date

1-23-15

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARDC-0066
Location:	EVESHAM TOWNSHIP P.D.	Calib. No.:	00011
Calibration File No.:	00246	Cert. Date:	08/20/2014
Certification File No.:	00247	Lin. Date:	08/20/2014
Linearity File No.:	00248	Soln. Date:	08/20/2014
Solution File No.:	00249	File Date:	08/20/2014
Sequential File No.:	00249		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDDC-0001
Control Solution %:	0.100%			Expires:	07/18/2015
Solution Control Lot:	13G115			Bottle No.:	1300

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:08D		
Control 1 EC	0.101%	16:09D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:10D		
Control 2 EC	0.101%	16:11D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:11D		
Control 3 EC	0.100%	16:12D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDLJPI-0033 MGA

Changed By:

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: *Sgt. Michelle Goncalves #6040*

Badge No.: 6040

Date: 08/20/2014



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

KIM GUADAGNO
 Lt. Governor

JOHN J. HOFFMAN
 Acting Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.

ANALYSIS DATE: 7/29/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 13G115

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1213 to 0.1220 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 18, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 7th day of August, 2013.

Joseph R. Fuentes
 Notary



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable

