

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Evesham Township PD 984 Tuckerton Road Marlton, NJ 08053		2. Contact: Sgt. Ron Ritter	4. Date: 6-21-2012
		3. Phone Number: 856-983-1111	5. County: Burlington
6. Alcotest Instrument Serial Number: ARDC-0066	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
During the control test the instrument lost power.

10. Comments:
Instrument is being sent back to Draeger for further evaluation.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Det. I M. Goncalves #6040

Name & Badge Number (Print)

Det I M Goncalves #6040 6/21/12

Signature

Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Evesham Twp , PD , Marlton , NJ

Date Received: 06/26/12

Date given to service: 06/26/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR DC-0066

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
1815113	FUSE	1	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Installed fuse to restore power to unit , ran 50x .1% test , then calibrated unit , then ran 50X .1 % test with no lose of power issues,also left unit off all night, turned unit on next morning no power issue. Unit was also taken apart and completely inspected.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 06-27-12

Packing Slip

Customer no Order No Order date
 150055418 10558795 06/28/2012

Packing slip no. Ship date
 80698150 06/28/2012

Please reference on inquiries

ship to

150055418
 EVESHAM TWP POLICE DEPT
 ATTN: SGT. RON RITTER
 BURLINGTON COUNTY
 984 TUCKERTON RD
 MARLTON, NJ 08053
 USA

bill to

Your Purchase Order Number

7110: ARDC-0066
 Any questions? Please contact:
 MARA HERRERA

Customer

EVESHAM TWP POLICE DEPT
 ATTN: SGT. RON RITTER
 BURLINGTON COUNTY
 984 TUCKERTON RD
 MARLTON, NJ 08053
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARDC-0066			
0010	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0020	1 EA	1815113 FUSE, 2.00A 250V FAST G : Ordered / Back ordered : 0 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150055418 10558795 06/28/2012

Packing slip no. Ship date
80698150 06/28/2012

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			