

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Evesham Township PD 984 Tuckerton Road Marlton, NJ 08053	2. Contact: Sgt. Ron Ritter	4. Date: 7-24-2012
	3. Phone Number: 856-983-1111	5. County: Burlington
6. Alcotest Instrument Serial Number: ARUL-0061	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
The installation of the Eco-Surge Tripp-Lite surge protector.

10. Comments:
The installation of the Eco-Surge Tripp-Lite surge protector was conducted as per Draegers recommendation. Model number TLP808NETG. The instruments phone line is also routed through the surge protector.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Det. I Goncalves #6040

Name & Badge Number (Pnnt)

Det. I Goncalves #6040 7/24/12

Signature

Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Evesham Township PD 984 Tuckerton Road Marlton, NJ 08053	2. Contact: Sgt. Ron Ritter	4. Date: 7-24-2012
3. Phone Number: 856-983-1111		5. County: Burlington
6. Alcotest Instrument Serial Number: ARDC-0066	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
The installation of the Eco-Surge Tripp-Lite surge protector.

10. Comments:
The installation of the Eco-Surge Tripp-Lite surge protector was conducted as per Draegers recommendation. Model number TLP808NETG. The instruments phone line is also routed through the surge protector.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Det. I M. Goncalves #6040

Name & Badge Number (Print)

Det. I M. Goncalves #6040 7/24/12

Signature

Date

DEPARTMENT OF
Traffic and Public Safety
Whose is to certify that

Michelle L. McKelvey
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTS ACCORDING TO CHAPTER 1602 OF THE LAWS OF 1996 IN THE OPERATION OF THE **Breathalyzer**
A METHOD TO DETERMINE BLOOD ALCOHOL CONTENT
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY ON 16th day of Aug
TWO THOUSAND AND 06

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES Feb. 3-7-2003

Refresher Course

DATE	PLACE	INSTRUCTOR
1. <u>12/10/02</u>	<u>OCPA</u>	<u>CP</u>
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2538 (Rev. 05/02)

DEPARTMENT OF
Traffic and Public Safety
Whose is to certify that

MICHELLE L. MCKELVEY
NEW JERSEY STATE POLICE

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTS ACCORDING TO CHAPTER 1602 OF THE LAWS OF 1996 IN THE OPERATION OF THE **ALCO TEST 7110 MKIII-C**
A METHOD TO DETERMINE BLOOD ALCOHOL CONTENT
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY ON 10th day of August
TWO THOUSAND AND FIVE

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2538 (Rev. 05/02)

DEPARTMENT OF
Traffic and Public Safety
Whose is to certify that

Michelle L. Goncalves
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTS ACCORDING TO CHAPTER 1602 OF THE LAWS OF 1996 IN THE OPERATION OF THE **Alcotest 7110 MKIII-C**
A METHOD TO DETERMINE BLOOD ALCOHOL CONTENT
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY ON 15th day of December
TWO THOUSAND AND FIVE

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2538 (Rev. 05/02)

DEPARTMENT OF
Traffic and Public Safety
Whose is to certify that

Michelle Goncalves
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTS ACCORDING TO CHAPTER 1602 OF THE LAWS OF 1996 IN THE OPERATION OF THE **Breath Test Coordination Instructor**
A METHOD TO DETERMINE BLOOD ALCOHOL CONTENT
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY ON 23rd day of March
TWO THOUSAND AND Twelve

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2538 (Rev. 05/02)

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that
Michelle L. McKelvey
 IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTERS PURSUANT TO CHAPTER 1407
 OF THE LAWS OF THE STATE OF NEW JERSEY AND TO OFFER ME BY
 A METHOD TO DETERMINE BREATH ALCOHOL
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 16th DAY OF Aug
 TWO THOUSAND AND 06

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES Feb. 3-7-2003
 Refresher Course

DATE	PLACE	INSTRUCTOR
1. <u>12/10/03</u>	<u>OCRA</u>	<u>C. [Signature]</u>
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2508 (Rev. 05/02)

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that
MICHELLE L. MCKELVEY
 NEW JERSEY STATE POLICE
 IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTERS PURSUANT TO CHAPTER 1407
 OF THE LAWS OF THE STATE OF NEW JERSEY AND TO OFFER ME BY
 A METHOD TO DETERMINE BREATH ALCOHOL
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 10th DAY OF August
 TWO THOUSAND AND FIVE

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES _____
 Refresher Course

DATE	PLACE	INSTRUCTOR
1. _____	_____	_____
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2508 (Rev. 05/02)

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that
Michelle L. Gonçalves
 New Jersey State Police
 IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTERS PURSUANT TO CHAPTER 1407
 OF THE LAWS OF THE STATE OF NEW JERSEY AND TO OFFER ME BY
 A METHOD TO DETERMINE BREATH ALCOHOL
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 15th DAY OF December
 TWO THOUSAND AND ELECT

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES _____
 Refresher Course

DATE	PLACE	INSTRUCTOR
1. _____	_____	_____
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2509 (Rev. 05/04)

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that
Michelle Gonçalves
 New Jersey State Police
 IS QUALIFIED AND COMPETENT TO CONDUCT COURSE IN BREATH TESTERS PURSUANT TO CHAPTER 1407
 OF THE LAWS OF THE STATE OF NEW JERSEY AND TO OFFER ME BY
 A METHOD TO DETERMINE BREATH ALCOHOL
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 23rd DAY OF March
 TWO THOUSAND AND Twelve

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES _____
 Refresher Course

DATE	PLACE	INSTRUCTOR
1. _____	_____	_____
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2509 (Rev. 05/04)

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Evesham Twp. Police Departmen 984 Tuckerton Road Marlton, NJ 08053	2. Contact: Ptl. B. Lacarte #31	4. Date: 07-26-2010
	3. Phone Number: 1-856-983-1116	5. County: Burlington
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARUL-0061</p>	7. Simulator Component Serial Number:	8. Temperature Probe Component Serial Number:


9. Reason for Service:
Responded to Department in reference to instrument returning from Draeger Safety Diagnostics Inc.

10. Comments:
Fuel cell replaced, plates replaced, packing slip received.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. H. Weyman #6289
 Name & Badge Number (Print)


 Signature #6289 7/26/2010
 Date

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Evesham Twp. Police Departmen 984 Tuckerton Road Marlton, NJ 08053	2. Contact: Ptl. B. Lacarte #31 3. Phone Number: 1-856-983-1116	4. Date: 07-15-2010 5. County: Burlington
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARUL-0061</p>	7. Simulator Component Serial Number:	8. Temperature Probe Component Serial Number:


9. Reason for Service:
Arrived at PD to perform calibration. Upon arrival, received complaint of instrument experiencing repeated control test failures.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several repeated control test failures due to low (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. H. Weyman #6289
 Name & Badge Number (Print)


 Signature 7/15/2010
 Date

Packing Slip

Customer no Order No Order date
 150055418 10420260 07/21/2010

Packing slip no. Ship date
 80534591 07/21/2010

Please reference on inquiries

ship to

150055418
 EVESHAM TWP POLICE DEPT
 ATTN: NICK MATTEO
 984 TUCKERTON RD
 BURLINGTON COUNTY
 MARLTON, NJ 08053
 USA

bill to

Your Purchase Order Number

7110 WARRANTY: ARUL-0061
 Any questions? Please contact:
 MARA HERRERA

Customer

EVESHAM TWP POLICE DEPT
 ATTN: NICK MATTEO
 984 TUCKERTON RD
 BURLINGTON COUNTY
 MARLTON, NJ 08053
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110: ARUL-0061
0010	1 EA	6808486 PLATES,7110 Ordered / Back ordered : 0 / 0 EA REPLACED
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150055418 10420260 07/21/2010

Packing slip no. Ship date
80534591 07/21/2010

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Evesham Township PD 984 Tuckerton Road Marlton, NJ 08053		2. Contact: Sgt. Ron Ritter	4. Date: 7-24-2012
		3. Phone Number: 856-983-1111	5. County: Burlington
6. Alcotest Instrument Serial Number: ARDC-0066	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
The above Alcotest was returned from outside evaluation and/or repair. A successful calibration was conducted utilizing the above Alcotest.


10. Comments:
See attached packing slip and Drager Return and Repair Form.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Det. I M. Goncalves #6040

Name & Badge Number (Print)


Signature Date