

ALCOTEST CHECKLIST

Municipality: FLORENCE TWP Alcotest Ser.#: ARTN-0053

County: BURLINGTON Date of Calibration: 02-08-2022

1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.

2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.
Ser. #: DDUM 53-0225

3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.
Ser. #: DDWT P2-144

4. **Digital Temperature Measuring System Report of Calibration.**
Ser. #: 200 357 843

5. **NIST - Traceable Digital Thermometer Readings.**

6. A. Alcotest 7110 Calibration Record
 B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
 C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
 D. **Alcotest Card** of operator/coordinator who performed tests.
 E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].
Ser. #: DDEE P2-060

7. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:

<input checked="" type="checkbox"/> A.	0.04% Solution.	<u>20260</u>
<input checked="" type="checkbox"/> B.	0.08% Solution.	<u>20270</u>
<input checked="" type="checkbox"/> C.	0.10% Solution.	<u>20220</u>
<input checked="" type="checkbox"/> D.	0.16% Solution.	<u>20280</u>

8. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:

A. 0.04% used in Calibration/Linearity Testing. DDWE 53-0211
 B. 0.08% used in Calibration/Linearity Testing. DDXC 53-0020
 C. 0.10% used in Calibration/Linearity Testing.
[Same as CU34 unit on instrument.]
 D. 0.16% used in Calibration/Linearity Testing. DDMK 53-0008

9. A. New Standard Solution Report following Calibration.
 B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
 C. Certificate of Analysis 0.10% solution for same.
Lot #: 20320
 D. Alcotest card of operator/coordinator who completed change.

Dräger

Alcotest 7110

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48864, 49 FR 48864, and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your state's specifications.

Certification Date:

Serial Number:

03/25/2020

ARTAL-0053

Dräger, Inc. 

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- X-Cal 2000 (Alcosim)
- Other: _____

Serial Number:

DDUMS3-D225

Certification Date:

Technician:

Re-Certification Due Date:

11.30.21

MB

11.30.22

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

Certification Date:

Next Certification Due:

DDWJP2-144

11.30.21

11.30.22

Probe Value:

101

Draeger, Inc.

MB



Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4000-11349797

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by VWR International LLC Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA, 19087

Instrument Identification:

Model: 61220-601,

S/N: 200357843

Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Thermistor Module	A27129	04 Feb 2021	1000451212
Temperature Calibration Bath	A42238		
Temperature Calibration Bath	B01375		
Temperature Probe	5394	21 Feb 2021	C0220030
Temperature Calibration Bath	B16388		
Temperature Probe	5267	21 Feb 2021	C0220028
Temperature Calibration Bath	B3A444		
Thermistor Module	B96381	16 Jul 2020	B9626028
Temperature Probe	5398	16 Jul 2020	B9605083
Thermistor Module	B96382	19 Aug 2020	B9628006
Temperature Probe	5410	13 Sep 2020	B9801031

Certificate Information:

Technician: 420

Procedure: CAL-06

Cal Date: 15 Jun 2020

Cal Due Date: 15 Jun 2022

Test Conditions: 52.44%RH 23.46°C 1018mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C	N.A.	N.A.		0.000	0.001	Y	-0.05	0.05	0.0087	>4:1
°C	N.A.	N.A.		25.001	25.001	Y	24.951	25.051	0.0087	>4:1
°C	N.A.	N.A.		50.002	50.001	Y	49.952	50.052	0.0087	>4:1
°C	N.A.	N.A.		100.001	99.998	Y	99.951	100.051	0.0087	>4:1

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy= $\pm(\text{Max}-\text{Min})/2$; Min=As Left Nominal(Rounded) - Tolerance; Max=As Left Nominal(Rounded) + Tolerance;

Nicol Rodriguez

Nicol Rodriguez, Quality Manager

Marisa Elms

Marisa Elms, Technical Manager

Note -

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometer change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by (AZLA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2015 Quality Certified by DNV GL, Certificate No. CERT-01805-2006-AQ-HOU-ANAB.
International Laboratory Accreditation Cooperation - Multilateral Recognition Arrangement (ILAC-MRA).



Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4000-11349797

Traceable® Certificate of Calibration for Digital Thermometer

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

Issue Date: 15 Jun 2020

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by (AZLA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2015 Quality Certified by DNV GL, Certificate No. CERT-01805-2008-AD-HOU-ANAB.
International Laboratory Accreditation Cooperation - Multilateral Recognition Arrangement (ILAC-MRA).

**Alcotest 7110 MKIII-C Calibration
NIST-Traceable Digital Thermometer Readings**

Coordinator:

Tpr I Dennis J Lotz
Name

7045
Badge No.

Location:

Florence Township P.O.
Agency

ARTN-0053
Alcotest Serial No.

Equipment:

260 357 843
Digital NIST Temperature Measuring System Serial No.

Simulator Solution Concentration	CU-34 Simulator Serial No.	Time Simulators Started to Heat	Time Temp. Reading Obtained	Temp. Reading on NIST Traceable Thermometer
0.04%	DDWE 83-0211	08:06 ^s	09:10 ^s	33.9 ^o C
0.08%	DDXC 83-0020	08:08 ^s	09:11 ^s	33.9 ^o C
0.10%	DDUM 83-0225	08:08 ^s	09:13 ^s	33.9 ^o C
0.16%	DDMK 83-0008	08:08 ^s	09:15 ^s	33.9 ^o C

Pursuant to law and the "Chemical Breath Testing Regulations" established at N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity and consistent with the "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on Alcotest 7110 MKIII-C instruments. Pursuant to and consistent with the current "Calibration Check Procedure for Alcotest 7110", I performed a Calibration Check Procedure on the Alcotest 7110 MKIII-C instrument identified on this certificate. Pursuant to the current "Calibration Check Procedure for Alcotest 7110", I used the Digital NIST-traceable Temperature Measuring System identified on this certificate to confirm that the temperatures of the 0.10%, 0.04%, 0.08%, and 0.16% Simulator Solutions used in the respective CU-34 Simulators identified on this certificate, were 34.0 degrees Celsius \pm 0.2 degrees Celsius. I hereby certify that I truthfully recorded on this certificate the temperatures of each of the simulator solutions as shown on the Digital NIST-traceable Temperature Measuring System thermometer. I am aware that if any of the foregoing statements made by me are willfully false, I am subject to punishment.

Tpr I Dennis J Lotz 7045
Coordinator's Signature

2-8-22
Date

Alcotest 7110 Calibration Record

Equipment


Alcotest 7110 MKIII-C
Location: FLORENCE TOWNSHIP P.D.
Serial No.: ARTN-0053
Calibration File No.: 01161 Calib. Date: 02/08/2022 Calib. No.: 00047
Certification File No.: 01144 Cert. Date: 09/08/2021 Cert. No.: 00038
Linearity File No.: 01145 Lin. Date: 09/08/2021 Lin. No.: 00037
Solution File No.: 01158 Soln. Date: 01/16/2022 Soln. No.: 00269
Sequential File No.: 01161 File Date: 02/08/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 05/06/2022
Solution Control Lot: 20220 Bottle No.: 0321

Coordinator

Last Name: LUTZ First Name: DENNIS MI: J
Signature: Tpr I [Signature] 7045 Badge No.: 7045
Date: 02/08/2022

*Black Key Temperature Probe Serial.....# DDEEP2-060 (DL)
*Digital NIST Temperature Measuring System Serial.....# 200 357 843 (DL)



Pursuant to law; and the "Chemical Breath Testing Regulations" N.J.A.C. 17:27, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01161 Calib. Date: 02/08/2022 Calib. No.: 00047
Certification File No.: 01162 Cert. Date: 02/08/2022 Cert. No.: 00039
Linearity File No.: 01145 Lin. Date: 09/08/2021 Lin. No.: 00037
Solution File No.: 01158 Soln. Date: 01/16/2022 Soln. No.: 00269
Sequential File No.: 01162 File Date: 02/08/2022


Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 05/06/2022
Solution Control Lot: 20220 Bottle No.: 0321

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:27S		
Control 1 EC	0.099%	09:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28S		
Control 2 EC	0.099%	09:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		
Control 3 EC	0.099%	09:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31S		

All tests within acceptable tolerance.

Coordinator

Last Name: LUTZ

Signature: 

First Name: DENNIS

MI: J

Badge No.: 7045

Date: 02/08/2022

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01161 Calib. Date: 02/08/2022 Calib. No.: 00047
Certification File No.: 01162 Cert. Date: 02/08/2022 Cert. No.: 00039
Linearity File No.: 01163 Lin. Date: 02/08/2022 Lin. No.: 00038
Solution File No.: 01158 Soln. Date: 01/16/2022 Soln. No.: 00269
Sequential File No.: 01163 File Date: 02/08/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0211
Control Solution %: 0.040% Expires: 06/08/2022
Solution Control Lot: 20260 Bottle No.: 0165

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0020
Control Solution %: 0.080% Expires: 06/11/2022
Solution Control Lot: 20270 Bottle No.: 0212

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDMK S3-0008
Control Solution %: 0.160% Expires: 06/17/2022
Solution Control Lot: 20280 Bottle No.: 0060

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:47S		
Control 1 EC	0.041%	09:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.040%	09:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49S		
Control 2 EC	0.041%	09:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.039%	09:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:52S		
Control 3 EC	0.082%	09:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.080%	09:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54S		
Control 4 EC	0.081%	09:54S	33.9°C	*** TEST PASSED ***
Control 4 IR	0.080%	09:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		
Control 5 EC	0.161%	09:56S	34.0°C	*** TEST PASSED ***
Control 5 IR	0.161%	09:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		
Control 6 EC	0.159%	09:59S	34.0°C	*** TEST PASSED ***
Control 6 IR	0.160%	09:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00S		

All tests within acceptable tolerance.

Coordinator

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: _____

Tpc I Dent 7045

Badge No.: 7045

Date: 02/08/2022

DEPARTMENT OF
Traffic and Public Safety
Wishes to be certified that

Dennis J. Lutz

Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE, IN ACCORDANCE WITH CHAPTER 170B OF THE LAWS OF 1966 IN THE OPERATION OF THE
 A METHOD TO DETERMINE INTOXICATION
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29th DAY OF January
 TWO THOUSAND AND Nineteen
 / s/ Dennis J. Lutz / s/ [Signature]
 SUPERINTENDENT ATTORNEY GENERAL
 NEW JERSEY STATE POLICE STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2938 (Rev. 07/07)

DEPARTMENT OF
Traffic and Public Safety
Wishes to be certified that

Dennis J. Lutz
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT COURSE, IN ACCORDANCE WITH CHAPTER 170B OF THE LAWS OF 1966 IN THE OPERATION OF THE
 A METHOD TO DETERMINE INTOXICATION
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 1st DAY OF October
 TWO THOUSAND AND Nine
 / s/ [Signature] / s/ [Signature]
 SUPERINTENDENT ATTORNEY GENERAL
 NEW JERSEY STATE POLICE STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 2-3-11	OCPA	Wm Horn
2. 1/24/13	OCPA	Adam Standa
3. 11-23-15	GCPA	M. Gonzalez
4. 4/6/17	LAKEHURST	Adam Standa
5. 8/22/19	NJSP Gallenay	[Signature]
6. 7-19-21	Lakehurst	[Signature]
7.		
8.		
9.		

S.P. 2938 (Rev. 07/07)



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

SHEILA Y. OLIVER
Lt. Governor

GURBIR S. GREWAL
Attorney General

PATRICK J. CALLAHAN
Colonel

CERTIFICATION OF ANALYSIS
0.040 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0469 to 0.0499 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/29/2020

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 20260

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0481 to 0.0486 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 08, 2022.

As Assistant Chief Forensic Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Assistant Chief Forensic Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 18th day of August, 2020.
Maureen Keckel
Notary



An Internationally Accredited Agency

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 832-2000

PHILIP D MURPHY
Governor

SHEILA Y OLIVER
Lt. Governor

GURBIR S. GREWAL
Attorney General

PATRICK J. CALLAHAN
Colonel

CERTIFICATION OF ANALYSIS
0.080 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0939 to 0.0997 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 08/07/2020

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 20270

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0968 to 0.0974 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 11, 2022.

As Assistant Chief Forensic Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Assistant Chief Forensic Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 18th day of August, 2020.
Notary



An Internationally Accredited Agency

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-3000

PHILIP D. MURPHY
Governor

SHEILA Y. OLIVER
Lt. Governor

GURBIR S. GREWAL
Attorney General

PATRICK J. CALLAHAN
Colonel

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 05/22/2020

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 20220

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1204 to 0.1227 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is May 06, 2022.

As Assistant Chief Forensic Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Assistant Chief Forensic Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 27th day of May, 2020.
Notary



An Internationally Accredited Agency
New Jersey is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

SHEILA Y. OLIVER
Lt. Governor

GURBIR S. GREWAL
Attorney General

PATRICK J. CALLAHAN
Colonel

CERTIFICATION OF ANALYSIS
0.160 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1878 to 0.1994 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/17/2020

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 20280


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1949 to 0.1977 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 17, 2022.

As Assistant Chief Forensic Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Assistant Chief Forensic Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 28 day of July, 2020.
Notary



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- X-Cal 2000 (Alcosim)
- Other: _____

Serial Number:

DDWES3-D211

Certification Date:

Technician:

Re-Certification Due Date:

3-18-21

MB

3-18-22

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- X-Cal 2000 (Alcosim)
- Other: _____

Serial Number:

DDXCS3-0020

Certification Date:

Technician:

Re-Certification Due Date:

6-21-21

AM

6-21-22

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

Model: ALCOTEST CU34

Model: MARK IIA

X-Cal 2000 (Alcosim)

Other: _____

Serial Number:

DDMKS3-0008

Certification Date:

6.21.21

Technician:

CPA

Re-Certification Due Date:

6.21.22

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110

Serial Number Temp Probe:

DDEEP2-060

Certification Date:

6.21.21

Next Certification Due:

6.21.22

Probe Value:

106

Draeger, Inc.

AM

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	01161	Calib. Date:	02/08/2022
Certification File No.:	01162	Cert. Date:	02/08/2022
Linearity File No.:	01163	Lin. Date:	02/08/2022
Solution File No.:	01164	Soln. Date:	02/08/2022
Sequential File No.:	01164	File Date:	02/08/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0225
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	0576

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:12S		
Control 1 EC	0.099%	11:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:13S		
Control 2 EC	0.099%	11:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		
Control 3 EC	0.099%	11:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2-1441 DU

Changed By:

Last Name: LUTZ	First Name: DENNIS	MI: J
Signature: <u>TP I Dent 7045</u>	Badge No.:	7045
	Date:	02/08/2022



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 832-2000

PHILIP D. MURPHY
Governor

SHEILA Y. OLIVER
Lt. Governor

GURBIR S. GREWAL
Attorney General

PATRICK J. CALLAHAN
Colonel

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/27/2020

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 20320

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1211 to 0.1222 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 06, 2022.

As Assistant Chief Forensic Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Assistant Chief Forensic Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 29 day of July, 2020.

Notary



An Internationally Accredited Agency

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable

