

# ALCOTEST CHECKLIST

Municipality: FLORENCE TWP. Alcotest Ser.#: ARTN-0053

County: BURLINGTON Date of Calibration: 03-28-2011

1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.

2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.  
Ser. #: DDUM 53-0222

3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.  
Ser. #: DDWI Pa-293

4. **Digital Temperature Measuring System Report of Calibration.**  
Ser. #: 101733 534

5.  A. Alcotest 7110 Calibration Record  
 B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**  
 C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**  
 D. **Alcotest Card** of operator/coordinator who performed tests.  
 E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].  
Ser. #: DDWA Pa-016

6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:

<input checked="" type="checkbox"/>	A.	0.04% Solution.	<u>10</u>	<u>A</u>	<u>073</u>
<input checked="" type="checkbox"/>	B.	0.08% Solution.	<u>10</u>	<u>A</u>	<u>074</u>
<input checked="" type="checkbox"/>	C.	0.10% Solution.	<u>10</u>	<u>F</u>	<u>080</u>
<input checked="" type="checkbox"/>	D.	0.16% Solution.	<u>10</u>	<u>A</u>	<u>075</u>

7. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:

- A. 0.04% used in Calibration/Linearity Testing. DDWE 53-0196
- B. 0.08% used in Calibration/Linearity Testing. DDWE 53-0203
- C. 0.10% used in Calibration/Linearity Testing DDUM 53-0222 [Same as instrument's CU34 unit].
- D. 0.16% used in Calibration/Linearity Testing DDWE 53-0205

8.  A. New Standard Solution Report following Calibration.  
 B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].  
 C. Certificate of Analysis 0.10% solution for same.  
Lot #: 10 F 079  
 D. Alcotest card of operator/coordinator who completed change.

7-12-2011

**Dräger**

**Alcotest® 7110 MKIII-C**

**CERTIFICATE OF ACCURACY**

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

SERIAL NUMBER:

3/21/2011

ARTN-0053

Dräger Safety Diagnostics, Inc.

  
\_\_\_\_\_



**NEW JERSEY STATE POLICE**  
**Alcohol/Drug Testing Unit**

**Trooper Steve Dellanoce #6027**  
Breath Test Coordinator / Instructor

Hamilton Technology Complex  
Suite 400  
1200 Negron Drive  
Hamilton, NJ 08891

(609) 584-5000 ext. 5808  
cell: (609) 947-9211  
Fax: (609) 584-8359  
email: LPP6027@gw.njsp.org



Drägersafety

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Dräger Safety Diagnostics, Inc.**

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: \_\_\_\_\_

Serial Number:

DDUM53-0222

Certification Date

Technician

Re-Certification Due Date

11-17-10

mm

11-17-11

## Drägersafety

### ALCOTEST® 7110 TEMPERATURE PROBE

# CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDWTP2-293

11-17-10

11-17-11

Probe Value

105

Dräger Safety Diagnostics, Inc.  
Technical Service Department

mm



Calibration  
Certificate No. 1750.01

Calibration complies with ISO 9001  
ISO/IEC 17025 AND ANSI/NCSL Z540-1



Cert. No.: 4000-2966289

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: VWR International, P.O. Box 2158, Secaucus, NJ 07094 U.S.A.

Instrument Identification:

Model: 61220-601      S/N: 101733534      Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-179	A45240		
Thermistor Module	A17118	11/19/10	A9B21010
Temperature Probe	128	12/10/10	A9B23079
Temperature Calibration Bath TC-231	A78341		
Temperature Probe	3039	12/10/10	A9B23080-1
Temperature Calibration Bath TC-218	A73332		
Thermistor Module	A27129	7/09/10	1000264338
Temperature Probe	5202	3/11/11	B0310050
Temperature Calibration Bath TC-255	B01375		
Temperature Probe	157	7/27/10	A9708011-4

Certificate Information:

Technician: 104      Procedure: CAL-06      Cal Date: 6/08/10      Cal Due: 6/08/12  
Test Conditions: 24.0°C      42.0%RH      1015 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C		N.A.		0.002	-0.001	Y	-0.048	0.052	0.013	3.8:1
°C		N.A.		24.999	25.000	Y	24.949	25.049	0.013	3.8:1
°C		N.A.		60.001	59.998	Y	59.951	60.051	0.018	2.8:1
°C		N.A.		100.001	100.002	Y	99.951	100.051	0.013	3.8:1

This instrument was calibrated using instruments traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal/Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min = Nominal/Rounded - Tolerance; Max = Nominal/Rounded + Tolerance; Date=MM/DD/YY

*Nicol Rodriguez*  
Nicol Rodriguez, Quality Manager

*Wallace Berry*  
Wallace Berry, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change bits, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA  
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.  
Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-2006-AD-HOU-ANAB.  
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).

# Alcotest 7110 Calibration Record

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
Location: FLORENCE TOWNSHIP P.D.  
Calibration File No.: 00458 Calib. Date: 03/28/2011 Calib. No.: 00017  
Certification File No.: 00428 Cert. Date: 01/05/2011 Cert. No.: 00009  
Linearity File No.: 00429 Lin. Date: 01/05/2011 Lin. No.: 00009  
Solution File No.: 00456 Soln. Date: 03/21/2011 Soln. No.: 00105  
Sequential File No.: 00458 File Date: 03/28/2011


Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0222  
Control Solution %: 0.100% Expires: 06/14/2012  
Solution Control Lot: 10F080 Bottle No.: 0371

## Coordinator

Last Name: DELLANOCE First Name: JOSEPH MI: S  
Signature: Joe Dellanoce 6027 Badge No.: 6027  
Date: 03/28/2011

\*Black Key Temperature Probe Serial.....# DDWAP2-016 *js*

\*Digital NIST Temperature Measuring System Serial.....# 101733534 *js*



Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator in my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis which are utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to and consistent with the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

# Alcotest 7110 Calibration Certificate

## Part I - Control Tests

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
**Location:** FLORENCE TOWNSHIP P.D.  
**Calibration File No.:** 00458 **Calib. Date:** 03/28/2011 **Calib. No.:** 00017  
**Certification File No.:** 00459 **Cert. Date:** 03/28/2011 **Cert. No.:** 00010  
**Linearity File No.:** 00429 **Lin. Date:** 01/05/2011 **Lin. No.:** 00009  
**Solution File No.:** 00456 **Soln. Date:** 03/21/2011 **Soln. No.:** 00105  
**Sequential File No.:** 00459 **File Date:** 03/28/2011

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUM S3-0222  
**Control Solution %:** 0.100% **Expires:** 06/14/2012  
**Solution Control Lot:** 10F080 **Bottle No.:** 0371

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:48S	34.0°C	*** TEST PASSED ***
Control 1 EC	0.099%	12:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49S	34.0°C	*** TEST PASSED ***
Control 2 EC	0.099%	12:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50S	34.0°C	*** TEST PASSED ***
Control 3 EC	0.099%	12:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52S	34.0°C	*** TEST PASSED ***

All tests within acceptable tolerance.

**Coordinator**

Last Name: DELANOCE

Signature: 

Date: 03/28/2011

Pursuant to law, and the "Chemical Breath Test Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Inspector in my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I performed calibration checks on approved instruments employing infrared analysis and electrochemical analysis which are utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to and consistent with, the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

# Alcotest 7110 Calibration Certificate

## Part II - Linearity Tests

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
**Location:** FLORENCE TOWNSHIP P.D.  
**Calibration File No.:** 00458 **Calib. Date:** 03/28/2011 **Calib. No.:** 00017  
**Certification File No.:** 00459 **Cert. Date:** 03/28/2011 **Cert. No.:** 00010  
**Linearity File No.:** 00460 **Lin. Date:** 03/28/2011 **Lin. No.:** 00010  
**Solution File No.:** 00456 **Soln. Date:** 03/21/2011 **Soln. No.:** 00105  
**Sequential File No.:** 00460 **File Date:** 03/28/2011

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0196  
**Control Solution %:** 0.040% **Expires:** 01/12/2012  
**Solution Control Lot:** 10A073 **Bottle No.:** 0042

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0203  
**Control Solution %:** 0.080% **Expires:** 01/15/2012  
**Solution Control Lot:** 10A074 **Bottle No.:** 1031

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWE S3-0205  
**Control Solution %:** 0.160% **Expires:** 01/21/2012  
**Solution Control Lot:** 10A075 **Bottle No.:** 0842

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:07S	14:07D	
Control 1 EC	0.039%	13:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.040%	13:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09S		
Control 2 EC	0.039%	13:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.040%	13:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10S		
Control 3 EC	0.078%	13:11S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.078%	13:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12S		
Control 4 EC	0.078%	13:13S	34.0°C	*** TEST PASSED ***
Control 4 IR	0.078%	13:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14S		
Control 5 EC	0.155%	13:15S	34.0°C	*** TEST PASSED ***
Control 5 IR	0.156%	13:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:16S		
Control 6 EC	0.156%	13:17S	34.0°C	*** TEST PASSED ***
Control 6 IR	0.158%	13:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18S	14:18D gsp	

All tests within acceptable tolerance.

### Coordinator

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: Joseph Dellanoce #6027

Badge No.: 6027

Date: 03/28/2011

DEPARTMENT OF  
*Justice and Public Safety*

Joseph S. Dellanoce  
 New Jersey State Police

is challenged and competent to conduct professional duties as a member according to CHAPTER 102 OF  
 TITLE 17B OF THE STATUTES OF THE STATE OF NEW JERSEY AND TO CHAPTER 102 OF  
 TITLE 17B OF THE STATUTES OF THE STATE OF NEW JERSEY AS AMENDED BY CHAPTER 7110 OF TITLE 17B  
 AS AMENDED TO EXTEND THE AUTHORITY OF THE BOARD OF EXAMINERS TO EXTEND THE AUTHORITY OF THE BOARD OF EXAMINERS  
 TO EXTEND THE AUTHORITY OF THE BOARD OF EXAMINERS TO EXTEND THE AUTHORITY OF THE BOARD OF EXAMINERS

29th day of March

FIVE

*[Signature]*  
 SUPERVISOR  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
2 2/5/07	GC PA	C. [Signature]
2 12/7/09	DC PA	[Signature]
3		
4		
5		
6		
7		
8		
9		

S.P. 2010 (REV. 12/04)



DEPARTMENT OF  
**Traffic and Public Safety**  
What is to be certified that

Joseph S. Dellanoce  
 New Jersey State Police

IS QUALIFIED AND COMMITTED TO COMPLETE THE COURSE AND TO INSTRUCT PURSUANT TO CHAPTER 16 OF  
 THE LAWS OF NEW JERSEY IN THE OPERATION OF THE Breath Test Coordinator/Instructor  
 A METHOD TO DEVELOPE INTO A COURSE  
 COVERED UNDER MY NAME AT THE OFFICE, NEW JERSEY POLICE

14th DAY OF January

TWO THOUSAND AND EIGHT

  
 JOSEPH S. DELLANOCE  
 NEW JERSEY STATE POLICE

  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>12/17/09</u>	<u>OC.PA</u>	<u>Adam Standa</u>
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

DEPARTMENT OF  
**Public Safety**

This is to certify that  
**Joseph B. Ballanocè**  
New Jersey State Police

IS QUALIFIED AND COMPETENT TO PERFORM THE DUTIES AND FUNCTIONS ASSIGNED PURSUANT TO CHAPTER  
143 OF TITLE 17A OF LAWS OF 1975 OF THE STATE OF NEW JERSEY, AS AMENDED, IN THE  
METHODS TO DETERMINE DRUG CONCENTRATIONS IN BREATH BY THE **Breathalyzer**

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 13th day of Dec  
TWO THOUSAND AND 01

*Car. P. Kelly*  
SUPERINTENDENT  
NEW JERSEY STATE POLICE

*John P. ...*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

ORIGINAL COURSE DATES: 8/30 - 9/3/99

Refresher Course		
DATE	PLACE	INSTRUCTOR
1. <u>12/12/01</u>	<u>A.G.T.C.</u>	<u>Potter</u>
2. <u>2-2-07</u>	<u>CPA</u>	<u>C. B. ...</u>
3. <u>10-3-09</u>	<u>ACTE</u>	<u>Ch. Stanley</u>
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

SP-2938 (Rev. 4/00)



JOHN S. CORZINE  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
PO BOX 080  
TRENTON, NJ 08625-0080

ANNE MILGRAM  
Attorney General

January 14, 2008

Col. Joseph R. Fuentes, Superintendent  
Division of State Police  
Division Headquarters  
P.O. Box 7068  
West Trenton, New Jersey 08628

Re: Breath Test Coordinator/Instructor, Certification -  
Trooper Joseph S. Dellanoce #6027

Dear Colonel Fuentes:

Pursuant to the provisions of *N.J.A.C.* 13:51-2.1 (b) and (c), as adopted and promulgated under the provisions of *N.J.S.A.* 39:4-50.3, 39:3-10.25 and 12:7-56, I hereby approve Trooper Joseph S. Dellanoce #6027, as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,

  
Anne Milgram  
Attorney General

jk  
c. Trooper Joseph S. Dellanoce #6027, Alcohol/Drug Test Unit, Division of State Police  
Lt. Mark Kolodzieski, Unit Head, Alcohol/Drug Test Unit, Division of State Police



20074



# CERTIFICATE

This is to certify that


**Joseph S. Dellanoce #6027**

has successfully completed the two day Draeger Safety  
Diagnostics, Inc. Operator Training and Preventive  
Maintenance Course on the New Jersey specific  
Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and  
Maintenance Technician**

Completion of this course qualifies this individual to train  
and certify Operators in the proper use and operation as  
well as perform Preventive Maintenance on the New Jersey  
specific Alcotest® 7110 MKIII-C.

Date: 4/12/2007

  
\_\_\_\_\_  
Instructor: Hansueli Rysler

**Draeger**



# Dräger

## ALCOTEST® 7110 TEMPERATURE PROBE

### CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDWAP2-016

February 7, 2011

February 7, 2012

Probe Value

107

Dräger Safety Diagnostics, Inc.  
Technical Service Department





State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7058  
 WEST TRENTON NJ 08628-0068  
 (609) 882-2000

PAULA T. DOW  
 Acting Attorney General

COLONEL JOSEPH R. FUENTES  
 Superintendent

HRIS CHRISTIE  
 Governor

**CERTIFICATION OF ANALYSIS**  
**0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.


ANALYSIS DATE: 2/2/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 19A073


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0479 to 0.0481 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 12, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
 Kenneth W. Kowalek, M.S.  
 Assistant Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

  
 Notary

Linda L. Desantis  
 Notary Public, New Jersey  
 Commission Expires 8-17-14





State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 Post Office Box 7068  
 WEST TRENTON NJ 08628-0068  
 (609) 882-2000

PAULA T. DOW  
 Acting Attorney General

COLONEL JOSEPH R. FUENTES  
 Superintendent

HRIS CHRISTIE  
 Governor

**CERTIFICATION OF ANALYSIS**  
**0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 2/3/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A074

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0954 to 0.0958 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

*Kenneth W. Kawalek*  
 Kenneth W. Kawalek, M.S.  
 Assistant Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

*Linda L. DiStasio*  
 Notary.

Linda L. DiStasio  
 Notary Public, New Jersey  
 My Commission Expires 8-17-14



New Jersey Is An Equal Opportunity Employer





State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON NJ 08628-0068  
 (609) 882-2000

CIRIS CHRISTIE  
 Governor

PAULA T. DOW  
 Attorney General

KIM GUADAGNO  
 Lt. Governor

COLONEL JOSEPH R. FUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 7/8/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10F080

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1197 to 0.1207 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 14, 2012.


As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

*Ajit R. Tungare*

Ajit R. Tungare  
 Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 5<sup>th</sup> day of July, 2010.

*Dora R. Fuentes*  
 Notary



New Jersey Is An Equal Opportunity Employer  
 Printed on Recycled Paper and Recyclable







State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 Post Office Box 7068  
 West Trenton NJ 08628-0068  
 (609) 882-2000

PAULA T. DOW  
 Acting Attorney General

COLONEL JOSEPH R. FUENTES  
 Superintendent

CHRIS CHRISTIE  
 Governor

**CERTIFICATION OF ANALYSIS**  
**0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.


ANALYSIS DATE: 2/4/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A075


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1913 to 0.1919 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 21, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
 Kenneth W. Kawalick, M.S.  
 Assistant Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

  
 Notary

Linda L. Desantis  
 Notary Public, New Jersey  
 My Commission Expires 8-17-14



New Jersey Is An Equal Opportunity Employer





**Dräger**

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Draeger Safety Diagnostics, Inc.**

Model: ALCOTEST® CU34

Model: MARK IIA

Other: \_\_\_\_\_

Serial Number:

DDWE 53-0196

Certification Date

Technician

Re-Certification Due Date

2/28/11

JM

2/28/12



**Dräger**

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Dräger Safety Diagnostics, Inc.**

Model: ALCOTEST® CU34

Model: MARK II A

Other: \_\_\_\_\_

Serial Number:

DDWES3-0203

Certification Date

Technician

Re-Certification Due Date

February 7, 2011

DM

February 7, 2012



**Dräger**

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Draeger Safety Diagnostics, Inc.**

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: \_\_\_\_\_

Serial Number:

DDWES3-0205

Certification Date

Technician

Re-Certification Due Date

February 7, 2011

DM

February 7, 2012

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	Calib. No.:	00017
Calibration File No.:	00458	Cert. Date:	03/28/2011
Certification File No.:	00459	Lin. Date:	03/28/2011
Linearity File No.:	00460	Soln. Date:	03/28/2011
Solution File No.:	00461	File Date:	03/28/2011
Sequential File No.:	00461		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0222
Solution Control Lot:	10F079	Expires:	06/10/2012
		Bottle No.:	1093

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:25S	15:25D	
Control 1 EC	0.099%	14:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:26S		
Control 2 EC	0.099%	14:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27S		
Control 3 EC	0.099%	14:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:29S	15:29D	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWJP2-293,10

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature:                     *J. Dellanoce*                    

Badge No.: 6027

Date: 03/28/2011



State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 Post Office Box 7068  
 WEST TRENTON NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
 Governor

PAULA T. DOW  
 Attorney General

KIM GUADAGNO  
 Lt. Governor

COLONEL JOSEPH R. PUENTI  
 Superintendent

**CERTIFICATION OF ANALYSIS  
 0.10 PERCENT BREATHE ALCOHOL SIMULATOR SOLUTION**

**ACCEPTANCE SPECIFICATIONS FOR BREATHE ALCOHOL SIMULATOR SOLUTION:** Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

**MANUFACTURER:** Dräger Safety, Inc.


**ANALYSIS DATE:** 7/7/2010

**BREATHE ALCOHOL SIMULATOR SOLUTION LOT NUMBER:** 10FV79


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1183 to 0.1190 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 10, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
 Kenneth W. Kawalek, M.S.  
 Assistant Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 20<sup>th</sup> day of July, 2010.

  
 Notary

Linda L. Mufantus  
 Notary Public, New Jersey  
 My Commission Expires 8-17-14



New Jersey Is An Equal Opportunity Employer  
 Print on Recycled Paper and Recyclable

