

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	
Calibration File No.:	01068	Calib. Date: 12/19/2019
Certification File No.:	01069	Calib. No.: 00040
Linearity File No.:	01070	Cert. Date: 12/19/2019
Solution File No.:	01073	Cert. No.: 00032
Sequential File No.:	01073	Lin. Date: 12/19/2019
		Lin. No.: 00032
		Soln. Date: 01/18/2020
		Soln. No.: 00241
		File Date: 01/18/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUM S3-0222
Control Solution %:	0.100%			Expires:	09/09/2021
Solution Control Lot:	19240			Bottle No.:	0164

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:03S		
Control 1 EC	0.100%	03:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:04S		
Control 2 EC	0.099%	03:05S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:06S		
Control 3 EC	0.098%	03:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: YOUNG

First Name: BRIAN

MI: R

Signature: _____



Badge No.: 4034

Date: 01/18/2020