

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	01143	Calib. Date:	09/08/2021
Certification File No.:	01144	Cert. Date:	09/08/2021
Linearity File No.:	01145	Lin. Date:	09/08/2021
Solution File No.:	01155	Soln. Date:	12/17/2021
Sequential File No.:	01155	File Date:	12/17/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0222
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	0398

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:01S		
Control 1 EC	0.100%	05:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	05:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:02S		
Control 2 EC	0.100%	05:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:03S		
Control 3 EC	0.099%	05:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:05S		

All tests within acceptable tolerance.

On this date, I installed the above-indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: _____



Badge No.: 4047

Date: 12/17/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01143 Calib. Date: 09/08/2021 Calib. No.: 00046
Certification File No.: 01144 Cert. Date: 09/08/2021 Cert. No.: 00038
Linearity File No.: 01145 Lin. Date: 09/08/2021 Lin. No.: 00037
Solution File No.: 01151 Soln. Date: 11/18/2021 Soln. No.: 00267
Sequential File No.: 01151 File Date: 11/18/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0222
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 0787

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:28S		
Control 1 EC	0.100%	13:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29S		
Control 2 EC	0.099%	13:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:30S		
Control 3 EC	0.100%	13:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: YOUNG

First Name: BRIAN

MI: R

Signature: _____



Badge No.: 4034

Date: 11/18/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	
Calibration File No.:	01143	Calib. Date: 09/08/2021
Certification File No.:	01144	Cert. No.: 00046
Linearity File No.:	01145	Cert. Date: 09/08/2021
Solution File No.:	01150	Lin. Date: 09/08/2021
Sequential File No.:	01150	Lin. No.: 00037
		Soln. Date: 10/18/2021
		Soln. No.: 00266
		File Date: 10/18/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUM S3-0222
Solution Control Lot:	20320	Expires: 07/06/2022
		Bottle No.: 0044

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:10D		
Control 1 EC	0.100%	12:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.099%	12:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		
Control 3 EC	0.100%	12:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: 

Badge No.: 4047

Date: 10/18/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01125 Calib. Date: 03/22/2021 Calib. No.: 00045
Certification File No.: 01126 Cert. Date: 03/22/2021 Cert. No.: 00037
Linearity File No.: 01127 Lin. Date: 03/22/2021 Lin. No.: 00036
Solution File No.: 01142 Soln. Date: 08/17/2021 Soln. No.: 00264
Sequential File No.: 01142 File Date: 08/17/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 0400

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:31D		
Control 1 EC	0.098%	16:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:32D		
Control 2 EC	0.097%	16:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:33D		
Control 3 EC	0.098%	16:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:35D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SADAR

First Name: TIMOTHY

MI: J

Signature: 

Badge No.: 4035

Date: 08/17/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01125 Calib. Date: 03/22/2021 Calib. No.: 00045
Certification File No.: 01126 Cert. Date: 03/22/2021 Cert. No.: 00037
Linearity File No.: 01127 Lin. Date: 03/22/2021 Lin. No.: 00036
Solution File No.: 01139 Soln. Date: 07/14/2021 Soln. No.: 00263
Sequential File No.: 01139 File Date: 07/14/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 0796

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:38D		
Control 1 EC	0.099%	06:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:39D		
Control 2 EC	0.098%	06:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:40D		
Control 3 EC	0.099%	06:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: A/Sgt. Anthony Cardone

Badge No.: 4047

Date: 07/14/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01125 Calib. Date: 03/22/2021 Calib. No.: 00045
Certification File No.: 01126 Cert. Date: 03/22/2021 Cert. No.: 00037
Linearity File No.: 01127 Lin. Date: 03/22/2021 Lin. No.: 00036
Solution File No.: 01137 Soln. Date: 06/14/2021 Soln. No.: 00262
Sequential File No.: 01137 File Date: 06/14/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 0341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:07D		
Control 1 EC	0.100%	02:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09D		
Control 2 EC	0.098%	02:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:10D		
Control 3 EC	0.099%	02:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: YOUNG

First Name: BRIAN

MI: R

Signature: _____



Badge No.: 4034

Date: 06/14/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01125 Calib. Date: 03/22/2021 Calib. No.: 00045
Certification File No.: 01126 Cert. Date: 03/22/2021 Cert. No.: 00037
Linearity File No.: 01127 Lin. Date: 03/22/2021 Lin. No.: 00036
Solution File No.: 01135 Soln. Date: 05/14/2021 Soln. No.: 00261
Sequential File No.: 01135 File Date: 05/14/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 0569

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:18D		
Control 1 EC	0.099%	03:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:19D		
Control 2 EC	0.098%	03:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:20D		
Control 3 EC	0.099%	03:21D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: _____



Badge No.: 4047

Date: 05/14/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01125 Calib. Date: 03/22/2021 Calib. No.: 00045
Certification File No.: 01126 Cert. Date: 03/22/2021 Cert. No.: 00037
Linearity File No.: 01127 Lin. Date: 03/22/2021 Lin. No.: 00036
Solution File No.: 01131 Soln. Date: 04/17/2021 Soln. No.: 00260
Sequential File No.: 01131 File Date: 04/17/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 0773

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:25D		
Control 1 EC	0.099%	20:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27D		
Control 2 EC	0.099%	20:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:28D		
Control 3 EC	0.099%	20:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCULLY

First Name: DANIEL

MI: S

Signature: _____



Badge No.: 4065

Date: 04/17/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	01102	Calib. Date:	09/30/2020
Certification File No.:	01103	Cert. Date:	09/30/2020
Linearity File No.:	01104	Lin. Date:	09/30/2020
Solution File No.:	01120	Soln. Date:	03/06/2021
Sequential File No.:	01120	File Date:	03/06/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0225
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	0567

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:51S		
Control 1 EC	0.100%	22:51S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:51S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:52S		
Control 2 EC	0.099%	22:52S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:52S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:53S		
Control 3 EC	0.100%	22:54S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:54S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCULLY

First Name: DANIEL

MI: S

Signature: _____



Badge No.: 4065

Date: 03/06/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.		
Calibration File No.:	01102	Calib. Date:	09/30/2020
Certification File No.:	01103	Cert. Date:	09/30/2020
Linearity File No.:	01104	Lin. Date:	09/30/2020
Solution File No.:	01116	Soln. Date:	02/05/2021
Sequential File No.:	01116	File Date:	02/05/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0225
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	0043

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:00S		
Control 1 EC	0.101%	13:01S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:01S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02S		
Control 2 EC	0.100%	13:02S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:02S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03S		
Control 3 EC	0.100%	13:03S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:03S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: JACOBY

First Name: AL

MI: J

Signature: _____



Badge No.: 4002

Date: 02/05/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0053
Location: FLORENCE TOWNSHIP P.D.
Calibration File No.: 01102 Calib. Date: 09/30/2020 Calib. No.: 00044
Certification File No.: 01103 Cert. Date: 09/30/2020 Cert. No.: 00035
Linearity File No.: 01104 Lin. Date: 09/30/2020 Lin. No.: 00035
Solution File No.: 01114 Soln. Date: 01/12/2021 Soln. No.: 00256
Sequential File No.: 01114 File Date: 01/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 0509

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:54S		
Control 1 EC	0.101%	09:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55S		
Control 2 EC	0.100%	09:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57S		
Control 3 EC	0.101%	09:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: _____



Badge No.: 4047

Date: 01/12/2021

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0053
Location:	FLORENCE TOWNSHIP P.D.	
Calibration File No.:	01102	Calib. Date: 09/30/2020
Certification File No.:	01103	Calib. No.: 00044
Linearity File No.:	01104	Cert. Date: 09/30/2020
Solution File No.:	01113	Cert. No.: 00035
Sequential File No.:	01113	Lin. Date: 09/30/2020
		Lin. No.: 00035
		Soln. Date: 01/12/2021
		Soln. No.: 00255
		File Date: 01/12/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUM S3-0225
Solution Control Lot:	20320	Expires: 07/06/2022
		Bottle No.: 0448

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:25S		
Control 1 EC	0.093%	08:26S	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.094%	08:26S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	08:31S		
Control 2 EC	0.095%	08:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	08:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		
Control 3 EC	0.097%	08:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	08:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:33S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: ASgh Atz ch

Badge No.: 4047

Date: 01/12/2021