

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
Location: FLORENCE TOWNSHIP P.D.  
Calibration File No.: 01179 Calib. Date: 07/06/2022 Calib. No.: 00048  
Certification File No.: 01180 Cert. Date: 07/06/2022 Cert. No.: 00040  
Linearity File No.: 01181 Lin. Date: 07/06/2022 Lin. No.: 00039  
Solution File No.: 01205 Soln. Date: 11/10/2022 Soln. No.: 00282  
Sequential File No.: 01205 File Date: 11/10/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 0239

| Function          | Result | Time   | Temperature    | Comment(s)      |
|-------------------|--------|--------|----------------|-----------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)     |
| Ambient Air Blank | 0.000% | 15:31S |                |                 |
| Control 1 EC      | -.---% | 15:31S | 34.0°C         | CTRL GAS SUPPLY |
| Control 1 IR      | -.---% | 15:31S | 34.0°C         | CTRL GAS SUPPLY |
| Ambient Air Blank | -.---% | --:--  |                |                 |
| Control 2 EC      | -.---% | --:--  | N/A            |                 |
| Control 2 IR      | -.---% | --:--  | N/A            |                 |
| Ambient Air Blank | -.---% | --:--  |                |                 |
| Control 3 EC      | -.---% | --:--  | N/A            |                 |
| Control 3 IR      | -.---% | --:--  | N/A            |                 |
| Ambient Air Blank | -.---% | --:--  |                |                 |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CZEPIEL

First Name: NICHOLAS

MI: B

Signature: \_\_\_\_\_

Badge No.: 4002

Date: 11/10/2022

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |              |
|-------------------------|------------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARTN-0053    |
| Location:               | FLORENCE TOWNSHIP P.D. |              |              |
| Calibration File No.:   | 01179                  | Calib. Date: | 07/06/2022   |
| Certification File No.: | 01180                  | Cert. Date:  | 07/06/2022   |
| Linearity File No.:     | 01181                  | Lin. Date:   | 07/06/2022   |
| Solution File No.:      | 01204                  | Soln. Date:  | 11/03/2022   |
| Sequential File No.:    | 01204                  | File Date:   | 11/03/2022   |
|                         |                        |              |              |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDUM S3-0225 |
| Solution Control Lot:   | 22160                  | Expires:     | 05/09/2024   |
|                         |                        | Bottle No.:  | 0416         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s)          |
|-------------------|----------------|---------------|-------------------------------|------------------------------------|
| Ambient Air Blank | 0.000%         | 05:10S        |                               |                                    |
| Control 1 EC      | -.---%         | 05:10S        | 34.0°C                        | CTRL GAS SUPPLY<br>CTRL GAS SUPPLY |
| Control 1 IR      | -.---%         | 05:10S        | 34.0°C                        |                                    |
| Ambient Air Blank | -.---%         | --:--         |                               |                                    |
| Control 2 EC      | -.---%         | --:--         | N/A                           |                                    |
| Control 2 IR      | -.---%         | --:--         | N/A                           |                                    |
| Ambient Air Blank | -.---%         | --:--         |                               |                                    |
| Control 3 EC      | -.---%         | --:--         | N/A                           |                                    |
| Control 3 IR      | -.---%         | --:--         | N/A                           |                                    |
| Ambient Air Blank | -.---%         | --:--         |                               |                                    |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LLOYD

First Name: MATTHEW

MI: T

Signature: \_\_\_\_\_

Badge No.: 4059

Date: 11/03/2022

# Calibrating Unit

## New Standard Solution Report

|                         |                        |                          |
|-------------------------|------------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARTN-0053    |
| Location:               | FLORENCE TOWNSHIP P.D. |                          |
| Calibration File No.:   | 01179                  | Calib. Date: 07/06/2022  |
| Certification File No.: | 01180                  | Calib. No.: 00048        |
| Linearity File No.:     | 01181                  | Cert. Date: 07/06/2022   |
| Solution File No.:      | 01203                  | Cert. No.: 00040         |
| Sequential File No.:    | 01203                  | Lin. Date: 07/06/2022    |
|                         |                        | Lin. No.: 00039          |
|                         |                        | Soln. Date: 10/31/2022   |
|                         |                        | Soln. No.: 00280         |
|                         |                        | File Date: 10/31/2022    |
| <br>                    |                        |                          |
| Calibrating Unit:       | WET                    | Model No.: CU-34         |
| Control Solution %:     | 0.100%                 | Serial No.: DDUM S3-0225 |
| Solution Control Lot:   | 22160                  | Expires: 05/09/2024      |
|                         |                        | Bottle No.: 0418         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 07:12S        |                               |                           |
| Control 1 EC      | -.---%         | 07:12S        | 34.0°C                        | CTRL GAS SUPPLY           |
| Control 1 IR      | -.---%         | 07:12S        | 34.0°C                        | CTRL GAS SUPPLY           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |
| Control 2 EC      | -.---%         | --:--         | N/A                           |                           |
| Control 2 IR      | -.---%         | --:--         | N/A                           |                           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |
| Control 3 EC      | -.---%         | --:--         | N/A                           |                           |
| Control 3 IR      | -.---%         | --:--         | N/A                           |                           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: \_\_\_\_\_



Badge No.: 4047

Date: 10/31/2022

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
Location: FLORENCE TOWNSHIP P.D.  
Calibration File No.: 01179 Calib. Date: 07/06/2022 Calib. No.: 00048  
Certification File No.: 01180 Cert. Date: 07/06/2022 Cert. No.: 00040  
Linearity File No.: 01181 Lin. Date: 07/06/2022 Lin. No.: 00039  
Solution File No.: 01196 Soln. Date: 10/02/2022 Soln. No.: 00278  
Sequential File No.: 01196 File Date: 10/02/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225  
Control Solution %: 0.100% Expires: 04/04/2024  
Solution Control Lot: 22110 Bottle No.: 0657

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 18:41D |                |                     |
| Control 1 EC      | 0.103% | 18:42D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.103% | 18:42D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:42D |                |                     |
| Control 2 EC      | 0.100% | 18:43D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.102% | 18:43D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:44D |                |                     |
| Control 3 EC      | 0.101% | 18:44D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.102% | 18:44D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:45D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SILK

First Name: JASON

MI: J

Signature: \_\_\_\_\_



Badge No.: 4049

Date: 10/02/2022

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
Location: FLORENCE TOWNSHIP P.D.  
Calibration File No.: 01179 Calib. Date: 07/06/2022 Calib. No.: 00048  
Certification File No.: 01180 Cert. Date: 07/06/2022 Cert. No.: 00040  
Linearity File No.: 01181 Lin. Date: 07/06/2022 Lin. No.: 00039  
Solution File No.: 01190 Soln. Date: 09/04/2022 Soln. No.: 00277  
Sequential File No.: 01190 File Date: 09/04/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 0199

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 04:06D |                |                     |
| Control 1 EC      | 0.099% | 04:07D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.101% | 04:07D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:07D |                |                     |
| Control 2 EC      | 0.100% | 04:08D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.101% | 04:08D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:09D |                |                     |
| Control 3 EC      | 0.099% | 04:09D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 04:09D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:10D |                |                     |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BONILLA

First Name: NICOLE

MI: M

Signature: 

Badge No.: 4043

Date: 09/04/2022

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
Location: FLORENCE TOWNSHIP P.D.  
Calibration File No.: 01179 Calib. Date: 07/06/2022 Calib. No.: 00048  
Certification File No.: 01180 Cert. Date: 07/06/2022 Cert. No.: 00040  
Linearity File No.: 01181 Lin. Date: 07/06/2022 Lin. No.: 00039  
Solution File No.: 01185 Soln. Date: 08/04/2022 Soln. No.: 00276  
Sequential File No.: 01185 File Date: 08/04/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0225  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 0598

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 06:58D |                |                     |
| Control 1 EC      | 0.101% | 06:58D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.101% | 06:58D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:59D |                |                     |
| Control 2 EC      | 0.099% | 06:59D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 06:59D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:00D |                |                     |
| Control 3 EC      | 0.100% | 07:01D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 07:01D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:01D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: SADAR

First Name: TIMOMTHY

MI: J

Signature: \_\_\_\_\_



Badge No.: 4035

Date: 08/04/2022

# Calibrating Unit

## New Standard Solution Report

|                         |                        |                         |
|-------------------------|------------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARTN-0053   |
| Location:               | FLORENCE TOWNSHIP P.D. |                         |
| Calibration File No.:   | 01161                  | Calib. Date: 02/08/2022 |
| Certification File No.: | 01162                  | Calib. No.: 00047       |
| Linearity File No.:     | 01163                  | Cert. Date: 02/08/2022  |
| Solution File No.:      | 01174                  | Cert. No.: 00039        |
| Sequential File No.:    | 01174                  | Lin. Date: 02/08/2022   |
|                         |                        | Lin. No.: 00038         |
|                         |                        | Soln. Date: 06/02/2022  |
|                         |                        | Soln. No.: 00274        |
|                         |                        | File Date: 06/02/2022   |

|                             |                  |                          |
|-----------------------------|------------------|--------------------------|
| Calibrating Unit: WET       | Model No.: CU-34 | Serial No.: DDUM S3-0225 |
| Control Solution %: 0.100%  |                  | Expires: 02/17/2024      |
| Solution Control Lot: 22070 |                  | Bottle No.: 1212         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 07:32D        |                               |                           |
| Control 1 EC      | 0.103%         | 07:33D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.101%         | 07:33D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:34D        |                               |                           |
| Control 2 EC      | 0.101%         | 07:34D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.100%         | 07:34D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:35D        |                               |                           |
| Control 3 EC      | 0.101%         | 07:35D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.101%         | 07:35D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:36D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: \_\_\_\_\_



Badge No.: 4047

Date: 06/02/2022

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |              |
|-------------------------|------------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARTN-0053    |
| Location:               | FLORENCE TOWNSHIP P.D. | Calib. No.:  | 00047        |
| Calibration File No.:   | 01161                  | Calib. Date: | 02/08/2022   |
| Certification File No.: | 01162                  | Cert. Date:  | 02/08/2022   |
| Linearity File No.:     | 01163                  | Lin. Date:   | 02/08/2022   |
| Solution File No.:      | 01172                  | Soln. Date:  | 05/02/2022   |
| Sequential File No.:    | 01172                  | File Date:   | 05/02/2022   |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDUM S3-0225 |
| Solution Control Lot:   | 20320                  | Expires:     | 07/06/2022   |
|                         |                        | Bottle No.:  | 0337         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 01:16D |                |                     |
| Control 1 EC      | 0.100% | 01:17D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 01:17D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:17D |                |                     |
| Control 2 EC      | 0.099% | 01:18D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 01:18D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:19D |                |                     |
| Control 3 EC      | 0.098% | 01:19D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 01:19D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:20D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: \_\_\_\_\_



Badge No.: 4047

Date: 05/02/2022



# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |              |
|-------------------------|------------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARTN-0053    |
| Location:               | FLORENCE TOWNSHIP P.D. |              |              |
| Calibration File No.:   | 01161                  | Calib. Date: | 02/08/2022   |
| Certification File No.: | 01162                  | Cert. No.:   | 00047        |
| Linearity File No.:     | 01163                  | Cert. Date:  | 02/08/2022   |
| Solution File No.:      | 01169                  | Lin. Date:   | 02/08/2022   |
| Sequential File No.:    | 01169                  | Lin. No.:    | 00038        |
|                         |                        | Soln. Date:  | 04/02/2022   |
|                         |                        | Soln. No.:   | 00272        |
|                         |                        | File Date:   | 04/02/2022   |
| <br>                    |                        |              |              |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDUM S3-0225 |
| Solution Control Lot:   | 20320                  | Expires:     | 07/06/2022   |
|                         |                        | Bottle No.:  | 0741         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 02:04S |                |                     |
| Control 1 EC      | 0.100% | 02:05S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 02:05S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:06S |                |                     |
| Control 2 EC      | 0.098% | 02:06S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 02:06S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:07S |                |                     |
| Control 3 EC      | 0.098% | 02:08S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 02:08S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:08S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CARDONE

First Name: ANTHONY

MI: D

Signature: \_\_\_\_\_



Badge No.: 4047

Date: 04/02/2022

# Calibrating Unit

## New Standard Solution Report

|                         |                        |             |              |
|-------------------------|------------------------|-------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: | ARTN-0053    |
| Location:               | FLORENCE TOWNSHIP P.D. | Calib. No.: | 00047        |
| Calibration File No.:   | 01161                  | Cert. Date: | 02/08/2022   |
| Certification File No.: | 01162                  | Lin. Date:  | 02/08/2022   |
| Linearity File No.:     | 01163                  | Soln. Date: | 03/06/2022   |
| Solution File No.:      | 01166                  | File Date:  | 03/06/2022   |
| Sequential File No.:    | 01166                  |             |              |
| <br>                    |                        |             |              |
| Calibrating Unit:       | WET                    | Model No.:  | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.: | DDUM S3-0225 |
| Solution Control Lot:   | 20320                  | Expires:    | 07/06/2022   |
|                         |                        | Bottle No.: | 0575         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 06:08S |                |                     |
| Control 1 EC      | 0.098% | 06:09S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 06:09S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:10S |                |                     |
| Control 2 EC      | 0.097% | 06:10S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 06:10S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:11S |                |                     |
| Control 3 EC      | 0.099% | 06:11S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 06:11S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:12S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LLOYD

First Name: MATTHEW

MI: T

Signature: \_\_\_\_\_



Badge No.: 4059

Date: 03/06/2022

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0053  
Location: FLORENCE TOWNSHIP P.D.  
Calibration File No.: 01143 Calib. Date: 09/08/2021 Calib. No.: 00046  
Certification File No.: 01144 Cert. Date: 09/08/2021 Cert. No.: 00038  
Linearity File No.: 01145 Lin. Date: 09/08/2021 Lin. No.: 00037  
Solution File No.: 01158 Soln. Date: 01/16/2022 Soln. No.: 00269  
Sequential File No.: 01158 File Date: 01/16/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0222  
Control Solution %: 0.100% Expires: 07/06/2022  
Solution Control Lot: 20320 Bottle No.: 0330

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 18:31S        |                               |                           |
| Control 1 EC      | 0.101%         | 18:32S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 18:32S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 18:33S        |                               |                           |
| Control 2 EC      | 0.099%         | 18:33S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.101%         | 18:33S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 18:34S        |                               |                           |
| Control 3 EC      | 0.100%         | 18:35S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 18:35S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 18:35S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SILK

First Name: JASON

MI: J

Signature: \_\_\_\_\_



Badge No.: 4049

Date: 01/16/2022