

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWA-0186  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01370 Calib. Date: 11/12/2014 Calib. No.: 00022  
Certification File No.: 01371 Cert. Date: 11/12/2014 Cert. No.: 00018  
Linearity File No.: 01372 Lin. Date: 11/12/2014 Lin. No.: 00018  
Solution File No.: 01382 Soln. Date: 12/19/2014 Soln. No.: 00162  
Sequential File No.: 01382 File Date: 12/19/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 131119 Bottle No.: 1290

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:57S		
Control 1 EC	0.100%	08:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58S		
Control 2 EC	0.100%	08:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 3 EC	0.099%	09:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CRESCITELLI

First Name: KENNETH

MI: E

Signature: \_\_\_\_\_



Badge No.: 1120

Date: 12/19/2014

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWA-0186  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01342 Calib. Date: 05/27/2014 Calib. No.: 00021  
Certification File No.: 01343 Cert. Date: 05/27/2014 Cert. No.: 00017  
Linearity File No.: 01344 Lin. Date: 05/27/2014 Lin. No.: 00017  
Solution File No.: 01367 Soln. Date: 11/04/2014 Soln. No.: 00160  
Sequential File No.: 01367 File Date: 11/04/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 131119 Bottle No.: 1281

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:11S		
Control 1 EC	0.099%	12:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12S		
Control 2 EC	0.099%	12:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13S		
Control 3 EC	0.099%	12:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: DIPIETRO

First Name: VINCENT

MI: J

Signature: \_\_\_\_\_



Badge No.: 1125

Date: 11/04/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01342	Calib. Date: 05/27/2014
Certification File No.:	01343	Calib. No.: 00021
Linearity File No.:	01344	Cert. Date: 05/27/2014
Solution File No.:	01360	Cert. No.: 00017
Sequential File No.:	01360	Lin. Date: 05/27/2014
		Lin. No.: 00017
		Soln. Date: 09/27/2014
		Soln. No.: 00159
		File Date: 09/27/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 1280

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:50D		
Control 1 EC	0.100%	03:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:52D		
Control 2 EC	0.099%	03:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:53D		
Control 3 EC	0.099%	03:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: REILLY

First Name: JAMES

MI: W

Signature: \_\_\_\_\_



Badge No.: 1139

Date: 09/27/2014

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01342	Calib. Date: 05/27/2014
Certification File No.:	01343	Calib. No.: 00021
Linearity File No.:	01344	Cert. Date: 05/27/2014
Solution File No.:	01355	Cert. No.: 00017
Sequential File No.:	01355	Lin. Date: 05/27/2014
		Lin. No.: 00017
		Soln. Date: 08/24/2014
		Soln. No.: 00158
		File Date: 08/24/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 1277

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:16D		
Control 1 EC	0.100%	15:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18D		
Control 2 EC	0.099%	15:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 3 EC	0.099%	15:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TROVILLO

First Name: C.

MI: -

Signature: \_\_\_\_\_



Badge No.: 1141

Date: 08/24/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01342	Calib. Date: 05/27/2014
Certification File No.:	01343	Calib. No.: 00021
Linearity File No.:	01344	Cert. Date: 05/27/2014
Solution File No.:	01351	Cert. No.: 00017
Sequential File No.:	01351	Lin. Date: 05/27/2014
		Lin. No.: 00017
		Soln. Date: 07/13/2014
		Soln. No.: 00157
		File Date: 07/13/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0286
Control Solution %:	0.100%		Expires: 09/10/2015
Solution Control Lot:	13I119		Bottle No.: 1293

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:08D		
Control 1 EC	0.099%	06:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:09D		
Control 2 EC	0.099%	06:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10D		
Control 3 EC	0.099%	06:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TROVILLO

First Name: C.

MI: -

Signature: \_\_\_\_\_



Badge No.: 1141

Date: 07/13/2014

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWA-0186  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01306 Calib. Date: 12/04/2013 Calib. No.: 00020  
Certification File No.: 01307 Cert. Date: 12/04/2013 Cert. No.: 00016  
Linearity File No.: 01308 Lin. Date: 12/04/2013 Lin. No.: 00016  
Solution File No.: 01336 Soln. Date: 05/11/2014 Soln. No.: 00155  
Sequential File No.: 01336 File Date: 05/11/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 13I119 Bottle No.: 1291

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:25D		
Control 1 EC	0.097%	03:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:26D		
Control 2 EC	0.097%	03:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:28D		
Control 3 EC	0.097%	03:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: TROVILLO

First Name: C.

MI: -

Signature: \_\_\_\_\_

Badge No.: 1141

Date: 05/11/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01306	Calib. Date: 12/04/2013
Certification File No.:	01307	Cert. No.: 00020
Linearity File No.:	01308	Cert. Date: 12/04/2013
Solution File No.:	01330	Lin. No.: 00016
Sequential File No.:	01330	Lin. Date: 12/04/2013
		Soln. No.: 00154
		Soln. Date: 04/21/2014
		File Date: 04/21/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0289
Solution Control Lot:	.131119	Expires: 09/10/2015
		Bottle No.: 1280

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:50D		
Control 1 EC	0.098%	10:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52D		
Control 2 EC	0.098%	10:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		
Control 3 EC	0.098%	10:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CRESCITELLI

First Name: KENNETH

MI: E

Signature: 

Badge No.: 1120

Date: 04/21/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWA 0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01306	Calib. Date:	12/04/2013
Certification File No.:	01307	Cert. Date:	12/04/2013
Linearity File No.:	01308	Lin. Date:	12/04/2013
Solution File No.:	01324	Soln. Date:	03/22/2014
Sequential File No.:	01324	File Date:	03/22/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	13119	Expires:	09/10/2015
		Bottle No.:	1392

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator: (°C)	or Error(s)
Ambient Air Blank	0.000%	02:58S		
Control 1 EC	0.099%	02:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:59S		
Control 2 EC	0.099%	03:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:01S		
Control 3 EC	0.099%	03:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TROVILLO

First Name: C.

MI: -

Signature:



Badge No.: 1141

Date: 03/22/2014





**State of New Jersey**  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON, NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
*Governor*

KIM GUADAGNO  
*Lt. Governor*

JOHN J. HOFFMAN  
*Acting Attorney General*

COLONEL JOSEPH R. FUENTES  
*Superintendent*

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

**ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION:** Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

**MANUFACTURER:** Draeger Safety, Inc.

**ANALYSIS DATE:** 10/17/2013

**BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER:** 131119

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1204 to 0.1218 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 10, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouic, Ph.D.  
 Research Scientist  
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 21<sup>st</sup> day of October, 2013.

Mary E. Medaughen  
 Notary

MARY ELIZABETH MCLAUGHLIN  
 NOTARY PUBLIC OF NEW JERSEY  
 MY COMMISSION EXPIRES 12-24-2013



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer  
 Printed on Recycled Paper and Recyclable



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKJII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01306	Calib. Date:	12/04/2013
Certification File No.:	01307	Cert. Date:	12/04/2013
Linearity File No.:	01308	Lin. Date:	12/04/2013
Solution File No.:	01319	Soln. Date:	02/18/2014
Sequential File No.:	01319	File Date:	02/18/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	12D101	Expires:	04/25/2014
		Bottle No.:	0550

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:30S		
Control 1 EC	0.100%	10:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31S		
Control 2 EC	0.100%	10:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32S		
Control 3 EC	0.099%	10:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: CRESCITELLI

First Name: KENNETH

MI: E

Signature: \_\_\_\_\_



Badge No.: 1120

Date: 02/18/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01306	Calib. Date: 12/04/2013
Certification File No.:	01307	Calib. No.: 00020
Linearity File No.:	01308	Cert. Date: 12/04/2013
Solution File No.:	01315	Cert. No.: 00016
Sequential File No.:	01315	Lin. Date: 12/04/2013
		Lin. No.: 00016
		Soln. Date: 01/15/2014
		Soln. No.: 00151
		File Date: 01/15/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0289
Control Solution %:	0.100%		Expires: 04/25/2014
Solution Control Lot:	12D101		Bottle No.: 5350

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:12S		
Control 1 EC	0.099%	09:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13S		
Control 2 EC	0.099%	09:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15S		
Control 3 EC	0.099%	09:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: CRESCITELLI

First Name: KENNETH

MI: E

Signature: \_\_\_\_\_



Badge No.: 1120

Date: 01/15/2014



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

JEFFREY S. CHIESA
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc. ANALYSIS DATE: 5/8/2012

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 12D101

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1204 to 0.1216 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 25, 2014.

As Forensic Scientist III for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Signature of Monica L. Tramontin
Monica L. Tramontin
Forensic Scientist III
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 15th day of May, 2012.

Signature of Linda L. Desantis
Notary

Linda L. Desantis
Notary Public, New Jersey
My Commission Expires 8-17-14



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable

