

Calibrating Unit

New Standard Solution Report



Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01433	Calib. Date:	09/10/2015
Certification File No.:	01434	Cert. Date:	09/10/2015
Linearity File No.:	01435	Lin. Date:	09/10/2015
Solution File No.:	01461	Soln. Date:	12/04/2015
Sequential File No.:	01461	File Date:	12/04/2015
Calibrating Unit: WET		Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0286
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	0181

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:10S		
Control 1 EC	0.101%	01:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:12S		
Control 2 EC	0.100%	01:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13S		
Control 3 EC	0.100%	01:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS	First Name: OFC. A.	MI: -
Signature: <u></u>	Badge No.: 1151	Date: 12/04/2015
Signature: <u></u>	Badge No.: 1151	Date: 12/04/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01433 Calib. Date: 09/10/2015 Calib. No.: 00024
Certification File No.: 01434 Cert. Date: 09/10/2015 Cert. No.: 00020
Linearity File No.: 01435 Lin. Date: 09/10/2015 Lin. No.: 00020
Solution File No.: 01458 Soln. Date: 11/24/2015 Soln. No.: 00174
Sequential File No.: 01458 File Date: 11/24/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 0182


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:18S		
Control 1 EC	0.101%	05:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:20S		
Control 2 EC	0.100%	05:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:21S		
Control 3 EC	0.100%	05:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:22S		


All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS First Name: OFC.A. MI: -

Signature:  Badge No.: 1151
Date: 11/24/2015

Signature:  Badge No.: 1151
Date: 11/24/2015

Calibrating Unit

New Standard Solution Report



Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01433	Calib. Date: 09/10/2015
Certification File No.:	01434	Calib. No.: 00024
Linearity File No.:	01435	Cert. Date: 09/10/2015
Solution File No.:	01447	Cert. No.: 00020
Sequential File No.:	01447	Lin. Date: 09/10/2015
		Lin. No.: 00020
		Soln. Date: 10/18/2015
		Soln. No.: 00173
		File Date: 10/18/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	14F130	Expires: 06/20/2016
		Bottle No.: 0198

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:38D		
Control 1 EC	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 2 EC	0.100%	09:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		
Control 3 EC	0.099%	09:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS	First Name: OFC. A.	MI: -
Signature: <u></u>	Badge No.: 1151	Date: 10/18/2015
Signature: <u></u>	Badge No.: 1151	Date: 10/18/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01433 Calib. Date: 09/10/2015 Calib. No.: 00024
Certification File No.: 01434 Cert. Date: 09/10/2015 Cert. No.: 00020
Linearity File No.: 01435 Lin. Date: 09/10/2015 Lin. No.: 00020
Solution File No.: 01436 Soln. Date: 09/10/2015 Soln. No.: 00172
Sequential File No.: 01436 File Date: 09/10/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286
Control Solution %: 0.100% Expires: 06/20/2016
Solution Control Lot: 14F130 Bottle No.: 0193

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:57D		
Control 1 EC	0.100%	09:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		
Control 2 EC	0.100%	09:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		
Control 3 EC	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔWAPZ - 018 **(KA)**

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TPR. I K W. #6704

Badge No.: 6704

Date: 09/10/2015

Signature: TPR. I K W. #6704

Date: 09/10/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01403	Calib. Date:	04/22/2015
Certification File No.:	01404	Cert. Date:	04/22/2015
Linearity File No.:	01405	Lin. Date:	04/22/2015
Solution File No.:	01427	Soln. Date:	08/09/2015
Sequential File No.:	01427	File Date:	08/09/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	131119	Expires:	09/10/2015
		Bottle No.:	1290

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:02D		
Control 1 EC	0.101%	21:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	21:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		
Control 2 EC	0.100%	21:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05D		
Control 3 EC	0.100%	21:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS

First Name: OFC. A.

MI: -

Signature: 

Badge No.: 1151
Date: 08/09/2015

Signature: 

Badge No.: 1151
Date: 08/09/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01403	Calib. Date:	04/22/2015
Certification File No.:	01404	Cert. No.:	00023
Linearity File No.:	01405	Cert. Date:	04/22/2015
Solution File No.:	01420	Lin. No.:	00019
Sequential File No.:	01420	Lin. Date:	04/22/2015
		Soln. No.:	00170
		Soln. Date:	07/02/2015
		File Date:	07/02/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0 100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	1299

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:33D		
Control 1 EC	0.100%	08:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35D		
Control 2 EC	0.099%	08:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:36D		
Control 3 EC	0.100%	08:37D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CRESCITELLI

First Name: KENNETH

MI: E

Signature: _____



Badge No.: 1120

Date: 07/02/2015

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01403	Calib. Date: 04/22/2015
Certification File No.:	01404	Calib. No.: 00023
Linearity File No.:	01405	Cert. Date: 04/22/2015
Solution File No.:	01419	Cert. No.: 00019
Sequential File No.:	01419	Lin. Date: 04/22/2015
		Lin. No.: 00019
		Soln. Date: 07/01/2015
		Soln. No.: 00169
		File Date: 07/01/2015

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWH S3-0289
Control Solution %: 0.100%		Expires: 09/10/2015
Solution Control Lot: 131119		Bottle No.: 1295

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:02D		
Control 1 EC	-.---%	18:02D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	18:02D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CRESCITELLI

First Name: KENNETH

MI: E

Signature:  1120

Badge No.: 1120

Date: 07/01/2015

Calibrating Unit New Standard Solution Report



Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01403	Calib. Date:	04/22/2015
Certification File No.:	01404	Cert. Date:	04/22/2015
Linearity File No.:	01405	Lin. Date:	04/22/2015
Solution File No.:	01415	Soln. Date:	05/25/2015
Sequential File No.:	01415	File Date:	05/25/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	1283

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:05D		
Control 1 EC	0.101%	09:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06D		
Control 2 EC	0.101%	09:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:08D		
Control 3 EC	0.101%	09:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS	First Name: OFC.A.	MI: -
Signature: <u> #1151</u>	Badge No.: 1151	Date: 05/25/2015
Signature: <u> #1151</u>	Badge No.: 1151	Date: 05/25/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01370	Calib. Date:	11/12/2014
Certification File No.:	01371	Cert. Date:	11/12/2014
Linearity File No.:	01372	Lin. Date:	11/12/2014
Solution File No.:	01402	Soln. Date:	04/20/2015
Sequential File No.:	01402	File Date:	04/20/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWH S3-0289
Control Solution %:	0.100%			Expires:	09/10/2015
Solution Control Lot:	13I119			Bottle No.:	1294


Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:43D		
Control 1 EC	0.096%	05:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:45D		
Control 2 EC	0.096%	05:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:46D		
Control 3 EC	0.097%	05:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CASAMASSIMA First Name: STEPHEN MI: N

Signature:  Badge No.: 1153
Date: 04/20/2015

Signature:  Badge No.: 1153
Date: 04/20/2015



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

JOHN J. HOFFMAN
 Acting Attorney General

KIM GUADAÑO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.

ANALYSIS DATE: 10/17/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 131119

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1204 to 0.1218 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is September 10, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 21st day of October, 2013.

Mary E. Medaughen
 Notary

MARY ELIZABETH McLAUGHLIN
 NOTARY PUBLIC OF NEW JERSEY
 MY COMMISSION EXPIRES 12-24-2013



"An Internationally Accredited Agency"

New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable



THOMAS

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01370	Calib. Date: 11/12/2014
Certification File No.:	01371	Calib. No.: 00022
Linearity File No.:	01372	Cert. Date: 11/12/2014
Solution File No.:	01397	Cert. No.: 00018
Sequential File No.:	01397	Lin. Date: 11/12/2014
		Lin. No.: 00018
		Soln. Date: 03/26/2015
		Soln. No.: 00165
		File Date: 03/26/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0289
Control Solution %:	0.100%		Expires: 09/10/2015
Solution Control Lot:	131119		Bottle No.: 1298

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:55S		
Control 1 EC	0.097%	04:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:56S		
Control 2 EC	0.097%	04:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:58S		
Control 3 EC	0.098%	04:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated 'NEW SOLUTION' in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DECESARI

First Name: MATTHEW

MI: P

Signature: _____



Badge No.: 1144

Date: 03/26/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01370 Calib. Date: 11/12/2014 Calib. No.: 00022
Certification File No.: 01371 Cert. Date: 11/12/2014 Cert. No.: 00018
Linearity File No.: 01372 Lin. Date: 11/12/2014 Lin. No.: 00018
Solution File No.: 01391 Soln. Date: 02/21/2015 Soln. No.: 00164
Sequential File No.: 01391 File Date: 02/21/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
Control Solution %: 0.100% Expires: 09/10/2015
Solution Control Lot: 131119 Bottle No.: 1297

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:32S		
Control 1 EC	0.097%	05:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	05:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:33S		
Control 2 EC	0.097%	05:34S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	05:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:35S		
Control 3 EC	0.098%	05:35S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	05:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:36S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS


First Name: OFC.A.

MI: -

Signature: 

Badge No.: 1151

Date: 02/21/2015

Signature: 

Badge No.: 1151

Date: 02/21/2015

Calibrating Unit New Standard Solution Report

Equipment: Alcotest 7110 MKIII-C
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01370 Calib. Date: 11/12/2014 Calib. No.: 00022
Certification File No.: 01371 Cert. Date: 11/12/2014 Cert. No.: 00018
Linearity File No.: 01372 Lin. Date: 11/12/2014 Lin. No.: 00018
Solution File No.: 01386 Soln. Date: 01/29/2015 Soln. No.: 00163
Sequential File No.: 01386 File Date: 01/29/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
Control Solution %: 0.100% Expires: 09/10/2015
Solution Control Lot: J31119 Bottle No.: 1287

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:32S		
Control 1 EC	0.101%	22:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:33S		
Control 2 EC	0.100%	22:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35S		
Control 3 EC	0.100%	22:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS

First Name: OFC.A.

MI: -

Signature: _____

OFC.A. Myers #1151

Badge No.: 1151

Date: 01/29/2015

Signature: _____

OFC.A. Myers #1151

Badge No.: 1151

Date: 01/29/2015