

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01641	Calib. Date: 08/02/2017
Certification File No.:	01642	Calib. No.: 00029
Linearity File No.:	01643	Cert. Date: 08/02/2017
Solution File No.:	01700	Cert. No.: 00024
Sequential File No.:	01700	Lin. Date: 08/02/2017
		Lin. No.: 00024
		Soln. Date: 12/23/2017
		Soln. No.: 00211
		File Date: 12/23/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0289
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 0072

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:05S		
Control 1 EC	0.099%	00:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06S		
Control 2 EC	0.099%	00:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08S		
Control 3 EC	0.099%	00:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GALLO

First Name: DAVE

MI: M

Signature: _____



#1167

Badge No.: 1167

Date: 12/23/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01641 Calib. Date: 08/02/2017 Calib. No.: 00029
Certification File No.: 01642 Cert. Date: 08/02/2017 Cert. No.: 00024
Linearity File No.: 01643 Lin. Date: 08/02/2017 Lin. No.: 00024
Solution File No.: 01692 Soln. Date: 11/29/2017 Soln. No.: 00210
Sequential File No.: 01692 File Date: 11/29/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 0071

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:36S		
Control 1 EC	0.097%	09:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37S		
Control 2 EC	0.097%	09:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39S		
Control 3 EC	0.097%	09:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 11/29/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01641 Calib. Date: 08/02/2017 Calib. No.: 00029
Certification File No.: 01642 Cert. Date: 08/02/2017 Cert. No.: 00024
Linearity File No.: 01643 Lin. Date: 08/02/2017 Lin. No.: 00024
Solution File No.: 01687 Soln. Date: 11/18/2017 Soln. No.: 00209
Sequential File No.: 01687 File Date: 11/18/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
Control Solution %: 0.100% Expires: 12/02/2018
Solution Control Lot: 16390 Bottle No.: 1365

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:21S		
Control 1 EC	0.099%	20:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:22S		
Control 2 EC	0.097%	20:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:24S		
Control 3 EC	0.097%	20:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: _____

Cpl. A. Myers #1151

Badge No.: 1151

Date: 11/18/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01641 Calib. Date: 08/02/2017 Calib. No.: 00029
Certification File No.: 01642 Cert. Date: 08/02/2017 Cert. No.: 00024
Linearity File No.: 01643 Lin. Date: 08/02/2017 Lin. No.: 00024
Solution File No.: 01685 Soln. Date: 11/11/2017 Soln. No.: 00208
Sequential File No.: 01685 File Date: 11/11/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 0246

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:20S		
Control 1 EC	0.095%	04:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:21S		
Control 2 EC	0.095%	04:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23S		
Control 3 EC	0.095%	04:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	04:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____

 1143

Badge No.: 1143

Date: 11/11/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01641 Calib. Date: 08/02/2017 Calib. No.: 00029
Certification File No.: 01642 Cert. Date: 08/02/2017 Cert. No.: 00024
Linearity File No.: 01643 Lin. Date: 08/02/2017 Lin. No.: 00024
Solution File No.: 01683 Soln. Date: 11/07/2017 Soln. No.: 00207
Sequential File No.: 01683 File Date: 11/07/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 0067

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:57S		
Control 1 EC	0.097%	04:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:58S		
Control 2 EC	0.096%	04:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:00S		
Control 3 EC	0.096%	05:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 11/07/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01641	Calib. Date:	08/02/2017
Certification File No.:	01642	Cert. No.:	00029
Linearity File No.:	01643	Cert. Date:	08/02/2017
Solution File No.:	01675	Lin. No.:	00024
Sequential File No.:	01675	Soln. Date:	10/19/2017
		Soln. No.:	00206
		File Date:	10/19/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	16140	Expires:	06/07/2018
		Bottle No.:	0492

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:00D		
Control 1 EC	0.097%	02:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:02D		
Control 2 EC	0.097%	02:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:03D		
Control 3 EC	0.097%	02:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____

Badge No.: 1143

Date: 10/19/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01641	Calib. Date:	08/02/2017
Certification File No.:	01642	Cert. Date:	08/02/2017
Linearity File No.:	01643	Lin. Date:	08/02/2017
Solution File No.:	01674	Soln. Date:	10/19/2017
Sequential File No.:	01674	File Date:	10/19/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	16140	Expires:	06/07/2018
		Bottle No.:	0028

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:26D		
Control 1 EC	0.096%	00:27D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.093%	00:27D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	00:27D		
Control 2 EC	0.098%	00:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	00:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:29D		
Control 3 EC	0.096%	00:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	00:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____

Badge No.: 1143

Date: 10/19/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01641 Calib. Date: 08/02/2017 Calib. No.: 00029
Certification File No.: 01642 Cert. Date: 08/02/2017 Cert. No.: 00024
Linearity File No.: 01643 Lin. Date: 08/02/2017 Lin. No.: 00024
Solution File No.: 01666 Soln. Date: 09/29/2017 Soln. No.: 00204
Sequential File No.: 01666 File Date: 09/29/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
Control Solution %: 0.100% Expires: 12/02/2018
Solution Control Lot: 16390 Bottle No.: 1266

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:48D		
Control 1 EC	0.098%	08:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50D		
Control 2 EC	0.098%	08:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51D		
Control 3 EC	0.099%	08:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 09/29/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01641	Calib. Date: 08/02/2017
Certification File No.:	01642	Calib. No.: 00029
Linearity File No.:	01643	Cert. Date: 08/02/2017
Solution File No.:	01660	Cert. No.: 00024
Sequential File No.:	01660	Lin. Date: 08/02/2017
		Lin. No.: 00024
		Soln. Date: 08/29/2017
		Soln. No.: 00203
		File Date: 08/29/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0289
Control Solution %:	0.100%		Expires: 12/02/2018
Solution Control Lot:	16390		Bottle No.: 1353

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:29D		
Control 1 EC	0.099%	05:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:30D		
Control 2 EC	0.098%	05:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:32D		
Control 3 EC	0.098%	05:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GUZMAN

First Name: MICHAEL

MI: C

Signature: _____



1156

Badge No.: 1156

Date: 08/29/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	01586	Calib. Date: 02/14/2017
Certification File No.:	01587	Calib. No.: 00028
Linearity File No.:	01588	Cert. Date: 02/14/2017
Solution File No.:	01634	Cert. No.: 00023
Sequential File No.:	01634	Lin. Date: 02/14/2017
		Lin. No.: 00023
		Soln. Date: 07/22/2017
		Soln. No.: 00201
		File Date: 07/22/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0286
Control Solution %:	0.100%		Expires: 05/17/2018
Solution Control Lot:	201605D		Bottle No.: 0160

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:01D		
Control 1 EC	0.101%	21:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02D		
Control 2 EC	0.100%	21:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	21:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		
Control 3 EC	0.101%	21:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 07/22/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01586 Calib. Date: 02/14/2017 Calib. No.: 00028
Certification File No.: 01587 Cert. Date: 02/14/2017 Cert. No.: 00023
Linearity File No.: 01588 Lin. Date: 02/14/2017 Lin. No.: 00023
Solution File No.: 01624 Soln. Date: 06/20/2017 Soln. No.: 00200
Sequential File No.: 01624 File Date: 06/20/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 0194

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:57D		
Control 1 EC	0.102%	04:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:59D		
Control 2 EC	0.101%	04:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:00D		
Control 3 EC	0.101%	05:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 06/20/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01586 Calib. Date: 02/14/2017 Calib. No.: 00028
Certification File No.: 01587 Cert. Date: 02/14/2017 Cert. No.: 00023
Linearity File No.: 01588 Lin. Date: 02/14/2017 Lin. No.: 00023
Solution File No.: 01603 Soln. Date: 04/17/2017 Soln. No.: 00198
Sequential File No.: 01603 File Date: 04/17/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286
Control Solution %: 0.100% Expires: 05/17/2018
Solution Control Lot: 201605D Bottle No.: 0163

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:44D		
Control 1 EC	0.101%	15:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45D		
Control 2 EC	0.101%	15:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		
Control 3 EC	0.101%	15:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: _____

Cpl. A. Myers #1151

Badge No.: 1151

Date: 04/17/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWA-0186
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 01586 Calib. Date: 02/14/2017 Calib. No.: 00028
Certification File No.: 01587 Cert. Date: 02/14/2017 Cert. No.: 00023
Linearity File No.: 01588 Lin. Date: 02/14/2017 Lin. No.: 00023
Solution File No.: 01595 Soln. Date: 03/17/2017 Soln. No.: 00197
Sequential File No.: 01595 File Date: 03/17/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 0389

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:24S		
Control 1 EC	0.102%	19:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:25S		
Control 2 EC	0.102%	19:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	19:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:26S		
Control 3 EC	0.101%	19:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	19:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GUZMAN

First Name: MICHAEL

MI: C

Signature: 

Badge No.: 1156

Date: 03/17/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWA-0186
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01537	Calib. Date:	08/16/2016
Certification File No.:	01538	Cert. Date:	08/16/2016
Linearity File No.:	01539	Lin. Date:	08/16/2016
Solution File No.:	01584	Soln. Date:	02/13/2017
Sequential File No.:	01584	File Date:	02/13/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWH S3-0289
Control Solution %:	0.100%			Expires:	08/04/2017
Solution Control Lot:	15H140			Bottle No.:	0198

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:40S		
Control 1 EC	0.102%	09:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42S		
Control 2 EC	0.102%	09:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	09:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43S		
Control 3 EC	0.102%	09:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____

Badge No.: 1143

Date: 02/13/2017