

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00641	Calib. Date:	06/28/2018
Certification File No.:	00642	Cert. Date:	06/28/2018
Linearity File No.:	00643	Lin. Date:	06/28/2018
Solution File No.:	00680	Soln. Date:	10/28/2018
Sequential File No.:	00680	File Date:	10/28/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	1074

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:55S		
Control 1 EC	0.098%	20:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:56S		
Control 2 EC	0.097%	20:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57S		
Control 3 EC	0.098%	20:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROJAS

First Name: ANTHONY

MI: J

Signature: \_\_\_\_\_

*Sgt. A. Vega #1152*

Badge No.: 1152

Date: 10/28/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00641	Calib. Date: 06/28/2018
Certification File No.:	00642	Calib. No.: 00030
Linearity File No.:	00643	Cert. Date: 06/28/2018
Solution File No.:	00673	Cert. No.: 00025
Sequential File No.:	00673	Lin. Date: 06/28/2018
		Lin. No.: 00025
		Soln. Date: 09/28/2018
		Soln. No.: 00194
		File Date: 09/28/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0289
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 1088

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:52D		
Control 1 EC	0.098%	08:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		
Control 2 EC	0.097%	08:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		
Control 3 EC	0.097%	08:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature:  1143

Badge No.: 1143

Date: 09/28/2018

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00641 Calib. Date: 06/28/2018 Calib. No.: 00030  
Certification File No.: 00642 Cert. Date: 06/28/2018 Cert. No.: 00025  
Linearity File No.: 00643 Lin. Date: 06/28/2018 Lin. No.: 00025  
Solution File No.: 00660 Soln. Date: 08/28/2018 Soln. No.: 00193  
Sequential File No.: 00660 File Date: 08/28/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 1119

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:39D		
Control 1 EC	0.099%	05:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:40D		
Control 2 EC	0.100%	05:41D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:42D		
Control 3 EC	0.099%	05:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: LOCILENTO

First Name: NICHOLAS

MI: M

Signature: \_\_\_\_\_

Badge No.: 1158

Date: 08/28/2018

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00641	Calib. Date:	06/28/2018
Certification File No.:	00642	Cert. Date:	06/28/2018
Linearity File No.:	00643	Lin. Date:	06/28/2018
Solution File No.:	00651	Soln. Date:	07/28/2018
Sequential File No.:	00651	File Date:	07/28/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	1011

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:16D		
Control 1 EC	0.100%	15:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17D		
Control 2 EC	0.099%	15:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18D		
Control 3 EC	0.099%	15:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature:  1143

Badge No.: 1143

Date: 07/28/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00593	Calib. Date: 01/03/2018
Certification File No.:	00594	Calib. No.: 00029
Linearity File No.:	00595	Cert. Date: 01/03/2018
Solution File No.:	00639	Cert. No.: 00024
Sequential File No.:	00639	Lin. Date: 01/03/2018
		Lin. No.: 00024
		Soln. Date: 06/13/2018
		Soln. No.: 00190
		File Date: 06/13/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0986

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:20D		
Control 1 EC	0.101%	14:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21D		
Control 2 EC	0.100%	14:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:23D		
Control 3 EC	0.100%	14:23D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_



Badge No.: 1143

Date: 06/13/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00593	Calib. Date: 01/03/2018
Certification File No.:	00594	Calib. No.: 00029
Linearity File No.:	00595	Cert. Date: 01/03/2018
Solution File No.:	00634	Cert. No.: 00024
Sequential File No.:	00634	Lin. Date: 01/03/2018
		Lin. No.: 00024
		Soln. Date: 05/12/2018
		Soln. No.: 00189
		File Date: 05/12/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	16390	Expires: 12/02/2018
		Bottle No.: 1404

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:54D		
Control 1 EC	0.100%	21:55D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:55D		
Control 2 EC	0.099%	21:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:57D		
Control 3 EC	0.099%	21:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROJAS

First Name: ANTHONY

MI: J

Signature: \_\_\_\_\_

Sgt. A. Rojas #1152

Badge No.: 1152

Date: 05/12/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00593	Calib. Date:	01/03/2018
Certification File No.:	00594	Cert. No.:	00029
Linearity File No.:	00595	Cert. Date:	01/03/2018
Solution File No.:	00630	Lin. No.:	00024
Sequential File No.:	00630	Lin. Date:	01/03/2018
		Soln. Date:	04/11/2018
		Soln. No.:	00188
		File Date:	04/11/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0286
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1354

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:30D		
Control 1 EC	0.100%	00:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:32D		
Control 2 EC	0.098%	00:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33D		
Control 3 EC	0.098%	00:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: *Cpl. A. Myers #1151*

Badge No.: 1151

Date: 04/11/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00593	Calib. Date:	01/03/2018
Certification File No.:	00594	Cert. No.:	00029
Linearity File No.:	00595	Cert. Date:	01/03/2018
Solution File No.:	00625	Lin. No.:	00024
Sequential File No.:	00625	Soln. Date:	03/24/2018
		Soln. No.:	00187
		File Date:	03/24/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0286
Solution Control Lot:	16140	Expires:	06/07/2018
		Bottle No.:	0648

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:50S		
Control 1 EC	0.099%	00:50S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:51S		
Control 2 EC	0.097%	00:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:52S		
Control 3 EC	0.098%	00:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PASCULLI

First Name: NICHOLAS

MI: P

Signature: \_\_\_\_\_

*G.P. Nicholas Pasculli #1150*

Badge No.: 1150

Date: 03/24/2018



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00593 Calib. Date: 01/03/2018 Calib. No.: 00029  
Certification File No.: 00594 Cert. Date: 01/03/2018 Cert. No.: 00024  
Linearity File No.: 00595 Lin. Date: 01/03/2018 Lin. No.: 00024  
Solution File No.: 00613 Soln. Date: 02/27/2018 Soln. No.: 00186  
Sequential File No.: 00613 File Date: 02/27/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286  
Control Solution %: 0.100% Expires: 12/02/2018  
Solution Control Lot: 16390 Bottle No.: 1364

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:11S		
Control 1 EC	0.098%	03:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:12S		
Control 2 EC	0.097%	03:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:14S		
Control 3 EC	0.098%	03:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CASAMASSIMA

First Name: STEPHEN

MI: N

Signature: 

Badge No.: 1153

Date: 02/27/2018

#1153

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00593 Calib. Date: 01/03/2018 Calib. No.: 00029  
Certification File No.: 00594 Cert. Date: 01/03/2018 Cert. No.: 00024  
Linearity File No.: 00595 Lin. Date: 01/03/2018 Lin. No.: 00024  
Solution File No.: 00605 Soln. Date: 01/29/2018 Soln. No.: 00185  
Sequential File No.: 00605 File Date: 01/29/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286  
Control Solution %: 0.100% Expires: 12/02/2018  
Solution Control Lot: 16390 Bottle No.: 1362

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:34S		
Control 1 EC	0.099%	00:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:36S		
Control 2 EC	0.098%	00:36S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:37S		
Control 3 EC	0.099%	00:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PASCULLI

First Name: NICHOLAS

MI: P

Signature: Cpt. Nicholas Pasculli #1150

Badge No.: 1150

Date: 01/29/2018