

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00762	Calib. Date: 10/04/2019
Certification File No.:	00763	Calib. No.: 00033
Linearity File No.:	00764	Cert. Date: 10/04/2019
Solution File No.:	00773	Lin. Date: 10/04/2019
Sequential File No.:	00773	Lin. No.: 00028
		Soln. Date: 11/03/2019
		Soln. No.: 00210
		File Date: 11/03/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0286
Control Solution %:	0.100%		Expires: 01/08/2021
Solution Control Lot:	19020		Bottle No.: 0778

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:27S		
Control 1 EC	0.097%	17:28S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:28S		
Control 2 EC	0.097%	17:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:30S		
Control 3 EC	0.097%	17:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JERNEGAN

First Name: DAVID

MI: A

Signature: _____



Badge No.: 1148

Date: 11/03/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00731	Calib. Date:	04/26/2019
Certification File No.:	00732	Cert. Date:	04/26/2019
Linearity File No.:	00733	Lin. Date:	04/26/2019
Solution File No.:	00761	Soln. Date:	09/30/2019
Sequential File No.:	00761	File Date:	09/30/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0833

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:57D		
Control 1 EC	0.099%	22:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:58D		
Control 2 EC	0.099%	22:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:59D		
Control 3 EC	0.099%	23:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NEHER

First Name: MICHAEL

MI: E

Signature: _____

Off. Michael Neher #1155

Badge No.: 1155

Date: 09/30/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00731	Calib. Date: 04/26/2019
Certification File No.:	00732	Calib. No.: 00032
Linearity File No.:	00733	Cert. Date: 04/26/2019
Solution File No.:	00756	Cert. No.: 00027
Sequential File No.:	00756	Lin. Date: 04/26/2019
		Lin. No.: 00027
		Soln. Date: 08/31/2019
		Soln. No.: 00207
		File Date: 08/31/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0289
Control Solution %:	0.100%		Expires: 01/31/2020
Solution Control Lot:	18050		Bottle No.: 0950

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:32D		
Control 1 EC	0.100%	05:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:33D		
Control 2 EC	0.100%	05:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:35D		
Control 3 EC	0.099%	05:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PIPER

First Name: ERIC

MI: S

Signature: _____



Badge No.: 1138

Date: 08/31/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0071
 Location: FRANKLIN TOWNSHIP P.D.
 Calibration File No.: 00731 Calib. Date: 04/26/2019 Calib. No.: 00032
 Certification File No.: 00732 Cert. Date: 04/26/2019 Cert. No.: 00027
 Linearity File No.: 00733 Lin. Date: 04/26/2019 Lin. No.: 00027
 Solution File No.: 00754 Soln. Date: 07/30/2019 Soln. No.: 00206
 Sequential File No.: 00754 File Date: 07/30/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289
 Control Solution %: 0.100% Expires: 01/08/2021
 Solution Control Lot: 19020 Bottle No.: 0264

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:44D		
Control 1 EC	0.099%	08:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:45D		
Control 2 EC	0.099%	08:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46D		
Control 3 EC	0.099%	08:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____

Badge No.: 1143

Date: 07/30/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00731	Calib. Date:	04/26/2019
Certification File No.:	00732	Cert. No.:	00032
Linearity File No.:	00733	Cert. Date:	04/26/2019
Solution File No.:	00746	Lin. No.:	00027
Sequential File No.:	00746	Lin. Date:	04/26/2019
		Soln. Date:	06/28/2019
		Soln. No.:	00205
		File Date:	06/28/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0779

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:52D		
Control 1 EC	0.099%	06:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:53D		
Control 2 EC	0.098%	06:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:54D		
Control 3 EC	0.099%	06:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 06/28/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00731	Calib. Date:	04/26/2019
Certification File No.:	00732	Cert. No.:	00032
Linearity File No.:	00733	Lin. Date:	04/26/2019
Solution File No.:	00739	Soln. No.:	00027
Sequential File No.:	00739	Soln. Date:	05/29/2019
		File Date:	05/29/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0784

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:09D		
Control 1 EC	0.100%	01:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:10D		
Control 2 EC	0.099%	01:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:12D		
Control 3 EC	0.099%	01:12D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: _____

Cpl. A. Myers 1151

Badge No.: 1151

Date: 05/29/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00686	Calib. Date: 11/20/2018
Certification File No.:	00687	Cert. No.: 00031
Linearity File No.:	00688	Cert. Date: 11/20/2018
Solution File No.:	00725	Lin. Date: 11/20/2018
Sequential File No.:	00725	Lin. No.: 00026
		Soln. Date: 04/12/2019
		Soln. No.: 00202
		File Date: 04/12/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0972

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:32D		
Control 1 EC	0.098%	00:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:34D		
Control 2 EC	0.097%	00:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:35D		
Control 3 EC	0.099%	00:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JERNEGAN

First Name: DAVID

MI: A

Signature: _____

Badge No.: 1148

Date: 04/12/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00686	Calib. Date: 11/20/2018
Certification File No.:	00687	Calib. No.: 00031
Linearity File No.:	00688	Cert. Date: 11/20/2018
Solution File No.:	00713	Cert. No.: 00026
Sequential File No.:	00713	Lin. Date: 11/20/2018
		Lin. No.: 00026
		Soln. Date: 03/05/2019
		Soln. No.: 00200
		File Date: 03/05/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0286
Control Solution %:	0.100%		Expires: 01/31/2020
Solution Control Lot:	18050		Bottle No.: 1020

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:37S		
Control 1 EC	0.098%	22:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38S		
Control 2 EC	0.098%	22:39S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:40S		
Control 3 EC	0.099%	22:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: *Cpl. A. Myers #1151*

Badge No.: 1151

Date: 03/05/2019