

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00807	Calib. Date: 05/18/2020
Certification File No.:	00808	Calib. No.: 00035
Linearity File No.:	00809	Cert. Date: 05/18/2020
Solution File No.:	00829	Cert. No.: 00029
Sequential File No.:	00829	Lin. Date: 05/18/2020
		Lin. No.: 00029
		Soln. Date: 09/14/2020
		Soln. No.: 00223
		File Date: 09/14/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0289
Control Solution %:	0.100%		Expires: 06/10/2021
Solution Control Lot:	19150		Bottle No.: 0783

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:11D		
Control 1 EC	0.097%	07:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:13D		
Control 2 EC	0.097%	07:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:14D		
Control 3 EC	0.098%	07:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____

Badge No.: 1143

Date: 09/14/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00807	Calib. Date:	05/18/2020
Certification File No.:	00808	Cert. Date:	05/18/2020
Linearity File No.:	00809	Lin. Date:	05/18/2020
Solution File No.:	00817	Soln. Date:	07/13/2020
Sequential File No.:	00817	File Date:	07/13/2020
		Calib. No.:	00035
		Cert. No.:	00029
		Lin. No.:	00029
		Soln. No.:	00221

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWH S3-0289
Control Solution %:	0.100%			Expires:	06/10/2021
Solution Control Lot:	19150			Bottle No.:	0453

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:26D		
Control 1 EC	0.098%	21:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27D		
Control 2 EC	0.098%	21:27D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:28D		
Control 3 EC	0.099%	21:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 07/13/2020