

# Calibrating Unit New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
 Location: FRANKLIN TOWNSHIP P.D.  
 Calibration File No.: 00927      Calib. Date: 11/03/2022      Serial No.: ARWC-0071  
 Certification File No.: 00928      Cert. Date: 11/03/2022      Calib. No.: 00041  
 Linearity File No.: 00929      Lin. Date: 11/03/2022      Cert. No.: 00034  
 Solution File No.: 00940      Soln. Date: 12/18/2022      Lin. No.: 00034  
 Sequential File No.: 00940      File Date: 12/18/2022      Soln. No.: 00249

Calibrating Unit: WET      Model No.: CU-34      Serial No.: DDWH S3-0289  
 Control Solution %: 0.100%      Expires: 04/04/2024  
 Solution Control Lot: 22110      Bottle No.: 0023

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:18S		
Control 1 EC	0.098%	06:19S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:19S		
Control 2 EC	0.096%	06:20S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21S		
Control 3 EC	0.097%	06:22S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROJAS

First Name: ANTHONY

MI: ]

Signature: \_\_\_\_\_

*Sgt Anthony #1150*

Badge No.: 1152

Date: 12/18/2022

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00927	Calib. No.: 00041
Certification File No.:	00928	Cert. No.: 00034
Linearity File No.:	00929	Lin. No.: 00034
Solution File No.:	00935	Soln. No.: 00248
Sequential File No.:	00935	File Date: 12/02/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0289
Solution Control Lot:	22110	Expires: 04/04/2024
		Bottle No.: 0455

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:45S		
Control 1 EC	0.099%	12:46S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47S		
Control 2 EC	0.098%	12:47S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48S		
Control 3 EC	0.099%	12:49S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_

*Sgt J. McGuigan 1143*

Badge No.: 1143

Date: 12/02/2022

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00904	Calib. Date: 06/30/2022
Certification File No.:	00905	Calib. No.: 00040
Linearity File No.:	00906	Cert. Date: 06/30/2022
Solution File No.:	00922	Cert. No.: 00033
Sequential File No.:	00922	Lin. Date: 06/30/2022
		Lin. No.: 00033
		Soln. Date: 10/06/2022
		Soln. No.: 00246
		File Date: 10/06/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	22110	Expires: 04/04/2024
		Bottle No.: 0124

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:36D		
Control 1 EC	0.099%	13:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38D		
Control 2 EC	0.097%	13:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39D		
Control 3 EC	0.098%	13:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_

*Sgt Jmg 1143*

Badge No.: 1143

Date: 10/06/2022

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
 Location: FRANKLIN TOWNSHIP P.D.  
 Calibration File No.: 00904  
 Certification File No.: 00905  
 Linearity File No.: 00906  
 Solution File No.: 00915  
 Sequential File No.: 00915

Serial No.: ARWC-0071

Calib. Date: 06/30/2022  
 Cert. Date: 06/30/2022  
 Lin. Date: 06/30/2022  
 Soln. Date: 09/09/2022  
 File Date: 09/09/2022

Calib. No.: 00040  
 Cert. No.: 00033  
 Lin. No.: 00033  
 Soln. No.: 00245

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 22110

Model No.: CU-34

Serial No.: DDWH S3-0286  
 Expires: 04/04/2024  
 Bottle No.: 0021

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:50D		*** TEST PASSED ***
Control 1 EC	0.100%	15:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51D		*** TEST PASSED ***
Control 2 EC	0.099%	15:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:53D		*** TEST PASSED ***
Control 3 EC	0.099%	15:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Badge No.: 1143

Date: 09/09/2022

Signature: \_\_\_\_\_

*Sgt Jeffrey 1143*

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00904	Calib. Date: 06/30/2022
Certification File No.:	00905	Calib. No.: 00040
Linearity File No.:	00906	Cert. Date: 06/30/2022
Solution File No.:	00911	Cert. No.: 00033
Sequential File No.:	00911	Lin. Date: 06/30/2022
		Lin. No.: 00033
		Soln. Date: 08/08/2022
		Soln. No.: 00244
		File Date: 08/08/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	22110	Expires: 04/04/2024
		Bottle No.: 0019

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:12D		
Control 1 EC	0.101%	02:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	02:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:14D		
Control 2 EC	0.100%	02:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:15D		
Control 3 EC	0.100%	02:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MYERS

First Name: SGT. A.

MI: -

Signature: \_\_\_\_\_

*Dgt. A. Myers #1151*

Badge No.: 1151

Date: 08/08/2022

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00881	Calib. Date: 02/01/2022
Certification File No.:	00882	Calib. No.: 00039
Linearity File No.:	00883	Cert. Date: 02/01/2022
Solution File No.:	00898	Cert. No.: 00032
Sequential File No.:	00898	Lin. Date: 02/01/2022
		Lin. No.: 00032
		Soln. Date: 06/10/2022
		Soln. No.: 00242
		File Date: 06/10/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	22110	Expires: 04/04/2024
		Bottle No.: 0125

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:24D		
Control 1 EC	0.101%	14:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:26D		
Control 2 EC	0.100%	14:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27D		
Control 3 EC	0.100%	14:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_

Sst Jeffrey 1143

Badge No.: 1143

Date: 06/10/2022

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00881	Calib. Date: 02/01/2022
Certification File No.:	00882	Calib. No.: 00039
Linearity File No.:	00883	Cert. Date: 02/01/2022
Solution File No.:	00896	Cert. No.: 00032
Sequential File No.:	00896	Lin. Date: 02/01/2022
		Lin. No.: 00032
		Soln. Date: 05/07/2022
		Soln. No.: 00241
		File Date: 05/07/2022

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0020

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:54D		
Control 1 EC	0.099%	13:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:55D		
Control 2 EC	0.099%	13:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:56D		
Control 3 EC	0.099%	13:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: GALLO

First Name: DAVID

MI: M

Signature: \_\_\_\_\_



Badge No.: 1167

Date: 05/07/2022

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00881 Calib. Date: 02/01/2022 Calib. No.: 00039  
Certification File No.: 00882 Cert. Date: 02/01/2022 Cert. No.: 00032  
Linearity File No.: 00883 Lin. Date: 02/01/2022 Lin. No.: 00032  
Solution File No.: 00891 Soln. Date: 04/07/2022 Soln. No.: 00240  
Sequential File No.: 00891 File Date: 04/07/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286  
Control Solution %: 0.100% Expires: 03/17/2023  
Solution Control Lot: 21090 Bottle No.: 0018

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:11D		
Control 1 EC	0.099%	09:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12D		
Control 2 EC	0.098%	09:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14D		
Control 3 EC	0.098%	09:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CASAMASSIMA

First Name: STEPHEN

MI: N.

Signature: 



Badge No.: 1153

Date: 04/07/2022



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00881	Calib. Date: 02/01/2022
Certification File No.:	00882	Calib. No.: 00039
Linearity File No.:	00883	Cert. Date: 02/01/2022
Solution File No.:	00888	Cert. No.: 00032
Sequential File No.:	00888	Lin. Date: 02/01/2022
		Lin. No.: 00032
		Soln. Date: 03/06/2022
		Soln. No.: 00239
		File Date: 03/06/2022

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0286
Control Solution %:	0.100%		Expires: 03/17/2023
Solution Control Lot:	21090		Bottle No.: 0726

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:41S		
Control 1 EC	0.101%	06:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:42S		
Control 2 EC	0.100%	06:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:43S		
Control 3 EC	0.100%	06:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GUZMAN

First Name: MICHAEL

MI: C

Signature: \_\_\_\_\_



Badge No.: 1156

Date: 03/06/2022

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00855 Calib. Date: 08/02/2021 Calib. No.: 00038  
Certification File No.: 00856 Cert. Date: 08/02/2021 Cert. No.: 00031  
Linearity File No.: 00857 Lin. Date: 08/02/2021 Lin. No.: 00031  
Solution File No.: 00878 Soln. Date: 01/04/2022 Soln. No.: 00237  
Sequential File No.: 00878 File Date: 01/04/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 03/17/2023  
Solution Control Lot: 21090 Bottle No.: 0432

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:48S		
Control 1 EC	0.099%	18:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:49S		
Control 2 EC	0.097%	18:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:51S		
Control 3 EC	0.097%	18:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:52S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GUZMAN

First Name: MICHAEL

MI: C

Signature: 

Badge No.: 1156

Date: 01/04/2022