

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01003 Calib. Date: 08/25/2023 Calib. No.: 00043  
Certification File No.: 01004 Cert. Date: 08/25/2023 Cert. No.: 00036  
Linearity File No.: 01005 Lin. Date: 08/25/2023 Lin. No.: 00036  
Solution File No.: 01030 Soln. Date: 12/22/2023 Soln. No.: 00264  
Sequential File No.: 01030 File Date: 12/22/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 02/01/2025  
Solution Control Lot: 23040 Bottle No.: 0535

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:01S		
Control 1 EC	0.096%	12:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03S		
Control 2 EC	0.095%	12:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:04S		
Control 3 EC	0.096%	12:05S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MURRAY

First Name: JOSEPH

MI: T

Signature: \_\_\_\_\_

*J Murray* #1154

Badge No.: 1154

Date: 12/22/2023

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	Calib. No.:	00043
Calibration File No.:	01003	Calib. Date:	08/25/2023
Certification File No.:	01004	Cert. Date:	08/25/2023
Linearity File No.:	01005	Lin. Date:	08/25/2023
Solution File No.:	01021	Soln. Date:	10/16/2023
Sequential File No.:	01021	File Date:	10/16/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	0596

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:25D		
Control 1 EC	0.097%	19:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:26D		
Control 2 EC	0.097%	19:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:28D		
Control 3 EC	0.097%	19:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HEIMERL

First Name: CHARLES

MI: A

Signature: \_\_\_\_\_



Badge No.: 11102

Date: 10/16/2023

# Calibrating Unit

## New Standard Solution Report

Franklin

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01003 Calib. Date: 08/25/2023 Calib. No.: 00043  
Certification File No.: 01004 Cert. Date: 08/25/2023 Cert. No.: 00036  
Linearity File No.: 01005 Lin. Date: 08/25/2023 Lin. No.: 00036  
Solution File No.: 01016 Soln. Date: 09/28/2023 Soln. No.: 00261  
Sequential File No.: 01016 File Date: 09/28/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 02/01/2025  
Solution Control Lot: 23040 Bottle No.: 0670

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:00D		
Control 1 EC	0.099%	19:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:01D		
Control 2 EC	0.097%	19:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:03D		
Control 3 EC	0.098%	19:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_



Badge No.: 1143

Date: 09/28/2023

# Calibrating Unit New Standard Solution Report

Franklin

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01003 Calib. Date: 08/25/2023 Calib. No.: 00043  
Certification File No.: 01004 Cert. Date: 08/25/2023 Cert. No.: 00036  
Linearity File No.: 01005 Lin. Date: 08/25/2023 Lin. No.: 00036  
Solution File No.: 01015 Soln. Date: 09/28/2023 Soln. No.: 00260  
Sequential File No.: 01015 File Date: 09/28/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 02/01/2025  
Solution Control Lot: 23040 Bottle No.: 0092

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:29D		
Control 1 EC	0.099%	14:30D	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	14:30D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	14:30D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_



Badge No.: 1143

Date: 09/28/2023

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00965	Calib. Date: 03/21/2023
Certification File No.:	00966	Calib. No.: 00042
Linearity File No.:	00967	Cert. Date: 03/21/2023
Solution File No.:	00997	Cert. No.: 00035
Sequential File No.:	00997	Lin. Date: 03/21/2023
		Lin. No.: 00035
		Soln. Date: 08/18/2023
		Soln. No.: 00258
		File Date: 08/18/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 0692

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:44D		
Control 1 EC	0.100%	09:45D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		
Control 2 EC	0.099%	09:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47D		
Control 3 EC	0.100%	09:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_



Badge No.: 1143

Date: 08/18/2023

# Calibrating Unit

## New Standard Solution Report

Franklin

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00965 Calib. Date: 03/21/2023 Calib. No.: 00042  
Certification File No.: 00966 Cert. Date: 03/21/2023 Cert. No.: 00035  
Linearity File No.: 00967 Lin. Date: 03/21/2023 Lin. No.: 00035  
Solution File No.: 00992 Soln. Date: 07/21/2023 Soln. No.: 00257  
Sequential File No.: 00992 File Date: 07/21/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0286  
Control Solution %: 0.100% Expires: 02/01/2025  
Solution Control Lot: 23040 Bottle No.: 0695

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:08D		
Control 1 EC	0.101%	13:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09D		
Control 2 EC	0.100%	13:10D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10D		
Control 3 EC	0.100%	13:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: 

Badge No.: 1143

Date: 07/21/2023

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00965	Calib. Date: 03/21/2023
Certification File No.:	00966	Calib. No.: 00042
Linearity File No.:	00967	Cert. Date: 03/21/2023
Solution File No.:	00983	Cert. No.: 00035
Sequential File No.:	00983	Lin. Date: 03/21/2023
		Lin. No.: 00035
		Soln. Date: 06/21/2023
		Soln. No.: 00256
		File Date: 06/21/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0286
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 0258

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:34D		
Control 1 EC	0.102%	19:35D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:36D		
Control 2 EC	0.100%	19:36D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	19:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:37D		
Control 3 EC	0.101%	19:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	19:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HEIMERL

First Name: CHARLES

MI: A

Signature: \_\_\_\_\_



Badge No.: 11102

Date: 06/21/2023

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	Calib. No.:	00042
Calibration File No.:	00965	Cert. Date:	03/21/2023
Certification File No.:	00966	Lin. Date:	03/21/2023
Linearity File No.:	00967	Soln. Date:	05/22/2023
Solution File No.:	00977	File Date:	05/22/2023
Sequential File No.:	00977		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0286
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	0591

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:00D		
Control 1 EC	0.101%	13:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		
Control 2 EC	0.101%	13:02D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03D		
Control 3 EC	0.100%	13:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_



Badge No.: 1143

Date: 05/22/2023



Franklin

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.	Calib. No.:	00042
Calibration File No.:	00965	Calib. Date:	03/21/2023
Certification File No.:	00966	Cert. Date:	03/21/2023
Linearity File No.:	00967	Lin. Date:	03/21/2023
Solution File No.:	00972	Soln. Date:	04/19/2023
Sequential File No.:	00972	File Date:	04/19/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0286
Solution Control Lot:	22110	Expires:	04/04/2024
		Bottle No.:	0245

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:44D		
Control 1 EC	0.100%	13:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45D		
Control 2 EC	0.099%	13:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46D		
Control 3 EC	0.099%	13:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_



Badge No.: 1143

Date: 04/19/2023

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00927 Calib. Date: 11/03/2022 Calib. No.: 00041  
Certification File No.: 00928 Cert. Date: 11/03/2022 Cert. No.: 00034  
Linearity File No.: 00929 Lin. Date: 11/03/2022 Lin. No.: 00034  
Solution File No.: 00961 Soln. Date: 03/10/2023 Soln. No.: 00252  
Sequential File No.: 00961 File Date: 03/10/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
Control Solution %: 0.100% Expires: 04/04/2024  
Solution Control Lot: 22110 Bottle No.: 0020

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:11S		
Control 1 EC	0.099%	14:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:12S		
Control 2 EC	0.098%	14:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14S		
Control 3 EC	0.099%	14:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: LT Jeffrey 1143

Badge No.: 1143

Date: 03/10/2023

Franklin PD

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
 Location: FRANKLIN TOWNSHIP P.D.  
 Calibration File No.: 00927 Calib. Date: 11/03/2022 Calib. No.: 00041  
 Certification File No.: 00928 Cert. Date: 11/03/2022 Cert. No.: 00034  
 Linearity File No.: 00929 Lin. Date: 11/03/2022 Lin. No.: 00034  
 Solution File No.: 00957 Soln. Date: 02/08/2023 Soln. No.: 00251  
 Sequential File No.: 00957 File Date: 02/08/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0289  
 Control Solution %: 0.100% Expires: 04/04/2024  
 Solution Control Lot: 22110 Bottle No.: 1183

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:54S		
Control 1 EC	0.100%	12:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56S		
Control 2 EC	0.099%	12:56S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:57S		
Control 3 EC	0.099%	12:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGUIGAN First Name: JEFFREY MI: M

Signature: Sgt Jeffrey 1143

Badge No.: 1143  
Date: 02/08/2023

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00927	Calib. Date:	11/03/2022
Certification File No.:	00928	Cert. Date:	11/03/2022
Linearity File No.:	00929	Lin. Date:	11/03/2022
Solution File No.:	00948	Soln. Date:	01/09/2023
Sequential File No.:	00948	File Date:	01/09/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0289
Solution Control Lot:	22110	Expires:	04/04/2024
		Bottle No.:	0119

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:26S		
Control 1 EC	0.099%	03:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:28S		
Control 2 EC	0.097%	03:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:29S		
Control 3 EC	0.099%	03:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HEIMERL

First Name: CHARLES

MI: A

Signature: \_\_\_\_\_



Badge No.: 11102

Date: 01/09/2023