

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01038 Calib. Date: 01/10/2024 Calib. No.: 00045  
Certification File No.: 01039 Cert. Date: 01/10/2024 Cert. No.: 00038  
Linearity File No.: 01040 Lin. Date: 01/10/2024 Lin. No.: 00038  
Solution File No.: 01061 Soln. Date: 04/03/2024 Soln. No.: 00270  
Sequential File No.: 01061 File Date: 04/03/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210  
Control Solution %: 0.100% Expires: 02/01/2025  
Solution Control Lot: 23040 Bottle No.: 0748

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:56S		
Control 1 EC	0.100%	19:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57S		
Control 2 EC	0.099%	19:58S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:59S		
Control 3 EC	0.099%	19:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HEIMERL

First Name: CHARLES

MI: A

Signature: 

Badge No.: 11102

Date: 04/03/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01038	Calib. Date:	01/10/2024
Certification File No.:	01039	Cert. Date:	01/10/2024
Linearity File No.:	01040	Lin. Date:	01/10/2024
Solution File No.:	01055	Soln. Date:	03/04/2024
Sequential File No.:	01055	File Date:	03/04/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0210
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	0696

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:43S		
Control 1 EC	0.100%	02:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	02:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:44S		
Control 2 EC	0.098%	02:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:46S		
Control 3 EC	0.099%	02:46S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GUZMAN

First Name: MICHAEL

MI: C

Signature: \_\_\_\_\_

Badge No.: 1156

Date: 03/04/2024

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0071  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 01038 Calib. Date: 01/10/2024 Calib. No.: 00045  
Certification File No.: 01039 Cert. Date: 01/10/2024 Cert. No.: 00038  
Linearity File No.: 01040 Lin. Date: 01/10/2024 Lin. No.: 00038  
Solution File No.: 01049 Soln. Date: 02/10/2024 Soln. No.: 00268  
Sequential File No.: 01049 File Date: 02/10/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210  
Control Solution %: 0.100% Expires: 02/01/2025  
Solution Control Lot: 23040 Bottle No.: 0686

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:45S		
Control 1 EC	0.099%	03:45S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:46S		
Control 2 EC	0.098%	03:47S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:48S		
Control 3 EC	0.098%	03:48S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: DAVEY

First Name: LOUIS

MI: C

Signature: \_\_\_\_\_



Badge No.: 1162

Date: 02/10/2024

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0071
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	01038	Calib. Date:	01/10/2024
Certification File No.:	01039	Cert. Date:	01/10/2024
Linearity File No.:	01040	Lin. Date:	01/10/2024
Solution File No.:	01048	Soln. Date:	02/10/2024
Sequential File No.:	01048	File Date:	02/10/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0210
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	0673

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:20S		
Control 1 EC	-.---%	00:21S	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	00:21S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DAVEY

First Name: LOUIS

MI: C

Signature: \_\_\_\_\_



Badge No.: 1162

Date: 02/10/2024