

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0074
Location:	FRANKLIN TOWNSHIP P.D.	
Calibration File No.:	00436	Calib. Date: 09/18/2018
Certification File No.:	00437	Calib. No.: 00029
Linearity File No.:	00438	Cert. Date: 09/18/2018
Solution File No.:	00445	Cert. No.: 00023
Sequential File No.:	00445	Lin. Date: 09/18/2018
		Lin. No.: 00023
		Soln. Date: 10/28/2018
		Soln. No.: 00143
		File Date: 10/28/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDBJ-0006
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 1087

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:47S		
Control 1 EC	0.099%	21:48S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:49S		
Control 2 EC	0.098%	21:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:50S		
Control 3 EC	0.099%	21:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: \_\_\_\_\_

*Cpl. A. Myers #1151*

Badge No.: 1151

Date: 10/28/2018

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXB-0074  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00415 Calib. Date: 04/11/2018 Calib. No.: 00028  
Certification File No.: 00416 Cert. Date: 04/11/2018 Cert. No.: 00022  
Linearity File No.: 00417 Lin. Date: 04/11/2018 Lin. No.: 00022  
Solution File No.: 00433 Soln. Date: 08/28/2018 Soln. No.: 00141  
Sequential File No.: 00433 File Date: 08/28/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 1051

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:21D		
Control 1 EC	0.100%	00:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:23D		
Control 2 EC	0.099%	00:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24D		
Control 3 EC	0.099%	00:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: \_\_\_\_\_

*Cpl. A. Myers #1151*

Badge No.: 1151

Date: 08/28/2018

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXB-0074  
Location: FRANKLIN TOWNSHIP P.D.  
Calibration File No.: 00415 Calib. Date: 04/11/2018 Calib. No.: 00028  
Certification File No.: 00416 Cert. Date: 04/11/2018 Cert. No.: 00022  
Linearity File No.: 00417 Lin. Date: 04/11/2018 Lin. No.: 00022  
Solution File No.: 00426 Soln. Date: 07/18/2018 Soln. No.: 00140  
Sequential File No.: 00426 File Date: 07/18/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210  
Control Solution %: 0.100% Expires: 12/02/2018  
Solution Control Lot: 16390 Bottle No.: 1347

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:41D		
Control 1 EC	0.100%	04:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:42D		
Control 2 EC	0.098%	04:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:43D		
Control 3 EC	0.099%	04:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: GALLO

First Name: DAVID

MI: M

Signature: \_\_\_\_\_



1167

Badge No.: 1167

Date: 07/18/2018