

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0074
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 00470 **Calib. Date:** 08/06/2019 **Calib. No.:** 00031
Certification File No.: 00471 **Cert. Date:** 08/06/2019 **Cert. No.:** 00025
Linearity File No.: 00472 **Lin. Date:** 08/06/2019 **Lin. No.:** 00025
Solution File No.: 00480 **Soln. Date:** 12/04/2019 **Soln. No.:** 00154
Sequential File No.: 00480 **File Date:** 12/04/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDBJ-0006
Control Solution %: 0.100% **Expires:** 01/08/2021
Solution Control Lot: 19020 **Bottle No.:** 0258

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:55S		
Control 1 EC	0.100%	11:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:56S		
Control 2 EC	0.099%	11:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58S		
Control 3 EC	0.099%	11:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MYERS

First Name: CPL. A.

MI: -

Signature: _____



Badge No.: 1151

Date: 12/04/2019

Sequential File No.: 00480 **File Date:** 12/04/2019 **Soln. No.:** 00154

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDBJ-0006
Control Solution %: 0.100% **Expires:** 01/08/2021
Solution Control Lot: 19020 **Bottle No.:** 0258

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:55S		
Control 1 EC	0.100%	11:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:56S		
Control 2 EC	0.099%	11:57S	34.0°C	*** TEST PASSED ***

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00470	Calib. Date:	08/06/2019
Certification File No.:	00471	Cert. Date:	08/06/2019
Linearity File No.:	00472	Lin. Date:	08/06/2019
Solution File No.:	00477	Soln. Date:	11/03/2019
Sequential File No.:	00477	File Date:	11/03/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDBJ-0006
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0773

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:14S		
Control 1 EC	0.098%	18:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:15S		
Control 2 EC	0.098%	18:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:16S		
Control 3 EC	0.098%	18:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MYERS	First Name: CPL. A.	MI: -
Signature: <u> 1151</u>	Badge No.: 1151	Date: 11/03/2019

Sequential File No.:	00477	File Date:	11/03/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDBJ-0006
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0773

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:14S		
Control 1 EC	0.098%	18:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:15S		
Control 2 EC	0.098%	18:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:16S	34.0°C	*** TEST PASSED ***

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0074
 Location: FRANKLIN TOWNSHIP P.D. (Buena)
 Calibration File No.: 00450 Calib. Date: 02/21/2019 Calib. No.: 00030
 Certification File No.: 00451 Cert. Date: 02/21/2019 Cert. No.: 00024
 Linearity File No.: 00452 Lin. Date: 02/21/2019 Lin. No.: 00024
 Solution File No.: 00469 Soln. Date: 07/14/2019 Soln. No.: 00150
 Sequential File No.: 00469 File Date: 07/14/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210
 Control Solution %: 0.100% Expires: 01/08/2021
 Solution Control Lot: 19020 Bottle No.: 0835

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:46D		
Control 1 EC	0.098%	20:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:48D		
Control 2 EC	0.097%	20:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:49D		
Control 3 EC	0.096%	20:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NEHER First Name: MICHAEL MI: E
 Signature: *Off. Michael Neher #1155* Badge No.: 1155
 Date: 07/14/2019

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: FRANKLIN TOWNSHIP P.D. Serial No.: ARXB-0074
Calibration File No.: 00450 Calib. Date: 02/21/2019 Calib. No.: 00030
Certification File No.: 00451 Cert. Date: 02/21/2019 Cert. No.: 00024
Linearity File No.: 00452 Lin. Date: 02/21/2019 Lin. No.: 00024
Solution File No.: 00467 Soln. Date: 06/10/2019 Soln. No.: 00149
Sequential File No.: 00467 File Date: 06/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210
Control Solution %: 0.100% Expires: 01/31/2020
Solution Control Lot: 18050 Bottle No.: 0942

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:01D		
Control 1 EC	0.097%	04:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:02D		
Control 2 EC	0.096%	04:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:03D		
Control 3 EC	0.097%	04:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CASAMASSIMA

First Name: STEPHEN

MI: N.

Signature: _____

Badge No.: 1153

Date: 06/10/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0074
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 00450 Calib. Date: 02/21/2019 Calib. No.: 00030
Certification File No.: 00451 Cert. Date: 02/21/2019 Cert. No.: 00024
Linearity File No.: 00452 Lin. Date: 02/21/2019 Lin. No.: 00024
Solution File No.: 00456 Soln. Date: 03/24/2019 Soln. No.: 00147
Sequential File No.: 00456 File Date: 03/24/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210
Control Solution %: 0.100% Expires: 01/31/2020
Solution Control Lot: 18050 Bottle No.: 0987

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:25S		
Control 1 EC	0.099%	00:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:27S		
Control 2 EC	0.099%	00:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:28S		
Control 3 EC	0.099%	00:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NEHER

First Name: MICHAEL

MI: E

Signature: _____

Off. Michael Neher #1155

Badge No.: 1155

Date: 03/24/2019