

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0074
 Location: FRANKLIN TOWNSHIP P.D.
 Calibration File No.: 00483 Calib. Date: 01/22/2020 Calib. No.: 00032
 Certification File No.: 00484 Cert. Date: 01/22/2020 Cert. No.: 00026
 Linearity File No.: 00485 Lin. Date: 01/22/2020 Lin. No.: 00026
 Solution File No.: 00496 Soln. Date: 05/31/2020 Soln. No.: 00159
 Sequential File No.: 00496 File Date: 05/31/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXJ S3-0210
 Control Solution %: 0.100% Expires: 06/10/2021
 Solution Control Lot: 19150 Bottle No.: 0794

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:32D		
Control 1 EC	0.098%	03:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:33D		
Control 2 EC	0.097%	03:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	03:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:34D		
Control 3 EC	0.096%	03:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PASCULLI

First Name: NICHOLAS

MI: P

Signature: _____

CP. Nicholas Pasculli #1050

Badge No.: 1150

Date: 05/31/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0074
Location:	FRANKLIN TOWNSHIP P.D.		
Calibration File No.:	00483	Calib. Date:	01/22/2020
Certification File No.:	00484	Cert. No.:	00032
Linearity File No.:	00485	Cert. Date:	01/22/2020
Solution File No.:	00494	Lin. Date:	01/22/2020
Sequential File No.:	00494	Lin. No.:	00026
		Soln. Date:	04/16/2020
		Soln. No.:	00158
		File Date:	04/16/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXJ S3-0210
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0082

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:50D		
Control 1 EC	0.097%	09:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:52D		
Control 2 EC	0.096%	09:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53D		
Control 3 EC	0.096%	09:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____

Badge No.: 1143

Date: 04/16/2020