

Calibrating Unit

New Standard Solution Report

Buena

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0074
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 00605 Calib. Date: 11/06/2023 Calib. No.: 00042
Certification File No.: 00606 Cert. Date: 11/06/2023 Cert. No.: 00035
Linearity File No.: 00607 Lin. Date: 11/06/2023 Lin. No.: 00035
Solution File No.: 00617 Soln. Date: 02/23/2024 Soln. No.: 00197
Sequential File No.: 00617 File Date: 02/23/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDBJ-0006
Control Solution %: 0.100% Expires: 02/01/2025
Solution Control Lot: 23040 Bottle No.: 0033

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:10S		
Control 1 EC	0.101%	13:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11S		
Control 2 EC	0.100%	13:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12S		
Control 3 EC	0.101%	13:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGUIGAN

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 1143

Date: 02/23/2024

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0074
Location: FRANKLIN TOWNSHIP P.D.
Calibration File No.: 00605 Calib. Date: 11/06/2023 Calib. No.: 00042
Certification File No.: 00606 Cert. Date: 11/06/2023 Cert. No.: 00035
Linearity File No.: 00607 Lin. Date: 11/06/2023 Lin. No.: 00035
Solution File No.: 00613 Soln. Date: 01/20/2024 Soln. No.: 00196
Sequential File No.: 00613 File Date: 01/20/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDBJ-0006
Control Solution %: 0.100% Expires: 02/01/2025
Solution Control Lot: 23040 Bottle No.: 0718

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:35S		
Control 1 EC	0.099%	14:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:36S		
Control 2 EC	0.098%	14:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37S		
Control 3 EC	0.099%	14:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MURRAY

First Name: JOSEPH

MI: T

Signature: _____



Badge No.: 1154

Date: 01/20/2024