

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Franklin Township Police Department 1571 Delsea Drive Franklinville, NJ 08322</b>	2. Contact: <b>Capt. Gordon Muller</b>	4. Date: <b>10/30/2020</b>
	3. Phone Number: <b>856-694-1415</b>	5. County: <b>Gloucester</b>
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARWC-0071</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>

9. Reason for Service:  
**Complaint received of the inability to remotely connect to the Alcotest instrument on (3) three consecutive occasions while attempting to retrieve data.**

10. Comments:  
**Operational condition of the instrument was assessed by confirming that the dedicated phone line was functioning properly. A separate test instrument was connected to the dedicated phone line and the test instrument remotely connected and successfully retrieved data. The instrument was placed out of service and returned to Draeger for evaluation.**

Last Known Sequential File #: 00835

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p>Tpr. I Matthew R. Watson #7078  <small>Name &amp; Badge Number (Print)</small></p>	<p><i>Tpr. I [Signature]</i>  <small>Signature</small></p>	<p><i>10/30/2020</i>  <small>Date</small></p>
---	--	---

# RETURN AND REPAIR FORM



## Customer Information

B: \_\_\_\_\_

S: \_\_\_\_\_

Company Name: Franklin Township Police Department, Franklinville, NJ

Date Received: 10/11/2020

Date given to service: 10/11/2020

Carrier:  FedEx  UPS  USPS      Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WC-0071

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	2.0	
MP Labor	Labor	1.5	

Repair Notes: Replaced modem board.

\_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

CAL W/QC AND OPS CHECK

Service Technician:  Date: 03/22/2021

Delivery receipt

Customer no. 150057092      Number of report 305597076      Date of receiving / /      Delivery receipt date 03/22/2021  
Please reference on inquiries

Consignee, 150057092  
FRANKLIN TOWNSHIP POLICE DEPARTMENT  
ATTN: CAPT. G. MULLER  
PO ARWC-0071  
1571 DELSEA DR  
FRANKLINVILLE NJ 08322-2391

Customer, 150057092  
FRANKLIN TOWNSHIP POLICE DEPARTMENT  
ATTN: CAPT. G. MULLER  
PO ARWC-0071  
1571 DELSEA DR  
FRANKLINVILLE NJ 08322-2391

Your order  
Date of order:  
11/03/2020  
Purch.ord.no.:  
ARWC-0071

Branch text

Draeger contact person  
Service Support-IDT  
800-437-2437 OPT 5  
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
001	ALCOTEST 7110 MK III USA	8314246	1030559688	ARWC-0071

Draeger Inc.  
Our Tax ID: 23-1699096  
3135 Quarry Road; Telford, PA 18969  
An Equal Opportunity Employer M / F / V  
Telephone 800-437-2437  
<http://www.draeger.com>

Remit to:  
LOCKBOX (Standard USPS)  
Draeger, Inc.  
PO Box 13369  
Newark, New Jersey  
07101-3362

Remit to:  
LOCKBOX (Overnight)  
FIS Lockbox Processing  
Lockbox #13369  
100 Grove Road  
Suite E  
West Deptford, NJ 08066

Remit US Wire Transfers to:  
Account Name: Draeger Inc.  
Account Number: 00-494-936  
Transit Routing: 021001033  
SWIFT: BKTRUS33  
Deutsche Bank Trust Company Americas  
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 150057092 Number of report 305597076 Date of receiving 11/03/2020 Please reference on inquiries

Customer FRANKLIN TOWNSHIP POLICE DEPARTMENT 1571 DELSEA DR FRANKLINVILLE NJ 08322-2391

Consignee FRANKLIN TOWNSHIP POLICE DEPARTMENT 1571 DELSEA DR FRANKLINVILLE NJ 08322-2391

Your order Date of order: 11/03/2020 Your reference: ARWC-0071

NL-Text Maintenance device list

Your contact person Service Support-IDT 800-437-2437 OPT 5 SS-IDT@draeger.com

Page 1

Item Quantity Part no. Description 1.5 H REPAIR / ARWC-0071 R001 Repair time - normal hours

STATE OF NEW JERSEY  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Franklin Township Police Department          1571 Delsea Drive          Franklinville, NJ 08322</b>	2. Contact: <b>Capt. Gordon Muller</b> 3. Phone Number: <b>856-694-1415</b>	4. Date: <b>04/08/2021</b> 5. County: <b>Gloucester</b>
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARWC-0071</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>

9. Reason for Service:  
**The alcotest instrument was returned from outside evaluation and placed back in service.**

10. Comments:  
**See Draeger Return and Repair form.**

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Matthew R. Watson #7078  
 Name & Badge Number (Print)

*Tpr. I Matthew R. Watson #7078* 04/08/2021  
 Signature Date