

# ALCOTEST CHECKLIST

Municipality: LONGPORT

Alcotest Ser. #: ARXA -0067

County: ATLANTIC

Calibration Date: 08-13-2010

1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.
2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used  
Ser. #: DDXC 53-0135
3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent  
Ser. #: DDWA P2-199
4. **Digital Temperature Measuring System Report of Calibration**  
Ser. #: 80637585
5.  A. Alcotest 7110 Calibration Record  
 B. Alcotest 7110 Calibration Certificate **Part I - Control Tests**  
 C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests**  
 D. **Alcotest Card** of operator/coordinator who performed tests  
 E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator]  
Ser. #: DDWA P2-016
6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:

<input checked="" type="checkbox"/> A.	0.04% Solution	<u>10</u>	<u>A</u>	<u>073</u>
<input checked="" type="checkbox"/> B.	0.08% Solution	<u>10</u>	<u>A</u>	<u>074</u>
<input checked="" type="checkbox"/> C.	0.10% Solution	<u>09</u>	<u>D</u>	<u>065</u>
<input checked="" type="checkbox"/> D.	0.16% Solution	<u>10</u>	<u>A</u>	<u>075</u>
7. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:

<input checked="" type="checkbox"/> A.	0.04% Calibration/Linearity Testing	<u>DDWE 53-0196</u>
<input checked="" type="checkbox"/> B.	0.08% Calibration/Linearity Testing	<u>DDWE 53-0203</u>
<input checked="" type="checkbox"/> C.	0.10% Calibration/Linearity Testing	<u>DDXC 53-0135</u> [Same as instrument's CU34 unit]
<input checked="" type="checkbox"/> D.	0.16% Calibration/Linearity Testing	<u>DDWE 53-0205</u>
8.  A. New Standard Solution Report - following Calibration  
 B. Calibrating CU34 Unit for same [same as CU34 unit on instrument]  
 C. Certificate of Analysis 0.10% solution for same  
Lot #: 09 A 062  
 D. Alcotest card of operator/coordinator who completed change

**Dräger**safety

**Alcotest<sup>®</sup> 7110 MKIII-C**

**CERTIFICATE OF ACCURACY**

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:


SERIAL NUMBER:

8-10-2010

ARXA-0067

Draeger Safety Diagnostics, Inc.

\_\_\_\_\_ *AS*



Drägersafety

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: \_\_\_\_\_

Serial Number:

DDXCS3-0135

Certification Date

Technician

Re-Certification Due Date

2/25/10

zm

2/25/11

## Drägersafety

### ALCOTEST® 7110 TEMPERATURE PROBE

# CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDWAP2-199

2/25/10

2/25/11

Probe Value

103

Draeger Safety Diagnostics, Inc.  
Technical Service Department

zm



Calibration  
Certificate No. 1750.01

Calibration complies with ISO 9001  
ISO/IEC 17025 AND ANSI/NCSL Z540-1



Cert. No.: 4000-2035927

Traceable® Certificate of Calibration for Digital Thermometer

Instrument Identification:

Model: 61220-601      S/N: 80637585      Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Probe	149	3/06/09	A82225037-3
Thermistor Module	A27129	8/22/09	1000248949
Temperature Calibration Bath TC218	A73332		
Temperature Calibration Bath TC155	93139		
Thermistor Module	A27129	8/22/09	1000248949
Temperature Probe	157	5/28/09	A8519038-4

Certificate Information:

Technician: 68      Procedure: CAL-06      Cal Date: 11/18/08      Cal Due: 11/18/10  
Test Conditions: 25.0°C      34.0 %RH      1031 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±uc	TUR
°C		N.A.		0.001	-0.002	Y	-0.049	0.051	0.013	3.8:1
°C		N.A.		25.001	25.003	Y	24.951	25.051	0.013	3.8:1
°C		N.A.		60.001	59.998	Y	59.951	60.051	0.018	2.8:1
°C		N.A.		100.001	100.001	Y	99.951	100.051	0.013	3.8:1

This instrument was calibrated using instruments traceable to National Institute of Standards and Technology. :

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±uc=Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Date=MM/DD/YY.

*Wallace Berry*  
Wallace Berry, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA  
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.  
Control Company is ISO 9001:2000 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-AQ-HDU.  
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).

# Alcotest 7110 Calibration Record

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00433 Calib. Date: 08/13/2010 Calib. No.: 00013  
Certification File No.: 00383 Cert. Date: 02/18/2010 Cert. No.: 00007  
Linearity File No.: 00384 Lin. Date: 02/18/2010 Lin. No.: 00007  
Solution File No.: 00431 Soln. Date: 08/10/2010 Soln. No.: 00096  
Sequential File No.: 00433 File Date: 08/13/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
Control Solution %: 0.100% Expires: 04/22/2011  
Solution Control Lot: 09D065 Bottle No.: 1341

## Coordinator

Last Name: DELLANOCE First Name: JOSEPH MI: S

Signature: Tr II J. Dellanoce # 6027 Badge No.: 6027  
Date: 08/13/2010

\*Black Key Temperature Probe Serial.....# DDWAP2-016 *JSD*

\*Digital NIST Temperature Measuring System Serial.....# 80637585 *JSD*

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

# Alcotest 7110 Calibration Certificate

## Part I - Control Tests

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00433 Calib. Date: 08/13/2010 Calib. No.: 00013  
Certification File No.: 00434 Cert. Date: 08/13/2010 Cert. No.: 00008  
Linearity File No.: 00384 Lin. Date: 02/18/2010 Lin. No.: 00007  
Solution File No.: 00431 Soln. Date: 08/10/2010 Soln. No.: 00096  
Sequential File No.: 00434 File Date: 08/13/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0135  
Control Solution %: 0.100% Expires: 04/22/2011  
Solution Control Lot: 09D065 Bottle No.: 1341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:06D		
Control 1 EC	0.099%	15:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		
Control 2 EC	0.099%	15:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09D		
Control 3 EC	0.099%	15:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:10D		

All tests within acceptable tolerance.

### Coordinator

Last Name: DELBANOCE First Name: JOSEPH MI: S  
Signature:  #6027 Badge No.: 6027  
Date: 08/13/2010

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to and consistent with, the current "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

# Alcotest 7110 Calibration Certificate

## Part II - Linearity Tests

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00433 Calib. Date: 08/13/2010 Calib. No.: 00013  
Certification File No.: 00434 Cert. Date: 08/13/2010 Cert. No.: 00008  
Linearity File No.: 00435 Lin. Date: 08/13/2010 Lin. No.: 00008  
Solution File No.: 00431 Soln. Date: 08/10/2010 Soln. No.: 00096  
Sequential File No.: 00435 File Date: 08/13/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0196  
Control Solution %: 0.040% Expires: 01/12/2012  
Solution Control Lot: 10A073 Bottle No.: 0144

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0203  
Control Solution %: 0.080% Expires: 01/15/2012  
Solution Control Lot: 10A074 Bottle No.: 0133

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0205  
Control Solution %: 0.160% Expires: 01/21/2012  
Solution Control Lot: 10A075 Bottle No.: 0059

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:25D		
Control 1 EC	0.040%	15:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.040%	15:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26D		
Control 2 EC	0.040%	15:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.040%	15:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28D		
Control 3 EC	0.081%	15:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.080%	15:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30D		
Control 4 EC	0.081%	15:31D	34.0°C	*** TEST PASSED ***
Control 4 IR	0.079%	15:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32D		
Control 5 EC	0.161%	15:32D	34.0°C	*** TEST PASSED ***
Control 5 IR	0.160%	15:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34D		
Control 6 EC	0.162%	15:34D	34.0°C	*** TEST PASSED ***
Control 6 IR	0.159%	15:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:36D		

All tests within acceptable tolerance.

### Coordinator

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

*Tpr II J. Dellanoce #6027*

Badge No.: 6027

Date: 08/13/2010

DEPARTMENT OF  
**Traffic and Public Safety**  
 This is to certify that

**Joseph S. Dellanoce**  
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL-BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1966 IN THE OPERATION OF THE **Alcotest 7110 MKIII-C**  
 A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29th DAY OF March

TWO THOUSAND AND Five

*Joseph S. Dellanoce*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*Mark J. Blum*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	2/15/07	GC/PA	C. POTER
2.	12/7/09	DC/PA	William S. ...
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 263B (Rev. 12/04)



*State of New Jersey*

OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
PO Box 080  
TRENTON, NJ 08625-0080

JON S. CORZINE  
*Governor*

ANNE MILGRAM  
*Attorney General*

January 14, 2008

Col. Joseph R. Fuentes, Superintendent  
Division of State Police  
Division Headquarters  
P.O. Box 7068  
West Trenton, New Jersey 08628

**Re: Breath Test Coordinator/Instructor, Certification -  
Trooper Joseph S. Dellanoce #6027**

Dear Colonel Fuentes:

Pursuant to the provisions of *N.J.A.C. 13:51-2.1 (b) and (c)*, as adopted and promulgated under the provisions of *N.J.S.A. 39:4-50.3, 39:3-10.25 and 12:7-56*, I hereby approve Trooper Joseph S. Dellanoce #6027, as a duly certified Breath Test Coordinator/Instructor. This approval is effective immediately.

Very truly yours,

  
Anne Milgram  
Attorney General

jk

c. Trooper Joseph S. Dellanoce #6027, Alcohol/Drug Test Unit, Division of State Police  
Lt. Mark Kolodzieski, Unit Head, Alcohol/Drug Test Unit, Division of State Police





# CERTIFICATE

This is to certify that


**Joseph S. Dellanoce #6027**

has successfully completed the two day Draeger Safety Diagnostics, Inc. Operator Training and Preventive Maintenance Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and  
Maintenance Technician**

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: 4/12/2007

  
\_\_\_\_\_  
Instructor: Hansueli Ryser

**Dräger**

DEPARTMENT OF  
**Public Safety**  
 This is to certify that  
**Joseph B. D'Alfonso**  
 New Jersey State Police

---

IS QUALIFIED AND COMPETENT TO INSTRUCT IN THE USE OF THE BREATHALYZER  
 IN ACCORDANCE WITH THE LAWS OF THE STATE OF NEW JERSEY PURSUANT TO CHAPTER  
 A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AND SEAL OF THE STATE OF NEW JERSEY THIS 13th DAY OF Dec.  
 TWO THOUSAND AND 01

*C. P. Kelly*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*J. P. ...*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES: 8/30 - 9/3/99

Refresher Course

DATE	PLACE	INSTRUCTOR
1. <u>12/12/01</u>	<u>A.G.P.C.</u>	<u>ROUSE</u>
2. <u>7-8-03</u>	<u>ACTV</u>	<u>A. ...</u>
3. <u>10/3-09</u>	<u>ACTV</u>	<u>A. ...</u>
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

SP-293B (Rev. 1/00)



**NEW JERSEY STATE POLICE**  
**Alcohol/Drug Testing Unit**

**Trooper Steve Dellanoce #6027**  
**Breath Test Coordinator / Instructor**

Hamilton Technology Complex  
Suite 400  
1200 Negron Drive  
Hamilton, NJ 08691

*(609) 584-5000 ext. 5608*  
*cell: (609) 947-9211*  
*Fax: (609) 584-9359*  
*email: LPP6027@gw.njsp.org*

**Dräger**safety

**ALCOTEST® 7110 TEMPERATURE PROBE**

**CERTIFICATE OF ACCURACY**

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDWAP2-016

Certification date:

2-3-10

Next Certification due:

2-3-11

Probe Value

106

Dräger Safety Diagnostics, Inc.  
Technical Service Department

mm



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

CHRIS CHRISTIE  
Governor

PAULA T. DOW  
Acting Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 2/2/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A073

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0479 to 0.0481 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 12, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Kenneth W. Klawalek, M.S.  
Assistant Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

  
Notary

Linda L. Desantis  
Notary Public, New Jersey  
My Commission Expires 8-17-14





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

CHRIS CHRISTIE  
Governor

PAULA T. DOW  
Acting Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

CERTIFICATION OF ANALYSIS  
0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.


MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 2/3/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A074


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0954 to 0.0958 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
Kenneth W. Kawalek, M.S.  
Assistant Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of February, 2010.

  
Notary

Linda L. Desantis  
Notary Public, New Jersey  
My Commission Expires 8-17-14





State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON NJ 08628-0068  
 (609) 882-2000

JON S. CORZINE  
 Governor

ANNE MILGRAM  
 Attorney General

COLONEL JOSEPH R. FUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 5/08/2009

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 09D065

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1201 to 0.1210 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 22, 2011.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare  
 Chief Forensic Scientist  
 Division of State Police

Sworn to and subscribed before me this 19<sup>th</sup> day of May, 2009.

Notary

Linda L. DeSantis  
 My Commission  
 Expires Aug. 17, 2009







State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

PAULA T. DOW
Acting Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 2/4/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10A075

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1913 to 0.1919 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 21, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Signature of Kenneth W. Kawalek, M.S.
Kenneth W. Kawalek, M.S.
Assistant Chief Forensic Scientist
Division of State Police


Sworn to and subscribed before me this 19th day of February, 2010.

Signature of Notary Linda L. Desantis

Notary

Linda L. Desantis
Notary Public, New Jersey
My Commission Expires 8-17-14





Drägersafety

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK II A

Other: \_\_\_\_\_

Serial Number:

DDWES3-0196

Certification Date


Technician

Re-Certification Due Date

2/3/10

DM

2/3/11



Dräger safety

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: \_\_\_\_\_

Serial Number:

DDWES3-0203

Certification Date


Technician

Re-Certification Due Date

2/3/10

DM

2/3/11



Draeger safety

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: \_\_\_\_\_

Serial Number:

DDWES3-0205

Certification Date

Technician

Re-Certification Due Date

2/3/10

DM

2/3/11

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00433	Calib. Date: 08/13/2010
Certification File No.:	00434	Calib. No.: 00013
Linearity File No.:	00435	Cert. Date: 08/13/2010
Solution File No.:	00436	Cert. No.: 00008
Sequential File No.:	00436	Lin. Date: 08/13/2010
		Lin. No.: 00008
		Soln. Date: 08/13/2010
		Soln. No.: 00097
		File Date: 08/13/2010

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0135
Control Solution %: 0.100%		Expires: 01/15/2011
Solution Control Lot: 09A062		Bottle No.: 0171

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:42D		
Control 1 EC	0.101%	16:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44D		
Control 2 EC	0.101%	16:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45D		
Control 3 EC	0.101%	16:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*TEMPERATURE PROBE SERIAL NUMBER: DDWAP2-199 JSD*

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

*Tr II JSD # 6027*

Badge No.: 6027

Date: 08/13/2010



JON S. CORZINE  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

ANNE MILGRAM  
Attorney General  
COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 2/11/2009

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 09A062

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1195 to 0.1197 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2011.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

*Ajit R. Tungare*  
Ajit R. Tungare  
Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 30<sup>th</sup> day of February, 2009.

*Linda L. DeSantis*  
Linda L. DeSantis  
My Commission  
Expires Aug. 17, 2009



New Jersey Is An Equal Opportunity Employer  
Printed on Recycled Paper and Recyclable

