

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0067
Location:	LONGPORT POLICE	
Calibration File No.:	00893	Calib. Date: 01/14/2020
Certification File No.:	00894	Calib. No.: 00036
Linearity File No.:	00895	Cert. Date: 01/14/2020
Solution File No.:	00899	Cert. No.: 00030
Sequential File No.:	00899	Lin. Date: 01/14/2020
		Lin. No.: 00030
		Soln. Date: 03/11/2020
		Soln. No.: 00238
		File Date: 03/11/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0154
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1419

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:10S		
Control 1 EC	0.099%	14:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:11S		
Control 2 EC	0.097%	14:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:13S		
Control 3 EC	0.097%	14:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature:  @ 331

Badge No.: 331

Date: 03/11/2020

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0067  
Location: LONGPORT POLICE  
Calibration File No.: 00893 Calib. Date: 01/14/2020 Calib. No.: 00036  
Certification File No.: 00894 Cert. Date: 01/14/2020 Cert. No.: 00030  
Linearity File No.: 00895 Lin. Date: 01/14/2020 Lin. No.: 00030  
Solution File No.: 00898 Soln. Date: 02/07/2020 Soln. No.: 00237  
Sequential File No.: 00898 File Date: 02/07/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0154  
Control Solution %: 0.100% Expires: 10/30/2021  
Solution Control Lot: 19300 Bottle No.: 1396

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:40S		
Control 1 EC	0.099%	14:41S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42S		
Control 2 EC	0.098%	14:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43S		
Control 3 EC	0.098%	14:44S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TIBBITT

First Name: DEAN

MI: M

Signature: \_\_\_\_\_



Badge No.: 331

Date: 02/07/2020