

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Longport Police Department 2305 Atlantic Avenue Longport, NJ 08403		2. Contact: Tpr. I Steve Dellanoce #6027	4. Date: 01/10/2014
		3. Phone Number: 609-947-9211	5. County: Atlantic
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXA-0067</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Alcotest instrument returned from outside maintenance and placed back in service.

10. Comments:
See Draeger Return and Repair form.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Steve Dellanoce #6027

Name & Badge Number (Print)

Signature

Tpr I Steve Dellanoce #6027 *1/10/14*

Date

Packing Slip

Customer no Order No Order date
 150062710 10672800 01/07/2014

Packing slip no. Ship date
 80827153 01/07/2014

Please reference on inquiries

ship to

150062710
 BOROUGH OF LONGPORT POLICE
 ATTN: CHIEF PACENTRILLI
 ATLANTIC COUNTY
 2305 ATLANTIC AVENUE
 LONGPORT, NJ 08403
 USA

bill to

Your Purchase Order Number

7110: ARXA-0067
 Any questions? Please contact:
 AMANDA DILLEY

Customer

BOROUGH OF LONGPORT POLICE
 ATTN: CHIEF PACENTRILLI
 ATLANTIC COUNTY
 2305 ATLANTIC AVENUE
 LONGPORT, NJ 08403
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110: ARXA-0067 REC'D ON 12-18-13
0010	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150062710 10672800 01/07/2014

Packing slip no. Ship date
80827153 01/07/2014

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	2 EA	6808486 PLATES, 7110--9510 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Longport P.D., Longport, NJDate Received: 12/18/13Date given to service: 12/18/13Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STDProduct: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 WholeSerial Number: AR XA-0067Printer Serial#: ARSim Serial#: DDProbe Serial#: DDWarranty Expires: May 2016Description: A B Plus Demo Screener Trade In

Accessories

 110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.Fuel Cell = AREL-0392CAL W/QC AND OPS CHECKService Technician: BCDate: 01/06/14

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Longport Police Department 2305 Atlantic Avenue Longport, NJ 08403	2. Contact: Tpr. I Steve Dellanoce #6027	4. Date: 12/16/2013
3. Phone Number: 609-947-9211		5. County: Atlantic
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXA-0067</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
While reviewing the performance history of the instrument low EC results on control tests were noted.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed low Electrochemical Sensor (EC) results (Fuel Cell) on control tests. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Please replace Fuel Cell.

Last Known Sequential File #: 00677

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Steve Dellanoce #6027

Name & Badge Number (Print)

Signature

Date

Tpr I Steve Dellanoce #6027 *12/16/13*