

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00780 Calib. Date: 10/31/2014 Calib. No.: 00027
Certification File No.: 00781 Cert. Date: 10/31/2014 Cert. No.: 00021
Linearity File No.: 00782 Lin. Date: 10/31/2014 Lin. No.: 00021
Solution File No.: 00787 Soln. Date: 12/23/2014 Soln. No.: 00178
Sequential File No.: 00787 File Date: 12/23/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 02/14/2016
Solution Control Lot: 14B127 Bottle No.: 0339

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:10S		
Control 1 EC	0.100%	13:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11S		
Control 2 EC	0.100%	13:12S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13S		
Control 3 EC	0.100%	13:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: Sgt. Frank Talarico 59

Badge No.: 59

Date: 12/23/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00780 **Calib. Date:** 10/31/2014 **Calib. No.:** 00027
Certification File No.: 00781 **Cert. Date:** 10/31/2014 **Cert. No.:** 00021
Linearity File No.: 00782 **Lin. Date:** 10/31/2014 **Lin. No.:** 00021
Solution File No.: 00784 **Soln. Date:** 11/20/2014 **Soln. No.:** 00177
Sequential File No.: 00784 **File Date:** 11/20/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWA S3-0012
Control Solution %: 0.100% **Expires:** 02/14/2016
Solution Control Lot: 14B127 **Bottle No.:** 0334

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:38S		
Control 1 EC	0.101%	17:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:40S		
Control 2 EC	0.099%	17:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:41S		
Control 3 EC	0.100%	17:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:43S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 11/20/2014

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00759 Calib. Date: 05/16/2014 Calib. No.: 00026
Certification File No.: 00760 Cert. Date: 05/16/2014 Cert. No.: 00020
Linearity File No.: 00761 Lin. Date: 05/16/2014 Lin. No.: 00020
Solution File No.: 00778 Soln. Date: 10/16/2014 Soln. No.: 00175
Sequential File No.: 00778 File Date: 10/16/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 02/14/2016
Solution Control Lot: 14B127 Bottle No.: 0326

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:48D		
Control 1 EC	0.097%	05:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:49D		
Control 2 EC	0.097%	05:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:51D		
Control 3 EC	0.096%	05:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	05:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 10/16/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00759 Calib. Date: 05/16/2014 Calib. No.: 00026
Certification File No.: 00760 Cert. Date: 05/16/2014 Cert. No.: 00020
Linearity File No.: 00761 Lin. Date: 05/16/2014 Lin. No.: 00020
Solution File No.: 00777 Soln. Date: 09/23/2014 Soln. No.: 00174
Sequential File No.: 00777 File Date: 09/23/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 02/14/2016
Solution Control Lot: 14B127 Bottle No.: 0332

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:49D		
Control 1 EC	0.100%	19:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:50D		
Control 2 EC	0.098%	19:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	19:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:52D		
Control 3 EC	0.098%	19:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	19:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Kull

Badge No.: 61

Date: 09/23/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00759	Calib. Date:	05/16/2014
Certification File No.:	00760	Cert. Date:	05/16/2014
Linearity File No.:	00761	Lin. Date:	05/16/2014
Solution File No.:	00775	Soln. Date:	08/31/2014
Sequential File No.:	00775	File Date:	08/31/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	14B127	Expires:	02/14/2016
		Bottle No.:	0327

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:06D		
Control 1 EC	0.099%	11:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08D		
Control 2 EC	0.097%	11:08D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:08D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09D		
Control 3 EC	0.097%	11:10D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:10D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 08/31/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00759	Calib. Date: 05/16/2014
Certification File No.:	00760	Cert. No.: 00020
Linearity File No.:	00761	Lin. Date: 05/16/2014
Solution File No.:	00773	Soln. No.: 00172
Sequential File No.:	00773	File Date: 08/06/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	14B127	Expires: 02/14/2016
		Bottle No.: 0331

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:10D		
Control 1 EC	0.100%	15:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12D		
Control 2 EC	0.100%	15:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13D		
Control 3 EC	0.099%	15:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____

Sgt. Tomas Whalen

Badge No.: 61

Date: 08/06/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00759	Calib. Date: 05/16/2014
Certification File No.:	00760	Cert. No.: 00026
Linearity File No.:	00761	Cert. Date: 05/16/2014
Solution File No.:	00771	Lin. No.: 00020
Sequential File No.:	00771	Lin. Date: 05/16/2014
		Soln. No.: 00171
		Soln. Date: 07/15/2014
		File Date: 07/15/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	13C111	Expires: 03/12/2015
		Bottle No.: 0661

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:38D		
Control 1 EC	0.097%	08:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40D		
Control 2 EC	0.097%	08:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41D		
Control 3 EC	0.098%	08:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WIALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 07/15/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00759	Calib. Date: 05/16/2014
Certification File No.:	00760	Calib. No.: 00026
Linearity File No.:	00761	Cert. Date: 05/16/2014
Solution File No.:	00764	Cert. No.: 00020
Sequential File No.:	00764	Lin. Date: 05/16/2014
		Lin. No.: 00020
		Soln. Date: 06/06/2014
		Soln. No.: 00169
		File Date: 06/06/2014

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA 53-0011
Solution Control Lot:	13C111	Expires: 03/12/2015
		Bottle No.: 0658

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:13D		
Control 1 EC	0.101%	20:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:15D		
Control 2 EC	0.099%	20:15D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:15D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:16D		
Control 3 EC	0.100%	20:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 06/06/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE			
Calibration File No.:	00739	Calib. Date:	11/18/2013	Calib. No.: 00025
Certification File No.:	00740	Cert. Date:	11/18/2013	Cert. No.: 00019
Linearity File No.:	00741	Lin. Date:	11/18/2013	Lin. No.: 00019
Solution File No.:	00757	Soln. Date:	05/11/2014	Soln. No.: 00167
Sequential File No.:	00757	File Date:	05/11/2014	

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0011
Control Solution %:	0.100%			Expires:	03/12/2015
Solution Control Lot:	13C111			Bottle No.:	0652

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:03D		
Control 1 EC	0.097%	20:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:05D		
Control 2 EC	0.097%	20:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:06D		
Control 3 EC	0.098%	20:07D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:07D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 05/11/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00739	Calib. Date: 11/18/2013
Certification File No.:	00740	Calib. No.: 00025
Linearity File No.:	00741	Cert. Date: 11/18/2013
Solution File No.:	00754	Cert. No.: 00019
Sequential File No.:	00754	Lin. Date: 11/18/2013
		Lin. No.: 00019
		Soln. Date: 04/16/2014
		Soln. No.: 00166
		File Date: 04/16/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWA S3-0011
Control Solution %:	0.100%		Expires: 03/12/2015
Solution Control Lot:	13C111		Bottle No.: 0654

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:25D		
Control 1 EC	0.099%	18:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	18:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:26D		
Control 2 EC	0.097%	18:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:28D		
Control 3 EC	0.097%	18:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:29D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 04/16/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00739	Calib. Date:	11/18/2013
Certification File No.:	00740	Cert. Date:	11/18/2013
Linearity File No.:	00741	Lin. Date:	11/18/2013
Solution File No.:	00753	Soln. Date:	03/25/2014
Sequential File No.:	00753	File Date:	03/25/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DIJWA S3-0011
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0653

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:16S		
Control 1 EC	0.100%	19:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:16S		
Control 2 EC	0.098%	19:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:19S		
Control 3 EC	0.098%	19:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: **WHALEN**

First Name: **TOMAS**

MI: **F**

Signature: Sgt. Tomas Whalen

Badge No.: **61**

Date: **03/25/2014**

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00739	Calib. Date: 11/18/2013
Certification File No.:	00740	Calib. No.: 00025
Linearity File No.:	00741	Cert. Date: 11/18/2013
Solution File No.:	00752	Cert. No.: 00019
Sequential File No.:	00752	Lin. Date: 11/18/2013
		Lin. No.: 00019
		Soln. Date: 03/25/2014
		Soln. No.: 00164
		File Date: 03/25/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	13C111	Expires: 03/12/2015
		Bottle No.: 0656

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:55S		
Control 1 EC	-.--%	17:56S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.--%	17:56S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.--%	--:--		
Control 2 EC	-.--%	--:--	N/A	
Control 2 IR	-.--%	--:--	N/A	
Ambient Air Blank	-.--%	--:--		
Control 3 EC	-.--%	--:--	N/A	
Control 3 IR	-.--%	--:--	N/A	
Ambient Air Blank	-.--%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 03/25/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00739	Calib. Date: 11/18/2013
Certification File No.:	00740	Calib. No.: 00025
Linearity File No.:	00741	Cert. Date: 11/18/2013
Solution File No.:	00750	Cert. No.: 00019
Sequential File No.:	00750	Lin. Date: 11/18/2013
		Lin. No.: 00019
		Soln. Date: 02/28/2014
		Soln. No.: 00163
		File Date: 02/28/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWA S3-0011
Control Solution %:	0.100%		Expires: 03/12/2015
Solution Control Lot:	13C111		Bottle No.: 0657

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:14S		
Control 1 EC	0.100%	20:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:16S		
Control 2 EC	0.097%	20:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17S		
Control 3 EC	0.098%	20:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 02/28/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII C	Serial No.: ARJIM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00739	Calib. Date: 11/18/2013
Certification File No.:	00740	Cert. No.: 00025
Linearity File No.:	00741	Cert. Date: 11/18/2013
Solution File No.:	00749	Lin. No.: 00019
Sequential File No.:	00749	Soln. No.: 00162
		File Date: 01/28/2014
Calibrating Unit:	W1ET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3 0011
Solution Control Lot:	137111	Expires: 03/12/2015
		Bottle No.: 0651

Function	Result	Time	Temperature	Comment(s)
	%BAC	(MM:SS)	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:28S		
Control 1 EC	0.100%	20:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:30S		
Control 2 EC	0.098%	20:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:32S		
Control 3 EC	0.097%	20:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MIF

Signature: 

Badge No.: 61

Date: 01/28/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00739	Calib. Date:	11/18/2013
Certification File No.:	00740	Cert. Date:	11/18/2013
Linearity File No.:	00741	Lin. Date:	11/18/2013
Solution File No.:	00746	Soln. Date:	01/09/2014
Sequential File No.:	00746	File Date:	01/09/2014
Calibrating Unit:	WJT	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0660

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:45S		
Control 1 EC	0.101%	22:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:47S		
Control 2 EC	0.098%	22:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:48S		
Control 3 EC	0.098%	22:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WIALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Kelly

Badge No.: 61

Date: 01/09/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM 0073
Location:	MERCANTVILLE POLICE	
Calibration File No.:	00739	Calib. Date: 11/18/2013
Certification File No.:	00740	Cert. No.: 00025
Linearity File No.:	00741	Cert. Date: 11/18/2013
Solution File No.:	00745	Lin. Date: 11/18/2013
Sequential File No.:	00745	Lin. No.: 00019
		Soln. No.: 00160
		File Date: 01/09/2014
Calibrating Unit:	WET	Model No.: CU 34
Control Solution %:	0.100%	Serial No.: JDDWA S3-0011
Solution Control Lot:	13C111	Expires: 03/12/2015
		Bottle No.: 0665

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:44S		
Control 1 EC	----%	19:45S	21.9°C	SIMULATOR TEMP. ERROR
Control 1 IR	----%	19:45S	21.9°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	----%	----		
Control 2 EC	----%	----	N/A	
Control 2 IR	----%	----	N/A	
Ambient Air Blank	----%	----		
Control 3 EC	----%	----	N/A	
Control 3 IR	----%	----	N/A	
Ambient Air Blank	----%	----		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: *Sgt Tomas Whalen*

Badge No.: 61

Date: 01/09/2014