

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00807	Calib. Date: 09/09/2015
Certification File No.:	00808	Calib. No.: 00029
Linearity File No.:	00809	Cert. Date: 09/09/2015
Solution File No.:	00821	Cert. No.: 00023
Sequential File No.:	00821	Lin. Date: 09/09/2015
		Lin. No.: 00023
		Soln. Date: 12/24/2015
		Soln. No.: 00194
		File Date: 12/24/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWA S3-0011
Control Solution %:	0.100%		Expires: 02/14/2016
Solution Control Lot:	14B127		Bottle No.: 0346

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:20S		
Control 1 EC	0.103%	08:20S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.104%	08:20S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:21S		
Control 2 EC	0.102%	08:22S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:22S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		
Control 3 EC	0.103%	08:23S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.103%	08:23S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 12/24/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00807	Calib. Date: 09/09/2015
Certification File No.:	00808	Calib. No.: 00029
Linearity File No.:	00809	Cert. Date: 09/09/2015
Solution File No.:	00819	Cert. No.: 00023
Sequential File No.:	00819	Lin. Date: 09/09/2015
		Lin. No.: 00023
		Soln. Date: 11/30/2015
		Soln. No.: 00193
		File Date: 11/30/2015
<b>Calibrating Unit:</b>	<b>WET</b>	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	14B127	Expires: 02/14/2016
		Bottle No.: 0350

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:26S		
Control 1 EC	0.102%	10:26S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:26S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:27S		
Control 2 EC	0.100%	10:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29S		
Control 3 EC	0.101%	10:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	10:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 11/30/2015

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00807	Calib. No.: 00029
Certification File No.:	00808	Cert. No.: 00023
Linearity File No.:	00809	Lin. No.: 00023
Solution File No.:	00817	Soln. No.: 00192
Sequential File No.:	00817	File Date: 11/07/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	14B127	Expires: 02/14/2016
		Bottle No.: 0347

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:22S		
Control 1 EC	0.103%	08:23S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.103%	08:23S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		
Control 2 EC	0.103%	08:24S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:24S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25S		
Control 3 EC	0.102%	08:26S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	08:26S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 11/07/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00807	Calib. Date: 09/09/2015
Certification File No.:	00808	Cert. No.: 00029
Linearity File No.:	00809	Cert. Date: 09/09/2015
Solution File No.:	00815	Lin. No.: 00023
Sequential File No.:	00815	Lin. Date: 09/09/2015
		Soln. No.: 00191
		Soln. Date: 10/11/2015
		File Date: 10/11/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	14B127	Expires: 02/14/2016
		Bottle No.: 0342

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:07D		
Control 1 EC	0.101%	20:08D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.103%	20:08D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09D		
Control 2 EC	0.101%	20:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		
Control 3 EC	0.101%	20:11D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	20:11D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 10/11/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00807	Calib. Date:	09/09/2015
Certification File No.:	00808	Cert. Date:	09/09/2015
Linearity File No.:	00809	Lin. Date:	09/09/2015
Solution File No.:	00811	Soln. Date:	09/14/2015
Sequential File No.:	00811	File Date:	09/14/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0011
Control Solution %:	0.100%			Expires:	02/14/2016
Solution Control Lot:	14B127			Bottle No.:	0341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:15D		
Control 1 EC	0.103%	13:16D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:16D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17D		
Control 2 EC	0.102%	13:18D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:18D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18D		
Control 3 EC	0.102%	13:19D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:19D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DUJTP2-168 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TPR. I K W. Q #6704

Badge No.: 6704

Date: 09/14/2015



**Dräger**

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Draeger Safety Diagnostics, Inc.**

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: \_\_\_\_\_

Serial Number:

DDWA 53-0011

Certification Date

8-6-15

Technician

BC

Re-Certification Due Date

8-6-16

# Dräger

## ALCOTEST® 7110 TEMPERATURE PROBE

### CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe, within 12 months of the certification date below or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDUS P2-168

Certification date:

8-5-15

Next Certification due:

8-5-16

Probe Value

104

**Draeger Safety Diagnostics, Inc.**  
Technical Service Department

BC

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
**Location:** MERCHANTVILLE POLICE  
**Calibration File No.:** 00792 **Calib. Date:** 04/10/2015 **Calib. No.:** 00028  
**Certification File No.:** 00793 **Cert. Date:** 04/10/2015 **Cert. No.:** 00022  
**Linearity File No.:** 00794 **Lin. Date:** 04/10/2015 **Lin. No.:** 00022  
**Solution File No.:** 00806 **Soln. Date:** 08/19/2015 **Soln. No.:** 00188  
**Sequential File No.:** 00806 **File Date:** 08/19/2015

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWA S3-0012  
**Control Solution %:** 0.100% **Expires:** 02/14/2016  
**Solution Control Lot:** 14B127 **Bottle No.:** 0336

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:26D		
Control 1 EC	0.102%	10:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:27D		
Control 2 EC	0.100%	10:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 3 EC	0.100%	10:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 08/19/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE	Calib. No.:	00028
Calibration File No.:	00792	Cert. Date:	04/10/2015
Certification File No.:	00793	Lin. Date:	04/10/2015
Linearity File No.:	00794	Soln. Date:	07/21/2015
Solution File No.:	00804	File Date:	07/21/2015
Sequential File No.:	00804		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0012
Control Solution %:	0.100%			Expires:	02/14/2016
Solution Control Lot:	14B127			Bottle No.:	0328

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:19D		
Control 1 EC	0.101%	10:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		
Control 2 EC	0.100%	10:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21D		
Control 3 EC	0.101%	10:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: \_\_\_\_\_



Badge No.: 59

Date: 07/21/2015



# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
Location: MERCHANTVILLE POLICE  
Calibration File No.: 00792 Calib. Date: 04/10/2015 Calib. No.: 00028  
Certification File No.: 00793 Cert. Date: 04/10/2015 Cert. No.: 00022  
Linearity File No.: 00794 Lin. Date: 04/10/2015 Lin. No.: 00022  
Solution File No.: 00803 Soln. Date: 06/23/2015 Soln. No.: 00186  
Sequential File No.: 00803 File Date: 06/23/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012  
Control Solution %: 0.100% Expires: 02/14/2016  
Solution Control Lot: 14B127 Bottle No.: 0330

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:43D		
Control 1 EC	0.101%	03:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:45D		
Control 2 EC	0.101%	03:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:46D		
Control 3 EC	0.101%	03:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: St. John J. J. 59

Badge No.: 59

Date: 06/23/2015

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
Location: MERCHANTVILLE POLICE  
Calibration File No.: 00792 Calib. Date: 04/10/2015 Calib. No.: 00028  
Certification File No.: 00793 Cert. Date: 04/10/2015 Cert. No.: 00022  
Linearity File No.: 00794 Lin. Date: 04/10/2015 Lin. No.: 00022  
Solution File No.: 00800 Soln. Date: 05/31/2015 Soln. No.: 00185  
Sequential File No.: 00800 File Date: 05/31/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012  
Control Solution %: 0.100% Expires: 02/14/2016  
Solution Control Lot: 14B127 Bottle No.: 0349

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:25D		
Control 1 EC	0.100%	05:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	05:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:27D		
Control 2 EC	0.101%	05:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	05:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28D		
Control 3 EC	0.101%	05:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:30D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 05/31/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00792	Calib. Date:	04/10/2015
Certification File No.:	00793	Cert. Date:	04/10/2015
Linearity File No.:	00794	Lin. Date:	04/10/2015
Solution File No.:	00798	Soln. Date:	05/08/2015
Sequential File No.:	00798	File Date:	05/08/2015

Calib. No.:	00028
Cert. No.:	00022
Lin. No.:	00022
Soln. No.:	00184

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0012
Control Solution %:	0.100%			Expires:	02/14/2016
Solution Control Lot:	14B127			Bottle No.:	0335

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:52D		
Control 1 EC	0.101%	19:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54D		
Control 2 EC	0.101%	19:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55D		
Control 3 EC	0.100%	19:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Kully

Badge No.: 61

Date: 05/08/2015

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
<b>Location:</b>	MERCHANTVILLE POLICE	
<b>Calibration File No.:</b>	00780	<b>Calib. Date:</b> 10/31/2014
<b>Certification File No.:</b>	00781	<b>Calib. No.:</b> 00027
<b>Linearity File No.:</b>	00782	<b>Cert. Date:</b> 10/31/2014
<b>Solution File No.:</b>	00791	<b>Cert. No.:</b> 00021
<b>Sequential File No.:</b>	00791	<b>Lin. Date:</b> 10/31/2014
		<b>Lin. No.:</b> 00021
		<b>Soln. Date:</b> 03/27/2015
		<b>Soln. No.:</b> 00182
		<b>File Date:</b> 03/27/2015
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDWA S3-0012
<b>Solution Control Lot:</b>	14B127	<b>Expires:</b> 02/14/2016
		<b>Bottle No.:</b> 0338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:57S		
Control 1 EC	0.099%	16:58S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:59S		
Control 2 EC	0.100%	17:00S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:01S		
Control 3 EC	0.100%	17:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 03/27/2015

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
<b>Location:</b>	MERCHANTVILLE POLICE	
<b>Calibration File No.:</b>	00780	<b>Calib. Date:</b> 10/31/2014
<b>Certification File No.:</b>	00781	<b>Calib. No.:</b> 00027
<b>Linearity File No.:</b>	00782	<b>Cert. Date:</b> 10/31/2014
<b>Solution File No.:</b>	00790	<b>Cert. No.:</b> 00021
<b>Sequential File No.:</b>	00790	<b>Lin. Date:</b> 10/31/2014
		<b>Lin. No.:</b> 00021
		<b>Soln. Date:</b> 03/01/2015
		<b>Soln. No.:</b> 00181
		<b>File Date:</b> 03/01/2015

<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34	<b>Serial No.:</b> DDWA S3-0012
<b>Control Solution %:</b>	0.100%		<b>Expires:</b> 02/14/2016
<b>Solution Control Lot:</b>	14B127		<b>Bottle No.:</b> 0329

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:36S		
Control 1 EC	0.101%	20:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:38S		
Control 2 EC	0.100%	20:38S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39S		
Control 3 EC	0.099%	20:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Kull

Badge No.: 61

Date: 03/01/2015

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00780	Calib. Date: 10/31/2014
Certification File No.:	00781	Cert. No.: 00027
Linearity File No.:	00782	Cert. Date: 10/31/2014
Solution File No.:	00789	Lin. No.: 00021
Sequential File No.:	00789	Lin. Date: 10/31/2014
		Soln. No.: 00180
		Soln. Date: 02/05/2015
		File Date: 02/05/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	14B127	Expires: 02/14/2016
		Bottle No.: 0344

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:09S		
Control 1 EC	0.102%	01:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:11S		
Control 2 EC	0.101%	01:11S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:12S		
Control 3 EC	0.101%	01:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 02/05/2015

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
**Location:** MERCHANTVILLE POLICE  
**Calibration File No.:** 00780 **Calib. Date:** 10/31/2014 **Calib. No.:** 00027  
**Certification File No.:** 00781 **Cert. Date:** 10/31/2014 **Cert. No.:** 00021  
**Linearity File No.:** 00782 **Lin. Date:** 10/31/2014 **Lin. No.:** 00021  
**Solution File No.:** 00788 **Soln. Date:** 01/13/2015 **Soln. No.:** 00179  
**Sequential File No.:** 00788 **File Date:** 01/13/2015

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWA S3-0012  
**Control Solution %:** 0.100% **Expires:** 02/14/2016  
**Solution Control Lot:** 14B127 **Bottle No.:** 0343

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:57S		
Control 1 EC	0.100%	05:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:58S		
Control 2 EC	0.097%	05:59S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:00S		
Control 3 EC	0.098%	06:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 01/13/2015