

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00844	Calib. Date: 07/28/2016
Certification File No.:	00845	Calib. No.: 00031
Linearity File No.:	00846	Cert. Date: 07/28/2016
Solution File No.:	00855	Cert. No.: 00025
Sequential File No.:	00855	Lin. Date: 07/28/2016
		Lin. No.: 00025
		Soln. Date: 11/22/2016
		Soln. No.: 00210
		File Date: 11/22/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 1178

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:58S		
Control 1 EC	0.103%	06:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:59S		
Control 2 EC	0.099%	07:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:01S		
Control 3 EC	0.100%	07:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: _____

Sgt. Tal Talarico 59

Badge No.: 59

Date: 11/22/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00844	Calib. Date:	07/28/2016
Certification File No.:	00845	Cert. Date:	07/28/2016
Linearity File No.:	00846	Lin. Date:	07/28/2016
Solution File No.:	00854	Soln. Date:	11/22/2016
Sequential File No.:	00854	File Date:	11/22/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	1176

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:31S		
Control 1 EC	0.105%	03:32S	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.102%	03:32S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	03:33S		
Control 2 EC	0.101%	03:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	03:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:35S		
Control 3 EC	0.101%	03:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:36S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: _____

Sgt. Frank Talarico 59

Badge No.: 59

Date: 11/22/2016

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00844	Calib. Date: 07/28/2016
Certification File No.:	00845	Calib. No.: 00031
Linearity File No.:	00846	Cert. Date: 07/28/2016
Solution File No.:	00853	Cert. No.: 00025
Sequential File No.:	00853	Lin. Date: 07/28/2016
		Lin. No.: 00025
		Soln. Date: 10/31/2016
		Soln. No.: 00208
		File Date: 10/31/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 1198

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:45S		
Control 1 EC	0.100%	04:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	04:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:46S		
Control 2 EC	0.100%	04:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:48S		
Control 3 EC	0.101%	04:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: _____



Badge No.: 59

Date: 10/31/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00844	Calib. Date: 07/28/2016
Certification File No.:	00845	Cert. No.: 00031
Linearity File No.:	00846	Cert. Date: 07/28/2016
Solution File No.:	00851	Lin. No.: 00025
Sequential File No.:	00851	Lin. Date: 07/28/2016
		Soln. No.: 00207
		Soln. Date: 10/11/2016
		File Date: 10/11/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWA S3-0012
Control Solution %:	0.100%		Expires: 08/04/2017
Solution Control Lot:	15H140		Bottle No.: 1188

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:02D		
Control 1 EC	0.101%	21:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04D		
Control 2 EC	0.101%	21:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05D		
Control 3 EC	0.101%	21:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: _____

Sgt. Frank Talarico 59

Badge No.: 59

Date: 10/11/2016

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00844	Calib. Date: 07/28/2016
Certification File No.:	00845	Calib. No.: 00031
Linearity File No.:	00846	Cert. Date: 07/28/2016
Solution File No.:	00849	Cert. No.: 00025
Sequential File No.:	00849	Lin. Date: 07/28/2016
		Lin. No.: 00025
		Soln. Date: 09/18/2016
		Soln. No.: 00206
		File Date: 09/18/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 1184

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:42D		
Control 1 EC	0.100%	02:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:44D		
Control 2 EC	0.099%	02:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45D		
Control 3 EC	0.099%	02:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: _____

Frank Talarico 59

Badge No.: 59

Date: 09/18/2016

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00844 Calib. Date: 07/28/2016 Calib. No.: 00031
Certification File No.: 00845 Cert. Date: 07/28/2016 Cert. No.: 00025
Linearity File No.: 00846 Lin. Date: 07/28/2016 Lin. No.: 00025
Solution File No.: 00848 Soln. Date: 08/21/2016 Soln. No.: 00205
Sequential File No.: 00848 File Date: 08/21/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 1200

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:56D		
Control 1 EC	0.095%	11:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		
Control 2 EC	0.095%	11:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 3 EC	0.096%	11:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CLINE

First Name: JONATHAN

MI: M

Signature: _____



Badge No.: 65

Date: 08/21/2016

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00826	Calib. Date:	02/24/2016
Certification File No.:	00827	Cert. Date:	02/24/2016
Linearity File No.:	00828	Lin. Date:	02/24/2016
Solution File No.:	00843	Soln. Date:	07/05/2016
Sequential File No.:	00843	File Date:	07/05/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0011
Control Solution %:	0.100%			Expires:	08/04/2017
Solution Control Lot:	15H140			Bottle No.:	1187

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:48D		
Control 1 EC	0.103%	16:49D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:49D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:49D		
Control 2 EC	0.100%	16:50D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:50D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:51D		
Control 3 EC	0.098%	16:52D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:52D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:52D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 07/05/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00826	Calib. Date: 02/24/2016
Certification File No.:	00827	Cert. No.: 00030
Linearity File No.:	00828	Cert. Date: 02/24/2016
Solution File No.:	00841	Lin. No.: 00024
Sequential File No.:	00841	Lin. Date: 02/24/2016
		Soln. Date: 06/12/2016
		Soln. No.: 00202
		File Date: 06/12/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 1189

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:35D		
Control 1 EC	0.103%	05:36D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:36D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:37D		
Control 2 EC	0.101%	05:37D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:37D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:38D		
Control 3 EC	0.101%	05:39D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:39D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: Sgt. Tal Talarico 59

Badge No.: 59

Date: 06/12/2016

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00826 Calib. Date: 02/24/2016 Calib. No.: 00030
Certification File No.: 00827 Cert. Date: 02/24/2016 Cert. No.: 00024
Linearity File No.: 00828 Lin. Date: 02/24/2016 Lin. No.: 00024
Solution File No.: 00840 Soln. Date: 05/15/2016 Soln. No.: 00201
Sequential File No.: 00840 File Date: 05/15/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 1194

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:02D		
Control 1 EC	0.104%	01:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:03D		
Control 2 EC	0.102%	01:04D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:04D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:05D		
Control 3 EC	0.101%	01:05D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:05D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: Sgt. Frank Talarico 59

Badge No.: 59

Date: 05/15/2016

Signature: Sgt. Frank Talarico 59

Badge No.: 59

Date: 05/15/2016

Signature: Sgt. Frank Talarico 59

Badge No.: 59

Date: 05/15/2016

Signature: Sgt. Frank Talarico 59

Badge No.: 59

Date: 05/15/2016

Signature: Sgt. Frank Talarico 59

Badge No.: 59

Date: 05/15/2016

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00826 Calib. Date: 02/24/2016 Calib. No.: 00030
Certification File No.: 00827 Cert. Date: 02/24/2016 Cert. No.: 00024
Linearity File No.: 00828 Lin. Date: 02/24/2016 Lin. No.: 00024
Solution File No.: 00839 Soln. Date: 04/22/2016 Soln. No.: 00200
Sequential File No.: 00839 File Date: 04/22/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 1191

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:50D		
Control 1 EC	0.100%	17:51D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:51D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52D		
Control 2 EC	0.097%	17:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:53D		
Control 3 EC	0.098%	17:54D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:54D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN First Name: TOMAS MI: F

Signature: Sgt. Tom Whale Badge No.: 61
Date: 04/22/2016

Signature: Sgt. Tom Whale Badge No.: 61
Date: 04/22/2016

Signature: Sgt. Tom Whale Badge No.: 61
Date: 04/22/2016

Signature: Sgt. Tom Whale Badge No.: 61
Date: 04/22/2016

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00826	Calib. Date: 02/24/2016
Certification File No.:	00827	Calib. No.: 00030
Linearity File No.:	00828	Cert. Date: 02/24/2016
Solution File No.:	00837	Cert. No.: 00024
Sequential File No.:	00837	Lin. Date: 02/24/2016
		Lin. No.: 00024
		Soln. Date: 03/26/2016
		Soln. No.: 00199
		File Date: 03/26/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWA S3-0011
Control Solution %:	0.100%		Expires: 08/04/2017
Solution Control Lot:	15H140		Bottle No.: 1182

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:07S		
Control 1 EC	0.101%	07:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:09S		
Control 2 EC	0.097%	07:09S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:09S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:10S		
Control 3 EC	0.097%	07:11S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:11S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN First Name: TOMAS MI: F

Signature: Sgt. Tomas Whalen Badge No.: 61 Date: 03/26/2016

Signature: Sgt. Tomas Whalen Badge No.: 61 Date: 03/26/2016

Signature: Sgt. Tomas Whalen Badge No.: 61 Date: 03/26/2016

Sequential File No.:	00837	File Date:	03/26/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	1182

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:07S		
Control 1 EC	0.101%	07:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:09S		
Control 2 EC	0.097%	07:09S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:09S	34.1°C	*** TEST PASSED ***

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00826	Calib. No.: 00030
Certification File No.:	00827	Cert. No.: 00024
Linearity File No.:	00828	Lin. No.: 00024
Solution File No.:	00834	Soln. No.: 00198
Sequential File No.:	00834	File Date: 03/20/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA 83-0011
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 1193

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:15S		
Control 1 EC	0.103%	07:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:17S		
Control 2 EC	0.099%	07:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:18S		
Control 3 EC	0.099%	07:19S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:19S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: Sgt. Frank Talarico 59

Badge No.: 59

Date: 03/20/2016