

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00894	Calib. Date: 12/07/2017
Certification File No.:	00895	Calib. No.: 00034
Linearity File No.:	00896	Cert. Date: 12/07/2017
Solution File No.:	00908	Cert. No.: 00028
Sequential File No.:	00908	Lin. Date: 12/07/2017
		Lin. No.: 00028
		Soln. Date: 03/16/2018
		Soln. No.: 00234
		File Date: 03/16/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 1141

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:36S		
Control 1 EC	0.102%	07:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.097%	07:39S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		
Control 3 EC	0.097%	07:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 03/16/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00894	Calib. Date:	12/07/2017
Certification File No.:	00895	Cert. No.:	00034
Linearity File No.:	00896	Cert. Date:	12/07/2017
Solution File No.:	00905	Lin. No.:	00028
Sequential File No.:	00905	Lin. Date:	12/07/2017
		Soln. Date:	02/22/2018
		Soln. No.:	00233
		File Date:	02/22/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	1054

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:55S		
Control 1 EC	0.101%	03:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:57S		
Control 2 EC	0.100%	03:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58S		
Control 3 EC	0.100%	03:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 02/22/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00894 Calib. Date: 12/07/2017 Calib. No.: 00034
Certification File No.: 00895 Cert. Date: 12/07/2017 Cert. No.: 00028
Linearity File No.: 00896 Lin. Date: 12/07/2017 Lin. No.: 00028
Solution File No.: 00903 Soln. Date: 01/29/2018 Soln. No.: 00232
Sequential File No.: 00903 File Date: 01/29/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0086

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:51S		
Control 1 EC	0.100%	01:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:52S		
Control 2 EC	0.099%	01:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:54S		
Control 3 EC	0.099%	01:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 01/29/2018

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 00894 Calib. Date: 12/07/2017 Calib. No.: 00034
Certification File No.: 00895 Cert. Date: 12/07/2017 Cert. No.: 00028
Linearity File No.: 00896 Lin. Date: 12/07/2017 Lin. No.: 00028
Solution File No.: 00901 Soln. Date: 01/18/2018 Soln. No.: 00231
Sequential File No.: 00901 File Date: 01/18/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0147

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:56S		
Control 1 EC	0.101%	04:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:57S		
Control 2 EC	0.096%	04:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:59S		
Control 3 EC	0.097%	05:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: Sgt. Tal Talarico

Badge No.: 59

Date: 01/18/2018