

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00929	Calib. Date: 10/29/2018
Certification File No.:	00930	Calib. No.: 00037
Linearity File No.:	00931	Cert. Date: 10/29/2018
Solution File No.:	00938	Cert. No.: 00030
Sequential File No.:	00938	Lin. Date: 10/29/2018
		Lin. No.: 00030
		Soln. Date: 12/17/2018
		Soln. No.: 00244
		File Date: 12/17/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 1393

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:46S		
Control 1 EC	0.099%	09:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48S		
Control 2 EC	0.100%	09:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49S		
Control 3 EC	0.099%	09:50S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 12/17/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00929	Calib. Date: 10/29/2018
Certification File No.:	00930	Calib. No.: 00037
Linearity File No.:	00931	Cert. Date: 10/29/2018
Solution File No.:	00937	Cert. No.: 00030
Sequential File No.:	00937	Lin. Date: 10/29/2018
		Lin. No.: 00030
		Soln. Date: 11/26/2018
		Soln. No.: 00243
		File Date: 11/26/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 1465

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:49S		
Control 1 EC	0.100%	08:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S		
Control 2 EC	0.100%	08:51S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52S		
Control 3 EC	0.100%	08:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: \_\_\_\_\_



Badge No.: 59

Date: 11/26/2018

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
Location: MERCHANTVILLE POLICE  
Calibration File No.: 00911 Calib. Date: 05/31/2018 Calib. No.: 00036  
Certification File No.: 00912 Cert. Date: 05/31/2018 Cert. No.: 00029  
Linearity File No.: 00913 Lin. Date: 05/31/2018 Lin. No.: 00029  
Solution File No.: 00928 Soln. Date: 10/21/2018 Soln. No.: 00241  
Sequential File No.: 00928 File Date: 10/21/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011  
Control Solution %: 0.100% Expires: 03/13/2020  
Solution Control Lot: 18090 Bottle No.: 1357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:05D		
Control 1 EC	0.101%	08:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06D		
Control 2 EC	0.101%	08:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07D		
Control 3 EC	0.101%	08:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	08:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 10/21/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00911	Calib. Date: 05/31/2018
Certification File No.:	00912	Calib. No.: 00036
Linearity File No.:	00913	Cert. Date: 05/31/2018
Solution File No.:	00925	Cert. No.: 00029
Sequential File No.:	00925	Lin. Date: 05/31/2018
		Lin. No.: 00029
		Soln. Date: 09/27/2018
		Soln. No.: 00240
		File Date: 09/27/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0219

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:43D		
Control 1 EC	0.101%	19:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:44D		
Control 2 EC	0.100%	19:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:45D		
Control 3 EC	0.100%	19:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 09/27/2018

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
Location: MERCHANTVILLE POLICE  
Calibration File No.: 00911 Calib. Date: 05/31/2018 Calib. No.: 00036  
Certification File No.: 00912 Cert. Date: 05/31/2018 Cert. No.: 00029  
Linearity File No.: 00913 Lin. Date: 05/31/2018 Lin. No.: 00029  
Solution File No.: 00922 Soln. Date: 08/21/2018 Soln. No.: 00239  
Sequential File No.: 00922 File Date: 08/21/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011  
Control Solution %: 0.100% Expires: 10/28/2018  
Solution Control Lot: 16330 Bottle No.: 0095

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:47D		
Control 1 EC	0.100%	15:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48D		
Control 2 EC	0.100%	15:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49D		
Control 3 EC	0.100%	15:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: \_\_\_\_\_



Badge No.: 59

Date: 08/21/2018

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0073  
Location: MERCHANTVILLE POLICE  
Calibration File No.: 00911 Calib. Date: 05/31/2018 Calib. No.: 00036  
Certification File No.: 00912 Cert. Date: 05/31/2018 Cert. No.: 00029  
Linearity File No.: 00913 Lin. Date: 05/31/2018 Lin. No.: 00029  
Solution File No.: 00917 Soln. Date: 07/24/2018 Soln. No.: 00238  
Sequential File No.: 00917 File Date: 07/24/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011  
Control Solution %: 0.100% Expires: 10/28/2018  
Solution Control Lot: 16330 Bottle No.: 1060

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:01D		
Control 1 EC	0.100%	08:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02D		
Control 2 EC	0.100%	08:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03D		
Control 3 EC	0.100%	08:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 07/24/2018

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00911	Calib. Date: 05/31/2018
Certification File No.:	00912	Calib. No.: 00036
Linearity File No.:	00913	Cert. Date: 05/31/2018
Solution File No.:	00916	Cert. No.: 00029
Sequential File No.:	00916	Lin. Date: 05/31/2018
		Lin. No.: 00029
		Soln. Date: 06/27/2018
		Soln. No.: 00237
		File Date: 06/27/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0032

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:58D		
Control 1 EC	0.101%	19:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:59D		
Control 2 EC	0.101%	20:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01D		
Control 3 EC	0.101%	20:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 06/27/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00894	Calib. Date: 12/07/2017
Certification File No.:	00895	Calib. No.: 00034
Linearity File No.:	00896	Cert. Date: 12/07/2017
Solution File No.:	00908	Cert. No.: 00028
Sequential File No.:	00908	Lin. Date: 12/07/2017
		Lin. No.: 00028
		Soln. Date: 03/16/2018
		Soln. No.: 00234
		File Date: 03/16/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 1141

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:36S		
Control 1 EC	0.102%	07:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.097%	07:39S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		
Control 3 EC	0.097%	07:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 03/16/2018



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00894	Calib. Date: 12/07/2017
Certification File No.:	00895	Calib. No.: 00034
Linearity File No.:	00896	Cert. Date: 12/07/2017
Solution File No.:	00905	Cert. No.: 00028
Sequential File No.:	00905	Lin. Date: 12/07/2017
		Lin. No.: 00028
		Soln. Date: 02/22/2018
		Soln. No.: 00233
		File Date: 02/22/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 1054

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:55S		
Control 1 EC	0.101%	03:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:57S		
Control 2 EC	0.100%	03:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58S		
Control 3 EC	0.100%	03:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 02/22/2018

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00894	Calib. Date: 12/07/2017
Certification File No.:	00895	Calib. No.: 00034
Linearity File No.:	00896	Cert. Date: 12/07/2017
Solution File No.:	00903	Cert. No.: 00028
Sequential File No.:	00903	Lin. Date: 12/07/2017
		Lin. No.: 00028
		Soln. Date: 01/29/2018
		Soln. No.: 00232
		File Date: 01/29/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0086

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:51S		
Control 1 EC	0.100%	01:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:52S		
Control 2 EC	0.099%	01:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:54S		
Control 3 EC	0.099%	01:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: \_\_\_\_\_



Badge No.: 59

Date: 01/29/2018

