

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01047	Calib. Date: 09/02/2021
Certification File No.:	01048	Calib. No.: 00044
Linearity File No.:	01049	Cert. Date: 09/02/2021
Solution File No.:	01059	Cert. No.: 00037
Sequential File No.:	01059	Lin. Date: 09/02/2021
		Lin. No.: 00037
		Soln. Date: 12/31/2021
		Soln. No.: 00295
		File Date: 12/31/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0076

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:45S		
Control 1 EC	0.102%	16:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		
Control 2 EC	0.102%	16:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	16:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:48S		
Control 3 EC	0.103%	16:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	16:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 12/31/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01047	Calib. Date: 09/02/2021
Certification File No.:	01048	Calib. No.: 00044
Linearity File No.:	01049	Cert. Date: 09/02/2021
Solution File No.:	01057	Cert. No.: 00037
Sequential File No.:	01057	Lin. Date: 09/02/2021
		Lin. No.: 00037
		Soln. Date: 12/09/2021
		Soln. No.: 00294
		File Date: 12/09/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0475

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:37S		
Control 1 EC	0.101%	09:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38S		
Control 2 EC	0.100%	09:39S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40S		
Control 3 EC	0.100%	09:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 12/09/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01047	Calib. Date: 09/02/2021
Certification File No.:	01048	Cert. No.: 00044
Linearity File No.:	01049	Cert. Date: 09/02/2021
Solution File No.:	01054	Lin. Date: 09/02/2021
Sequential File No.:	01054	Lin. No.: 00037
		Soln. Date: 11/15/2021
		Soln. No.: 00293
		File Date: 11/15/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0248

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:28S		
Control 1 EC	0.099%	14:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30S		
Control 2 EC	0.100%	14:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31S		
Control 3 EC	0.100%	14:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 11/15/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01047 Calib. Date: 09/02/2021 Calib. No.: 00044
Certification File No.: 01048 Cert. Date: 09/02/2021 Cert. No.: 00037
Linearity File No.: 01049 Lin. Date: 09/02/2021 Lin. No.: 00037
Solution File No.: 01053 Soln. Date: 10/22/2021 Soln. No.: 00292
Sequential File No.: 01053 File Date: 10/22/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0622

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:35D		
Control 1 EC	0.102%	22:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	22:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:36D		
Control 2 EC	0.102%	22:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	22:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38D		
Control 3 EC	0.102%	22:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	22:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 10/22/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
 Location: MERCHANTVILLE POLICE
 Calibration File No.: 01047 Calib. Date: 09/02/2021 Calib. No.: 00044
 Certification File No.: 01048 Cert. Date: 09/02/2021 Cert. No.: 00037
 Linearity File No.: 01049 Lin. Date: 09/02/2021 Lin. No.: 00037
 Solution File No.: 01052 Soln. Date: 09/22/2021 Soln. No.: 00291
 Sequential File No.: 01052 File Date: 09/22/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
 Control Solution %: 0.100% Expires: 03/17/2023
 Solution Control Lot: 21090 Bottle No.: 0132

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:41D		
Control 1 EC	0.102%	07:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	07:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42D		
Control 2 EC	0.101%	07:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44D		
Control 3 EC	0.101%	07:44D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	07:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN First Name: TOMAS MI: F

Signature:  Badge No.: 61
Date: 09/22/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01034 Calib. Date: 04/19/2021 Calib. No.: 00043
Certification File No.: 01035 Cert. Date: 04/19/2021 Cert. No.: 00036
Linearity File No.: 01036 Lin. Date: 04/19/2021 Lin. No.: 00036
Solution File No.: 01045 Soln. Date: 08/09/2021 Soln. No.: 00289
Sequential File No.: 01045 File Date: 08/09/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0301

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:47D		
Control 1 EC	0.096%	15:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	15:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49D		
Control 2 EC	0.096%	15:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:50D		
Control 3 EC	0.097%	15:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 08/09/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	01034	Calib. Date:	04/19/2021
Certification File No.:	01035	Cert. No.:	00043
Linearity File No.:	01036	Cert. Date:	04/19/2021
Solution File No.:	01044	Lin. Date:	04/19/2021
Sequential File No.:	01044	Lin. No.:	00036
		Soln. Date:	08/09/2021
		Soln. No.:	00288
		File Date:	08/09/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	0849

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:43D		
Control 1 EC	0.098%	07:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44D		
Control 2 EC	0.097%	07:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	07:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46D		
Control 3 EC	0.095%	07:46D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.093%	07:46D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	07:47D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 08/09/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01034 Calib. Date: 04/19/2021 Calib. No.: 00043
Certification File No.: 01035 Cert. Date: 04/19/2021 Cert. No.: 00036
Linearity File No.: 01036 Lin. Date: 04/19/2021 Lin. No.: 00036
Solution File No.: 01041 Soln. Date: 07/22/2021 Soln. No.: 00287
Sequential File No.: 01041 File Date: 07/22/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 1350

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:59D		
Control 1 EC	0.096%	13:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	13:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01D		
Control 2 EC	0.096%	13:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	13:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		
Control 3 EC	0.096%	13:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	13:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tom Whalen

Badge No.: 61

Date: 07/22/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01034 Calib. Date: 04/19/2021 Calib. No.: 00043
Certification File No.: 01035 Cert. Date: 04/19/2021 Cert. No.: 00036
Linearity File No.: 01036 Lin. Date: 04/19/2021 Lin. No.: 00036
Solution File No.: 01040 Soln. Date: 06/29/2021 Soln. No.: 00286
Sequential File No.: 01040 File Date: 06/29/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0954

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:42D		
Control 1 EC	0.099%	20:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:43D		
Control 2 EC	0.098%	20:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44D		
Control 3 EC	0.099%	20:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 06/29/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01034 Calib. Date: 04/19/2021 Calib. No.: 00043
Certification File No.: 01035 Cert. Date: 04/19/2021 Cert. No.: 00036
Linearity File No.: 01036 Lin. Date: 04/19/2021 Lin. No.: 00036
Solution File No.: 01039 Soln. Date: 06/04/2021 Soln. No.: 00285
Sequential File No.: 01039 File Date: 06/04/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0858

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:25D		
Control 1 EC	0.099%	22:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:26D		
Control 2 EC	0.097%	22:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27D		
Control 3 EC	0.098%	22:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tom Cully

Badge No.: 61

Date: 06/04/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01034 Calib. Date: 04/19/2021 Calib. No.: 00043
Certification File No.: 01035 Cert. Date: 04/19/2021 Cert. No.: 00036
Linearity File No.: 01036 Lin. Date: 04/19/2021 Lin. No.: 00036
Solution File No.: 01038 Soln. Date: 05/12/2021 Soln. No.: 00284
Sequential File No.: 01038 File Date: 05/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 1338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:01D		
Control 1 EC	0.100%	23:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02D		
Control 2 EC	0.098%	23:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03D		
Control 3 EC	0.099%	23:04D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:04D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:05D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 05/12/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01020 Calib. Date: 12/08/2020 Calib. No.: 00042
Certification File No.: 01021 Cert. Date: 12/08/2020 Cert. No.: 00035
Linearity File No.: 01022 Lin. Date: 12/08/2020 Lin. No.: 00035
Solution File No.: 01033 Soln. Date: 04/15/2021 Soln. No.: 00282
Sequential File No.: 01033 File Date: 04/15/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 1276

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:54D		
Control 1 EC	0.101%	15:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:55D		
Control 2 EC	0.101%	15:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:56D		
Control 3 EC	0.101%	15:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tom Whalen

Badge No.: 61

Date: 04/15/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	01020	Calib. Date:	12/08/2020
Certification File No.:	01021	Cert. Date:	12/08/2020
Linearity File No.:	01022	Lin. Date:	12/08/2020
Solution File No.:	01030	Soln. Date:	03/17/2021
Sequential File No.:	01030	File Date:	03/17/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0011
Control Solution %:	0.100%			Expires:	09/09/2021
Solution Control Lot:	19240			Bottle No.:	0206

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:14S		
Control 1 EC	0.102%	08:15S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16S		
Control 2 EC	0.100%	08:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		
Control 3 EC	0.101%	08:18S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 03/17/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01020 Calib. Date: 12/08/2020 Calib. No.: 00042
Certification File No.: 01021 Cert. Date: 12/08/2020 Cert. No.: 00035
Linearity File No.: 01022 Lin. Date: 12/08/2020 Lin. No.: 00035
Solution File No.: 01029 Soln. Date: 02/24/2021 Soln. No.: 00280
Sequential File No.: 01029 File Date: 02/24/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 1313

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:02S		
Control 1 EC	0.101%	03:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:04S		
Control 2 EC	0.101%	03:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	03:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:05S		
Control 3 EC	0.103%	03:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	03:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:07S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 02/24/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01020 Calib. Date: 12/08/2020 Calib. No.: 00042
Certification File No.: 01021 Cert. Date: 12/08/2020 Cert. No.: 00035
Linearity File No.: 01022 Lin. Date: 12/08/2020 Lin. No.: 00035
Solution File No.: 01028 Soln. Date: 01/30/2021 Soln. No.: 00279
Sequential File No.: 01028 File Date: 01/30/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0853

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:34S		
Control 1 EC	0.100%	08:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:36S		
Control 2 EC	0.101%	08:36S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37S		
Control 3 EC	0.101%	08:38S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 01/30/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01020	Calib. Date: 12/08/2020
Certification File No.:	01021	Cert. No.: 00042
Linearity File No.:	01022	Cert. Date: 12/08/2020
Solution File No.:	01026	Lin. No.: 00035
Sequential File No.:	01026	Lin. Date: 12/08/2020
		Soln. No.: 00278
		Soln. Date: 01/06/2021
		File Date: 01/06/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 0900

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:14S		
Control 1 EC	0.102%	20:15S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:15S		
Control 2 EC	0.101%	20:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17S		
Control 3 EC	0.101%	20:18S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 01/06/2021