

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01077	Calib. Date: 07/19/2022
Certification File No.:	01078	Calib. No.: 00046
Linearity File No.:	01079	Cert. Date: 07/19/2022
Solution File No.:	01093	Cert. No.: 00039
Sequential File No.:	01093	Lin. Date: 07/19/2022
		Lin. No.: 00039
		Soln. Date: 12/18/2022
		Soln. No.: 00313
		File Date: 12/18/2022

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0011
Control Solution %:	0.100%			Expires:	03/17/2023
Solution Control Lot:	21090			Bottle No.:	1072

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:02S		
Control 1 EC	0.099%	09:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03S		
Control 2 EC	0.099%	09:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05S		
Control 3 EC	0.100%	09:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 12/18/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01077	Calib. Date: 07/19/2022
Certification File No.:	01078	Cert. No.: 00046
Linearity File No.:	01079	Cert. Date: 07/19/2022
Solution File No.:	01092	Lin. No.: 00039
Sequential File No.:	01092	Lin. Date: 07/19/2022
		Soln. No.: 00312
		Soln. Date: 11/29/2022
		File Date: 11/29/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1488

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:05S		
Control 1 EC	0.098%	11:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07S		
Control 2 EC	0.098%	11:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08S		
Control 3 EC	0.099%	11:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 11/29/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01077 Calib. Date: 07/19/2022 Calib. No.: 00046
Certification File No.: 01078 Cert. Date: 07/19/2022 Cert. No.: 00039
Linearity File No.: 01079 Lin. Date: 07/19/2022 Lin. No.: 00039
Solution File No.: 01089 Soln. Date: 11/10/2022 Soln. No.: 00311
Sequential File No.: 01089 File Date: 11/10/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0802

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:47S		
Control 1 EC	0.101%	08:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 2 EC	0.100%	08:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S		
Control 3 EC	0.100%	08:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 11/10/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01077	Calib. Date: 07/19/2022
Certification File No.:	01078	Calib. No.: 00046
Linearity File No.:	01079	Cert. Date: 07/19/2022
Solution File No.:	01088	Cert. No.: 00039
Sequential File No.:	01088	Lin. Date: 07/19/2022
		Lin. No.: 00039
		Soln. Date: 10/18/2022
		Soln. No.: 00310
		File Date: 10/18/2022

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 0777

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:05D		
Control 1 EC	0.101%	21:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:07D		
Control 2 EC	0.099%	21:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:08D		
Control 3 EC	0.099%	21:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 10/18/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01077 Calib. Date: 07/19/2022 Calib. No.: 00046
Certification File No.: 01078 Cert. Date: 07/19/2022 Cert. No.: 00039
Linearity File No.: 01079 Lin. Date: 07/19/2022 Lin. No.: 00039
Solution File No.: 01087 Soln. Date: 09/29/2022 Soln. No.: 00309
Sequential File No.: 01087 File Date: 09/29/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0749

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:39D		
Control 1 EC	0.099%	02:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40D		
Control 2 EC	0.099%	02:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:42D		
Control 3 EC	0.099%	02:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tom Whalen

Badge No.: 61

Date: 09/29/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01077 Calib. Date: 07/19/2022 Calib. No.: 00046
Certification File No.: 01078 Cert. Date: 07/19/2022 Cert. No.: 00039
Linearity File No.: 01079 Lin. Date: 07/19/2022 Lin. No.: 00039
Solution File No.: 01085 Soln. Date: 09/01/2022 Soln. No.: 00308
Sequential File No.: 01085 File Date: 09/01/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0956

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:35D		
Control 1 EC	0.100%	09:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36D		
Control 2 EC	0.100%	09:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 3 EC	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CALLAWAY

First Name: BRIAN

MI: M

Signature: 

Badge No.: 67

Date: 09/01/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01077	Calib. Date: 07/19/2022
Certification File No.:	01078	Calib. No.: 00046
Linearity File No.:	01079	Cert. Date: 07/19/2022
Solution File No.:	01081	Cert. No.: 00039
Sequential File No.:	01081	Lin. Date: 07/19/2022
		Lin. No.: 00039
		Soln. Date: 08/03/2022
		Soln. No.: 00306
		File Date: 08/03/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0478

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:59D		
Control 1 EC	0.099%	09:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01D		
Control 2 EC	0.099%	09:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02D		
Control 3 EC	0.099%	09:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CALLAWAY

First Name: BRIAN

MI: M

Signature: 

Badge No.: 67

Date: 08/03/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01061	Calib. Date: 02/10/2022
Certification File No.:	01062	Calib. No.: 00045
Linearity File No.:	01063	Cert. Date: 02/10/2022
Solution File No.:	01076	Cert. No.: 00038
Sequential File No.:	01076	Lin. Date: 02/10/2022
		Lin. No.: 00038
		Soln. Date: 07/15/2022
		Soln. No.: 00304
		File Date: 07/15/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0135

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:19D		
Control 1 EC	0.099%	08:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:21D		
Control 2 EC	0.098%	08:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:22D		
Control 3 EC	0.099%	08:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 07/15/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01061 Calib. Date: 02/10/2022 Calib. No.: 00045
Certification File No.: 01062 Cert. Date: 02/10/2022 Cert. No.: 00038
Linearity File No.: 01063 Lin. Date: 02/10/2022 Lin. No.: 00038
Solution File No.: 01073 Soln. Date: 06/22/2022 Soln. No.: 00303
Sequential File No.: 01073 File Date: 06/22/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0363

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:15D		
Control 1 EC	0.098%	23:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:17D		
Control 2 EC	0.098%	23:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:18D		
Control 3 EC	0.098%	23:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 06/22/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01061	Calib. Date: 02/10/2022
Certification File No.:	01062	Calib. No.: 00045
Linearity File No.:	01063	Cert. Date: 02/10/2022
Solution File No.:	01071	Cert. No.: 00038
Sequential File No.:	01071	Lin. Date: 02/10/2022
		Lin. No.: 00038
		Soln. Date: 05/26/2022
		Soln. No.: 00302
		File Date: 05/26/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0476

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:41D		
Control 1 EC	0.101%	02:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:42D		
Control 2 EC	0.100%	02:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:44D		
Control 3 EC	0.100%	02:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 05/26/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01061	Calib. Date: 02/10/2022
Certification File No.:	01062	Cert. No.: 00045
Linearity File No.:	01063	Cert. Date: 02/10/2022
Solution File No.:	01070	Lin. No.: 00038
Sequential File No.:	01070	Soln. Date: 05/02/2022
		Soln. No.: 00301
		File Date: 05/02/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0075

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:52D		
Control 1 EC	0.101%	20:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:54D		
Control 2 EC	0.099%	20:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55D		
Control 3 EC	0.100%	20:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 05/02/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01061 Calib. Date: 02/10/2022 Calib. No.: 00045
Certification File No.: 01062 Cert. Date: 02/10/2022 Cert. No.: 00038
Linearity File No.: 01063 Lin. Date: 02/10/2022 Lin. No.: 00038
Solution File No.: 01068 Soln. Date: 04/08/2022 Soln. No.: 00300
Sequential File No.: 01068 File Date: 04/08/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0249

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:27D		
Control 1 EC	0.099%	20:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:28D		
Control 2 EC	0.099%	20:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:30D		
Control 3 EC	0.099%	20:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 04/08/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01061 Calib. Date: 02/10/2022 Calib. No.: 00045
Certification File No.: 01062 Cert. Date: 02/10/2022 Cert. No.: 00038
Linearity File No.: 01063 Lin. Date: 02/10/2022 Lin. No.: 00038
Solution File No.: 01067 Soln. Date: 04/08/2022 Soln. No.: 00299
Sequential File No.: 01067 File Date: 04/08/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0012
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 1244

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:10D		
Control 1 EC	0.084%	19:10D	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.083%	19:10D	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	19:20D		
Control 2 EC	0.090%	19:21D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.089%	19:21D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	19:22D		
Control 3 EC	0.092%	19:23D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.092%	19:23D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	19:24D		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 04/08/2022

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	01061	Calib. Date: 02/10/2022
Certification File No.:	01062	Calib. No.: 00045
Linearity File No.:	01063	Cert. Date: 02/10/2022
Solution File No.:	01066	Cert. No.: 00038
Sequential File No.:	01066	Lin. Date: 02/10/2022
		Lin. No.: 00038
		Soln. Date: 03/08/2022
		Soln. No.: 00298
		File Date: 03/08/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0017

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:48S		
Control 1 EC	0.098%	19:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:49S		
Control 2 EC	0.097%	19:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:50S		
Control 3 EC	0.099%	19:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 03/08/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0073
Location: MERCHANTVILLE POLICE
Calibration File No.: 01047 Calib. Date: 09/02/2021 Calib. No.: 00044
Certification File No.: 01048 Cert. Date: 09/02/2021 Cert. No.: 00037
Linearity File No.: 01049 Lin. Date: 09/02/2021 Lin. No.: 00037
Solution File No.: 01060 Soln. Date: 01/25/2022 Soln. No.: 00296
Sequential File No.: 01060 File Date: 01/25/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWA S3-0011
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0846

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:30S		
Control 1 EC	0.103%	20:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:32S		
Control 2 EC	0.101%	20:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:33S		
Control 3 EC	0.101%	20:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	20:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:35S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: 

Badge No.: 61

Date: 01/25/2022