

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department Merchantville Police Department 1 West Maple Ave Merchantville, NJ 08109	2. Contact Chief Bauer 3. Phone Number 856-662-0500	4. Date 3-21-2012 5. County Camden
6. Alcotest Instrument Serial Number ARUM-0073	7. Simulator Component Serial Number N/A	8. Temperature Probe Component Serial Number N/A

9. Reason for Service.
 The above Alcotest was returned from outside evaluation and/or repair. A successful calibration was conducted utilizing the above Alcotest.

10. Comments
 See attached packing slip and Drager Return and Repair Form.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F Sullivan #5103 <small>Name & Badge Number (Print)</small>	Tpr. I F Sullivan #5103 <small>Signature</small>	3-21-2012 <small>Date</small>
---	---	----------------------------------

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department Merchantville Police Department 1 West Maple Ave Merchantville, NJ 08109	2. Contact Chief Bauer 3. Phone Number: 856-662-0500	4. Date 2-7-2012 5. County Camden
6. Alcotest Instrument Serial Number ARUM-0073	7. Simulator Component Serial Number N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
Unable to connect to the 7110 modem. Can not retrieve the memory via modem.

10. Comments
Instrument is being sent back to Dräger for connectivity issue.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F. Sullivan #5103
 Name & Badge Number (Pnn):

Tpr I F Sullivan #5103
 Signature

2-7-2012
 Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Merchantville P.D., Merchantville, NJ

Date Received: 02/10/12 Date given to service: 02/10/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM - 0073
 Printer Serial#: AR
 SIm Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

- 110V A/C Adapter
- Regulator
- Printer Ribbon
- Printer Paper
- Mouthpieces
- 8510 Stylus
- 8510 Top Cover
- Carrying Case
- Dry Gas
- Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>mp Cal 71</u>	<u>Calibrator</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Replaced Capacitor at location C156 to restore modem connection.

Cal w/ QC / ops check

Service Technician: [Signature] Date: 3-9-12

Packing Slip

Customer no Order No Order date
 150054976 10535732 03/13/2012

Packing slip no. Ship date
 80670698 03/13/2012

Please reference on inquiries

slip to
 150054976
 MERCHANTVILLE BOROUGH POLICE
 CAMDEN COUNTY
 1 WEST MAPLE AVENUE
 MERCHANTVILLE, NJ 08109
 USA

bill to

Your Purchase Order Number
 7110: ARUM-0073
 Any questions? Please contact:
 MARA HERRERA

Customer
 MERCHANTVILLE BOROUGH POLICE
 CAMDEN COUNTY
 1 WEST MAPLE AVENUE
 MERCHANTVILLE, NJ 08109
 USA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Page 1 / 1

Item#	sh. Quant.	Part no Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110: ARUM-0073
0010	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA
0020	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426815
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337