

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARJM 0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00739	Calib. Date: 11/18/2013
Certification File No.:	00740	Calib. No.: 00025
Linearity File No.:	00741	Cert. Date: 11/18/2013
Solution File No.:	00744	Cert. No.: 00019
Sequential File No.:	00744	Lin. Date: 11/18/2013
		Lin. No.: 00019
		Soln. Date: 12/17/2013
		Soln. No.: 00159
		File Date: 12/17/2013

Calibrating Unit: WET	Model No.: CU-34	Serial No.: JDDWA S3-0011
Control Solution %: 0.100%		Expires: 03/12/2015
Solution Control Lot: 13C111		Bottle No.: 0663

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:59S		
Control 1 EC	0.100%	13:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01S		
Control 2 EC	0.097%	13:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02S		
Control 3 EC	0.098%	13:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WIALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Wialen

Badge No.: 61

Date: 12/17/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00739	Calib. Date:	11/18/2013
Certification File No.:	00740	Cert. Date:	11/18/2013
Linearity File No.:	00741	Lin. Date:	11/18/2013
Solution File No.:	00743	Soln. Date:	11/27/2013
Sequential File No.:	00743	File Date:	11/27/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0664

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:42S		
Control 1 EC	0.099%	10:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44S		
Control 2 EC	0.099%	10:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		
Control 3 EC	0.099%	10:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: 1

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 11/27/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00721	Calib. Date:	07/01/2013
Certification File No.:	00722	Cert. No.:	00024
Linearity File No.:	00723	Cert. Date:	07/01/2013
Solution File No.:	00738	Lin. No.:	00018
Sequential File No.:	00738	Soln. Date:	11/14/2013
		Soln. No.:	00156
		File Date:	11/14/2013
Calibrating Unit:	WFT	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3 0012
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0670

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:03S		
Control 1 EC	0.101%	19:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:05S		
Control 2 EC	0.099%	19:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:06S		
Control 3 EC	0.099%	19:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 11/14/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE	Calib. No.:	00024
Calibration File No.:	00721	Cert. Date:	07/01/2013
Certification File No.:	00722	Lin. Date:	07/01/2013
Linearity File No.:	00723	Soln. Date:	10/16/2013
Solution File No.:	00735	File Date:	10/16/2013
Sequential File No.:	00735		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0675

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:01D		
Control 1 EC	0.103%	09:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02D		
Control 2 EC	0.101%	09:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04D		
Control 3 EC	0.100%	09:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WIALEN

First Name: TOMAS

MI: V

Signature: 

Badge No.: 61

Date: 10/16/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00721	Calib. Date:	07/01/2013
Certification File No.:	00722	Cert. No.:	00024
Linearity File No.:	00723	Cert. Date:	07/01/2013
Solution File No.:	00734	Lin. No.:	00018
Sequential File No.:	00734	Lin. Date:	07/01/2013
		Soln. No.:	00154
		Soln. Date:	09/25/2013
		File Date:	09/25/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0667

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:17D		
Control 1 EC	0.100%	12:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18D		
Control 2 EC	0.099%	12:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		
Control 3 EC	0.100%	12:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 09/25/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00721	Calib. Date:	07/01/2013
Certification File No.:	00722	Cert. No.:	00024
Linearity File No.:	00723	Cert. Date:	07/01/2013
Solution File No.:	00730	Lin. No.:	00018
Sequential File No.:	00730	Soln. Date:	09/06/2013
		Soln. No.:	00153
		File Date:	09/06/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0672

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:51D		
Control 1 EC	0.101%	05:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:52D		
Control 2 EC	0.100%	05:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:54D		
Control 3 EC	0.100%	05:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WIALEN

First Name: TOMAS

MI: P

Signature: Sgt. Tomas Wialen

Badge No.: 61

Date: 09/06/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00721	Calib. Date:	07/01/2013
Certification File No.:	00722	Cert. Date:	07/01/2013
Linearity File No.:	00723	Lin. Date:	07/01/2013
Solution File No.:	00725	Soln. Date:	07/21/2013
Sequential File No.:	00725	File Date:	07/21/2013
Calibrating Unit:	WET	Model No.:	CU 34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0674

Function	Result	Time	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:58D		
Control 1 EC	0.101%	07:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00D		
Control 2 EC	0.100%	08:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01D		
Control 3 EC	0.099%	08:02D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:02D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

M: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 07/21/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00700	Calib. Date: 01/17/2013
Certification File No.:	00701	Calib. No.: 00023
Linearity File No.:	00702	Cert. Date: 01/17/2013
Solution File No.:	00719	Cert. No.: 00017
Sequential File No.:	00719	Lin. Date: 01/17/2013
		Lin. No.: 00017
		Soln. Date: 06/14/2013
		Soln. No.: 00149
		File Date: 06/14/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: YDDWA S3-0012
Solution Control Lot:	13C111	Expires: 03/12/2015
		Bottle No.: 0668

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:35D		
Control 1 EC	0.101%	00:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	00:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:37D		
Control 2 EC	0.101%	00:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:38D		
Control 3 EC	0.100%	00:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 06/14/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.	00700	Calib. Date:	01/17/2013
Certification File No.:	00701	Cert. Date:	01/17/2013
Linearity File No.:	00702	Lin. Date:	01/17/2013
Solution File No.:	00718	Soln. Date:	05/26/2013
Sequential File No.:	00718	File Date:	05/26/2013
Calibrating Unit:	WEI	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0799

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:01D		
Control 1 EC	0.102%	21:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02D		
Control 2 EC	0.100%	21:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		
Control 3 EC	0.101%	21:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MT: 1:

Signature: Sgt Tomas Whalen

Badge No.: 61

Date: 05/26/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00700	Calib. No.: 00023
Certification File No.:	00701	Cert. No.: 00017
Linearity File No.:	00702	Lin. No.: 00017
Solution File No.:	00715	Soln. No.: 00147
Sequential File No.:	00715	File Date: 05/08/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	11K096	Expires: 11/23/2013
		Bottle No.: 0800

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:55D		
Control 1 EC	0.100%	04:55D	34.0°C	*** TEST PASSED ***
Control 1 TR	0.100%	04:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:56D		
Control 2 EC	0.100%	04:57D	34.0°C	*** TEST PASSED ***
Control 2 TR	0.099%	04:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:57D		
Control 3 EC	0.100%	04:58D	34.0°C	*** TEST PASSED ***
Control 3 TR	0.100%	04:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHATEN

First Name: TOMAS

MI: P

Signature: _____

Sgt. Tomas Krill

Badge No.: 61

Date: 05/08/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcoest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00700	Calib. Date: 01/17/2013
Certification File No.:	00701	Calib. No.: 00023
Linearity File No.:	00702	Cert. Date: 01/17/2013
Solution File No.:	00712	Cert. No.: 00017
Sequential File No.:	00712	Lin. Date: 01/17/2013
		Lin. No.: 00017
		Soln. Date: 04/18/2013
		Soln. No.: 00146
		File Date: 04/18/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	11K096	Expires: 11/23/2013
		Bottle No.: 0795

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:57D		
Control 1 EC	0.103%	13:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:59D		
Control 2 EC	0.101%	14:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00D		
Control 3 EC	0.101%	14:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcoest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

M1: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 04/18/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM 0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00700	Calib. Date: 01/17/2013
Certification File No.:	00701	Cert. No.: 00023
Linearity File No.:	00702	Cert. No.: 00017
Solution File No.:	00711	Lin. No.: 00017
Sequential File No.:	00711	Soln. No.: 00145
		File Date: 04/18/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DLJWA S3-0012
Solution Control Lot:	11K096	Expires: 11/23/2013
		Bottle No.: 0780

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:51D		
Control 1 EC	-.---%	07:51D	24.7°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	07:51D	24.7°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHATLEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Wilh

Badge No.: 61

Date: 04/18/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00700	Calib. Date: 01/17/2013
Certification File No.:	00701	Calib. No.: 00023
Linearity File No.:	00702	Cert. Date: 01/17/2013
Solution File No.:	00710	Cert. No.: 00017
Sequential File No.:	00710	Lin. Date: 01/17/2013
		Lin. No.: 00017
		Soln. Date: 03/29/2013
		Soln. No.: 00144
		File Date: 03/29/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWA S3-0012
Control Solution %:	0.100%		Expires: 11/23/2013
Solution Control Lot:	TK096		Route No.: 0783

Function	Result	Time	Temperature	Comments)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:12S		
Control 1 EC	0.102%	07:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:13S		
Control 2 EC	0.099%	07:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:15S		
Control 3 EC	0.099%	07:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: _____



Badge No.: 61

Date: 03/29/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00700	Calib. Date:	01/17/2013
Certification File No.:	00701	Cert. Date:	01/17/2013
Linearity File No.:	00702	Lin. Date:	01/17/2013
Solution File No.:	00709	Soln. Date:	03/06/2013
Sequential File No.:	00709	File Date:	03/06/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0012
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0779

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:06S		
Control 1 EC	0.102%	17:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:07S		
Control 2 EC	0.099%	17:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:09S		
Control 3 EC	0.099%	17:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WHALEN

First Name: TOMAS

M: F

Signature: _____



Badge No.: 61

Date: 03/06/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00700	Calib. Date: 01/17/2013
Certification File No.:	00701	Calib. No.: 00023
Linearity File No.:	00702	Cert. Date: 01/17/2013
Solution File No.:	00707	Cert. No.: 00017
Sequential File No.:	00707	Lin. Date: 01/17/2013
		Lin. No.: 00017
		Soln. Date: 02/17/2013
		Soln. No.: 00142
		File Date: 02/17/2013

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0012
Solution Control Lot:	TK096	Expires: 11/23/2013
		Bottle No.: 0789

Function	Result	Time	Temperature	Comment(s)
	% BAC	HH:MM	Simulator (°C)	or Errors)
Ambient Air Blank	0.000%	03:24S		
Control 1 FC	0.100%	03:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:26S		
Control 2 FC	0.100%	03:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:27S		
Control 3 FC	0.099%	03:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MJE

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 02/17/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00701	Calib. Date: 01/17/2013
Certification File No.:	00701	Calib. No.: 00023
Linearity File No.:	00702	Cert. Date: 01/17/2013
Solution File No.:	00705	Cert. No.: 00017
Sequential File No.:	00705	Lin. Date: 01/17/2013
		Lin. No.: 00017
		Soln. Date: 02/07/2013
		Soln. No.: 00141
		File Date: 02/07/2013
Calibrating Unit:	WLT	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3 0012
Solution Control Lot:	11K096	Expires: 11/23/2013
		Bottle No.: 0784

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:37S		
Control 1 EC	0.101%	05:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:39S		
Control 2 EC	0.099%	05:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:40S		
Control 3 EC	0.101%	05:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 02/07/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE		
Calibration File No.:	00676	Calib. Date:	09/10/2012
Certification File No.:	00677	Cert. Date:	09/10/2012
Linearity File No.:	00678	Lin. Date:	09/10/2012
Solution File No.:	00699	Soln. Date:	01/16/2013
Sequential File No.:	00699	File Date:	01/16/2013
Calibrating Unit:	WET	Model No.:	CU 34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0794

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:44S		
Control 1 EC	0.098%	22:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45S		
Control 2 EC	0.098%	22:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:47S		
Control 3 EC	0.098%	22:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:48S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

Badge No.: 59

Date: 01/16/2013