

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00676	Calib. Date: 09/10/2012
Certification File No.:	00677	Calib. No.: 00022
Linearity File No.:	00678	Cert. Date: 09/10/2012
Solution File No.:	00696	Cert. No.: 00016
Sequential File No.:	00696	Lin. Date: 09/10/2012
		Lin. No.: 00016
		Soln. Date: 12/26/2012
		Soln. No.: 00138
		File Date: 12/26/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	11K096	Expires: 11/23/2013
		Bottle No.: 0796

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:36S		
Control 1 EC	0.100%	08:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:38S		
Control 2 EC	0.098%	08:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39S		
Control 3 EC	0.099%	08:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 12/26/2012



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII C Serial No.: ARUM 0073  
**Location:** MERCHANTVILLE POLICE  
**Calibration File No.:** 00676 **Calib. Date:** 09/10/2012 **Calib. No.:** 00022  
**Certification File No.:** 00677 **Cert. Date:** 09/10/2012 **Cert. No.:** 00016  
**Linearity File No.:** 00678 **Lin. Date:** 09/10/2012 **Lin. No.:** 00016  
**Solution File No.:** 00690 **Soln. Date:** 11/01/2012 **Soln. No.:** 00136  
**Sequential File No.:** 00690 **File Date:** 11/01/2012

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWA S3 0011  
**Control Solution %:** 0.100% **Expires:** 11/23/2013  
**Solution Control Lot:** 11K096 **Bottle No.:** 0797

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:44S		
Control 1 EC	0.101%	20:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:45S		
Control 2 EC	0.099%	20:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:46S		
Control 3 EC	0.098%	20:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Kulis

Badge No.: 61

Date: 11/01/2012

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
<b>Location:</b>	MERCHANTVILLE POLICE		
<b>Calibration File No.:</b>	00676	<b>Calib. Date:</b>	09/10/2012
<b>Certification File No.:</b>	00677	<b>Cert. No.:</b>	00022
<b>Linearity File No.:</b>	00678	<b>Cert. Date:</b>	09/10/2012
<b>Solution File No.:</b>	00686	<b>Lin. No.:</b>	00016
<b>Sequential File No.:</b>	00686	<b>Lin. Date:</b>	09/10/2012
		<b>Soln. No.:</b>	00135
		<b>Soln. Date:</b>	10/07/2012
		<b>File Date:</b>	10/07/2012
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDWA S3 0011
<b>Solution Control Lot:</b>	11K096	<b>Expires:</b>	11/23/2013
		<b>Bottle No.:</b>	0776

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:59D		
Control 1 EC	0.100%	03:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:01D		
Control 2 EC	0.099%	03:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:02D		
Control 3 EC	0.099%	03:03D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:03D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: 

59

Badge No.: 59

Date: 10/07/2012

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00635	Calib. Date: 03/21/2012
Certification File No.:	00636	Calib. No.: 00021
Linearity File No.:	00637	Cert. Date: 03/21/2012
Solution File No.:	00665	Cert. No.: 00015
Sequential File No.:	00665	Lin. Date: 03/21/2012
		Lin. No.: 00015
		Soln. Date: 08/23/2012
		Soln. No.: 00132
		File Date: 08/23/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	11K096	Expires: 11/23/2013
		Bottle No.: 0778

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:56D		
Control 1 EC	0.099%	10:57D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:57D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		
Control 2 EC	0.099%	10:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 3 EC	0.098%	11:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WIALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_

*Sgt. Tomas Wialen*

Badge No.: 61

Date: 08/23/2012



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM 0073
<b>Location:</b>	MERCHANTVILLE POLICE	
<b>Calibration File No.:</b>	00635	<b>Calib. Date:</b> 03/21/2012
<b>Certification File No.:</b>	00636	<b>Calib. No.:</b> 00021
<b>Linearity File No.:</b>	00637	<b>Cert. Date:</b> 03/21/2012
<b>Solution File No.:</b>	00656	<b>Cert. No.:</b> 00015
<b>Sequential File No.:</b>	00656	<b>Lin. Date:</b> 03/21/2012
		<b>Lin. No.:</b> 00015
		<b>Soln. Date:</b> 07/26/2012
		<b>Soln. No.:</b> 00130
		<b>File Date:</b> 07/26/2012
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDWA S3-0011
<b>Solution Control Lot:</b>	11K096	<b>Expires:</b> 11/23/2013
		<b>Bottle No.:</b> 0787

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:00D		
Control 1 EC	0.099%	20:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01D		
Control 2 EC	0.098%	20:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:03D		
Control 3 EC	0.099%	20:03D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:03D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 07/26/2012

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCRIANTVILLE POLICE	Calib. No.:	00021
Calibration File No.:	00635	Calib. Date:	03/21/2012
Certification File No.:	00636	Cert. Date:	03/21/2012
Linearity File No.:	00637	Lin. Date:	03/21/2012
Solution File No.:	00652	Soln. Date:	06/30/2012
Sequential File No.:	00652	File Date:	06/30/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0782

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:33D		
Control 1 EC	0.099%	03:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:34D		
Control 2 EC	0.099%	03:35D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:35D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:36D		
Control 3 EC	0.098%	03:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TALARICO

First Name: FRANK

MI: D

Signature: Frank Talarico

Badge No.: 59

Date: 06/30/2012





# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
Location:	MERCHANTVILLE POLICE	
Calibration File No.:	00635	Calib. Date: 03/21/2012
Certification File No.:	00636	Calib. No.: 00021
Linearity File No.:	00637	Cert. Date: 03/21/2012
Solution File No.:	00650	Cert. No.: 00015
Sequential File No.:	00650	Lin. Date: 03/21/2012
		Lin. No.: 00015
		Soln. Date: 06/19/2012
		Soln. No.: 00127
		File Date: 06/19/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DIDWA S3 0011
Solution Control Lot:	10H081	Expires: 08/04/2012
		Bottle No.: 0944

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:49D		
Control 1 EC	0.098%	08:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50D		
Control 2 EC	0.098%	08:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		
Control 3 EC	0.098%	08:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 06/19/2012

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0073
Location:	MERCHANTVILLE POLICE	Calib. No.:	00021
Calibration File No.:	00635	Calib. Date:	03/21/2012
Certification File No.:	00636	Cert. Date:	03/21/2012
Linearity File No.:	00637	Lin. Date:	03/21/2012
Solution File No.:	00645	Soln. Date:	04/30/2012
Sequential File No.:	00645	File Date:	04/30/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	10H081	Expires:	08/04/2012
		Bottle No.:	0945

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:01D		
Control 1 EC	0.098%	10:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		
Control 2 EC	0.096%	10:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		
Control 3 EC	0.097%	10:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Kull

Badge No.: 61

Date: 04/30/2012

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII C	Serial No.:	ARUM-0073
Location:	MERCILANTVILLE POLICE	Calib. No.:	00021
Calibration File No.:	00635	Cert. Date:	03/21/2012
Certification File No.:	00636	Lin. Date:	03/21/2012
Linearity File No.:	00637	Soln. Date:	04/30/2012
Solution File No.:	00644	File Date:	04/30/2012
Sequential File No.:	00644		
Calibrating Unit:	WET	Model No.:	CU 34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0011
Solution Control Lot:	10H081	Expires:	08/04/2012
		Bottle No.:	0930

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:31D		
Control 1 EC	----%	08:32D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	----%	08:32D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	----%	----		
Control 2 EC	----%	----	N/A	
Control 2 IR	----%	----	N/A	
Ambient Air Blank	----%	----		
Control 3 EC	----%	----	N/A	
Control 3 IR	----%	----	N/A	
Ambient Air Blank	----%	----		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 04/30/2012

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
<b>Location:</b>	MERCHANTVILLE POLICE	
Calibration File No.:	00635	Calib. Date: 03/21/2012
Certification File No.:	00636	Calib. No.: 00021
Linearity File No.:	00637	Cert. Date: 03/21/2012
Solution File No.:	00641	Cert. No.: 00015
Sequential File No.:	00641	Lin. Date: 03/21/2012
		Lin. No.: 00015
		Soln. Date: 04/14/2012
		Soln. No.: 00123
		File Date: 04/14/2012
<b>Calibrating Unit:</b>	WE1	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0011
Solution Control Lot:	10H081	Expires: 08/04/2012
		Bottle No.: 0941

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:57D		
Control 1 EC	0.099%	19:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:59D		
Control 2 EC	0.098%	19:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00D		
Control 3 EC	0.097%	20:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WIALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 04/14/2012

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0073
<b>Location:</b>	MERCHANTVILLE POLICE	
<b>Calibration File No.:</b>	00607	<b>Calib. Date:</b> 09/28/2011
<b>Certification File No.:</b>	00608	<b>Calib. No.:</b> 00018
<b>Linearity File No.:</b>	00609	<b>Cert. Date:</b> 09/28/2011
<b>Solution File No.:</b>	00630	<b>Cert. No.:</b> 00014
<b>Sequential File No.:</b>	00630	<b>Lin. Date:</b> 09/28/2011
		<b>Lin. No.:</b> 00014
		<b>Soln. Date:</b> 01/12/2012
		<b>Soln. No.:</b> 00120
		<b>File Date:</b> 01/12/2012
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b> DDWA S3-0012
<b>Solution Control Lot:</b>	1011081	<b>Expires:</b> 08/04/2012
		<b>Bottle No.:</b> 0934

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:58S		
Control 1 EC	0.099%	18:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:59S		
Control 2 EC	0.099%	19:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:01S		
Control 3 EC	0.099%	19:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: \_\_\_\_\_



Badge No.: 61

Date: 01/12/2012

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM 0073
<b>Location:</b>	MERCHANTVILLE POLICE		
<b>Calibration File No.:</b>	00607	<b>Calib. Date:</b>	09/28/2011
<b>Certification File No.:</b>	00608	<b>Cert. No.:</b>	00014
<b>Linearity File No.:</b>	00609	<b>Lin. No.:</b>	00014
<b>Solution File No.:</b>	00629	<b>Soln. No.:</b>	00119
<b>Sequential File No.:</b>	00629	<b>File Date:</b>	01/12/2012
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDWA S3-0012
<b>Solution Control Lot:</b>	1011081	<b>Expires:</b>	08/04/2012
		<b>Bottle No.:</b>	0946

Function	Result	Time	Temperature	Comments(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:35S		
Control 1 EC	0.093%	17:36S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.093%	17:36S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:36S		
Control 2 EC	0.092%	17:37S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.093%	17:37S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:38S		
Control 3 EC	0.093%	17:39S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.095%	17:39S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:39S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHALEN

First Name: TOMAS

MI: F

Signature: Sgt. Tomas Whalen

Badge No.: 61

Date: 01/12/2012