Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-1-2012 tuning fork Serial No. 392537
was tested and found to oscillate at 2019 cycles per second. Such
oscillation causes a doppler radar operating in the KA band to read 20 mph.
When operated over the temperature of -22°F to +140°F no correction is required

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-184B Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-1-2012 tuning fork Serial No. 392535
was tested and found to oscillate at 2020 cycles per second. Such
oscillation causes a doppler radar operating in the KA band to read 20 mph.
When operated over the temperature of –22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-184B Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-1-2012 tuning fork Serial No. 392560
was tested and found to oscillate at 2020 cycles per second. Such
oscillation causes a doppler radar operating in the KA band to read 20 mph.
When operated over the temperature of -22°F to +140°F no correction is required

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-028 MPD-184B Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-1-2012 tuning fork Serial No. 392524
was tested and found to oscillate at 2019 cycles per second. Such
oscillation causes a doppler radar operating in the KA band to read 20 mph.
When operated over the temperature of –22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-184B Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-1-2012 tuning fork Serial No. 392541 was tested and found to oscillate at 2018 cycles per second. Such oscillation causes a doppler radar operating in the KA band to read 20 mph. When operated over the temperature of –22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-1840 Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-8-2012 tuning fork Serial No. 392765
was tested and found to oscillate at 5048 cycles per second. Such
oscillation causes a doppler radar operating in the KA band to read 50 mph.
When operated over the temperature of −22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-184B Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-2-2012 tuning fork Serial No. 392594 was tested and found to oscillate at 5046 cycles per second. Such oscillation causes a doppler radar operating in the KA band to read 50 mph. When operated over the temperature of –22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-164B Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 7-31-2012 tuning fork Serial No. 392501 was tested and found to oscillate at 5045 cycles per second. Such oscillation causes a doppler radar operating in the KA band to read 50 mph. When operated over the temperature of -22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-184B Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-7-2012 tuning fork Serial No. 392758
was tested and found to oscillate at 5047 cycles per second. Such
oscillation causes a doppler radar operating in the KA band to read 50 mph.
When operated over the temperature of –22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-028 MPD-1848 Rev. 05/12

Tuning Fork
CERTIFICATE OF ACCURACY

This is to certify that on 8-19-2012 tuning fork Serial No. 392829 was tested and found to oscillate at 5045 cycles per second. Such oscillation causes a doppler radar operating in the KA band to read 50 mph. When operated over the temperature of -22°F to +140°F no correction is required.

MPH

316 East Ninth Street / Owensboro, KY 42303

Technician

GC-026 MPD-184B Rev. 05/12