



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

JON S. CORZINE
Governor

ANNE MILGRAM
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 06/28/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 97B046

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1222 to 0.1228 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 24, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajt R. Tungare (Handwritten signature)

Ajt R. Tungare
Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 13th day of July, 2007.

Linda L. DeSantis (Handwritten signature)
Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009



New Jersey is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





Draeger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
 Model: MARK IIA
 Other: _____

Serial Number:

DDUK S3-0133

Certification Date

JUL 25 2007

Technician

DB

Re-Certification Due Date

JUL 25 2008

Drägersafety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDUNP2-316

07-25-2007

07-25-2008

Probe Value

101

Dräger Safety Diagnostics, Inc.
Technical Service Department



Dräger safety

Alcotest[®] 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48654, 49 FR 48664 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

SERIAL NUMBER:

11 APR 05

APK-0039

Dräger Safety Diagnostics, Inc.
Durango, CO

[Signature]

Alcotest 7110 Calibration Record

Equipment

Alcotest 7110 MKIII-C
Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 00415
Calib. Date: 11/28/2007
Calib. No.: 00007
Certification File No.: 00309
Cert. Date: 04/25/2007
Cert. No.: 00004
Linearity File No.: 00310
Lin. Date: 04/25/2007
Lin. No.: 00003
Solution File No.: 00408
Soln. Date: 11/20/2007
Soln. No.: 00048
Sequential File No.: 00415
File Date: 11/28/2007

Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDUK S3-0133
Control Solution %: 0.100%
Expires: 02/22/2009
Solution Control Lot: 07B045
Bottle No.: 0083

Coordinator

Last Name: CROSS
First Name: WILLIAM
MI: M
Signature: TPR. I WILL M. CROSS #5342
Badge No.: 5342
Date: 11/28/2007

*Black Key Temperature Probe Serial# DDWAP2-001 WNR

*Ertco-Hart Digital Temperature Measuring System Serial# A23725 WNR

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 00415 Calib. Date: 11/28/2007 Calib. No.: 00007
Certification File No.: 00416 Cert. Date: 11/28/2007 Cert. No.: 00005
Linearity File No.: 00310 Lin. Date: 04/25/2007 Lin. No.: 00003
Solution File No.: 00408 Soln. Date: 11/20/2007 Soln. No.: 00048
Sequential File No.: 00416 File Date: 11/28/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 02/22/2009
Solution Control Lot: 07B045 Bottle No.: 0083

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:03S		
Control 1 EC	0.099%	09:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05S		
Control 2 EC	0.099%	09:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06S		
Control 3 EC	0.099%	09:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		

All tests within acceptable tolerance.

Coordinator

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: TPR I. William Cross #5342

Badge No.: 5342

Date: 11/28/2007

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 00415 Calib. Date: 11/28/2007 Calib. No.: 00007
Certification File No.: 00416 Cert. Date: 11/28/2007 Cert. No.: 00005
Linearity File No.: 00417 Lin. Date: 11/28/2007 Lin. No.: 00004
Solution File No.: 00408 Soln. Date: 11/20/2007 Soln. No.: 00048
Sequential File No.: 00417 File Date: 11/28/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDMK S3-0003
Control Solution %: 0.040% Expires: 01/26/2009
Solution Control Lot: 07A041 Bottle No.: 0045

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDRK S3-0005
Control Solution %: 0.080% Expires: 01/26/2009
Solution Control Lot: 07A042 Bottle No.: 0090

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDRF S3-0009
Control Solution %: 0.160% Expires: 01/26/2009
Solution Control Lot: 07A043 Bottle No.: 0019

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:35S		
Control 1 EC	0.040%	09:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.040%	09:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37S		
Control 2 EC	0.040%	09:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.039%	09:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39S		
Control 3 EC	0.081%	09:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.080%	09:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41S		
Control 4 EC	0.080%	09:42S	34.0°C	*** TEST PASSED ***
Control 4 IR	0.079%	09:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43S		
Control 5 EC	0.159%	09:44S	34.0°C	*** TEST PASSED ***
Control 5 IR	0.158%	09:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45S		
Control 6 EC	0.158%	09:46S	34.0°C	*** TEST PASSED ***
Control 6 IR	0.159%	09:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47S		

All tests within acceptable tolerance.

Coordinator

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: _____

TPR. I WILLIAM P. #5342

Badge No.: 5342

Date: 11/28/2007



JON S. CORZINE
Governor

State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

STUART RABNER
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 04/04/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07B045

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1222 to 0.1227 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 22, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 20th day of April, 2007.

Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009



New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 West Trenton NJ 08628-0068
 (609) 882-2000

RON B. CORZINE
 Governor

STUART RAHNER
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.040 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. **ANALYSIS DATE:** 03/06/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07A041

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0483 to 0.0489 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 26, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 29th day of March, 2007.

Linda L. DeSantis
 Notary

Linda L. DeSantis
 My Commission
 Expires Aug. 17, 2009





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7058
 West Trenton NJ 08628-0058
 (609) 882-2000

STUART RABNER
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

RON S. CORZINE
Governor

CERTIFICATION OF ANALYSIS
0.080 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 03/06/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07A042

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0961 to 0.0972 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 25, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungarc

Ajit R. Tungarc
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 29th day of March, 2007.

Linda L. DeSantis
 Notary

Linda L. DeSantis
 My Commission
 Expires Aug. 17, 2009



New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

STUART RAIBER
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

JOHN S. CORZINE
 Governor

CERTIFICATION OF ANALYSIS
0.160 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. **ANALYSIS DATE:** 03/06/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: D7A043

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1932 to 0.1938 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 26, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Linda L DeSantis
 My Commission
 Expires Aug. 17, 2008

Ajit R. Tungare
 Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 29th day of March, 2007.

Linda L. DeSantis
 Notary



New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable



Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 00415 Calib. Date: 11/28/2007 Calib. No.: 00007
Certification File No.: 00416 Cert. Date: 11/28/2007 Cert. No.: 00005
Linearity File No.: 00417 Lin. Date: 11/28/2007 Lin. No.: 00004
Solution File No.: 00418 Soln. Date: 11/28/2007 Soln. No.: 00049
Sequential File No.: 00418 File Date: 11/28/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133
Control Solution %: 0.100% Expires: 04/24/2009
Solution Control Lot: 07B046 Bottle No.: 0685

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:53S		
Control 1 EC	0.099%	10:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54S		
Control 2 EC	0.099%	10:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 3 EC	0.099%	10:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP2-316 WOL

Changed By:

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: *TPR. I WILL M. Cross #5342*

Badge No.: 5342

Date: 11/28/2007



Draegersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDMK53-0003

Certification Date

FEB 19 2007

Technician

Re-Certification Due Date

FEB 19 2008



Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST[®] CU34
 Model: MARK IIA
 Other: _____

Serial Number:

DDRK 23-0005


Certification Date

FEB 19 2007

Technician

Re-Certification Due Date

FEB 19 2008



Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST[®] CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDRF53-8009

Certification Date

FEB 19 2007

Technician

Re-Certification Due Date

FEB 19 2008



Drägersafety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDWAP2-001

Certification date:

02/19/2007

Next Certification due:

02/19/2008

Probe Value

101

Draeger Safety Diagnostics, Inc.
Technical Service Department

CRD

Drägersafety

Ertco-Hart Digital Temperature Measuring System

REPORT OF CALIBRATION

This is to certify that the Ertco-Hart Digital Temperature Measuring System has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). Draeger Safety Diagnostics, Inc. (DSDI) recommends accuracy verification of the Ertco-Hart Digital Temperature Measuring System within 12 months of the certification date below, or sooner, according to your state specification.

DSDI equipment used for temperature verification Serial Number: HH41 303176

Digital Units Serial Number: A23725

Probe Serial Number: 583984

Certification Date: 02/19/2007

Next Certification Due: 02/19/2008

At 34.00 °C digital unit displays 34.00 °C

Draeger Safety Diagnostics, Inc. Technician: CRD

DEPARTMENT OF
Water and Public Safety
This is to certify that

William M. Cross

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 248 OF THE LAWS OF 1986 IN THE OPERATION OF THE Breathalyzer METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 31st DAY OF Jan.
ONE THOUSAND NINE HUNDRED AND 97

Carl A. Williams
SUPERINTENDENT
NEW JERSEY STATE POLICE

Peter ...
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Water and Public Safety
This is to certify that

William M. Cross
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 248 OF THE LAWS OF 1986 IN THE OPERATION OF THE Alcotest 7110 MKIII-C METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF February
ONE THOUSAND AND 05

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	2-11-99	GLDA	Robert ...
2.	10-10-01	BCDA	L. ...
3.	4-17-03	BCDA	Mike ...
4.	6-22-05	ACTC	C. ...
5.	7-27-07	OLPA	A. ...
6.			
7.			
8.			
9.			

SP-2938

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev 12/04)



CERTIFICATE

This is to certify that


William M. Cross
Trooper II #5342

has successfully completed the two day Draeger Safety Diagnostics, Inc. Operator Training and Preventive Maintenance Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and
Maintenance Technician**

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: October 28, 2004



Instructor: Hansueli Ryser

Draegersafety

William M. Cross #5342
Department of Law and Public Safety
New Jersey State Police
P.O. Box 7068-River Road
West Trenton, NJ 08628
609-882-2000
lpp5342@gw.njsp.org

Enlistment
Date

December 23, 1995

Career
Experience

10 years, 9 months

May 2003 to present

Field Operations Section
Alcohol/Drug Test Unit

December 2001 to May 2003

Field Operations Section
Tactical Patrol Unit #1
Troop D

March 1999 to December 2001

Field Operations Section
Moorestown Barracks
Troop D

November 1997 to March 1999

Field Operations Section
Bellmawr Barracks
Troop A

October 1997 to November 1997

Field Operations Section
Red Lion Barracks
Troop A

January 1997 to October 1997

Field Operations Section
Tuckerton Barracks
Troop A

June 1996 to January 1997

Field Operations Section
Bridgeton Barracks
Troop A

December 1995 to June 1996

Field operations Section
Red Lion Barracks
Troop A

July 1995 to December 1995

New Jersey State Police Academy

Education

- May 1992 Bachelor of Arts degree, Political Science,
Kutztown University, Kutztown, PA
- June 1988 Graduated Haddon Township High School,
Haddon Township, NJ

Technical/ Professional Training

- February 2005 Alcotest® 7110 MKIII-C Certified
- October 2004 Draeger Safety Diagnostics, Inc. Operator Training
and Preventative Maintenance Course on the
New Jersey specific Alcotest® 7110 MKIII-C
- January 2004 Instructor Training in DWI Detection and
Standardized Field Sobriety Testing
- January 2004 Certified in Standardized Field Sobriety
Tests including Horizontal Gaze Nystagmus
- September 2003 Certified Breath Test Coordinator/Instructor
- August 2003 National Draeger, Inc. Breathalyzer® 900/900A
Maintenance Course
- July 2003 Methods of Instruction Course
- June 2003 Driving While Intoxicated/Standardized Field
Sobriety Testing Course
- May 2003 The Robert F. Borkenstein Course on Alcohol,
Drugs, and Highway Safety
- May 2003 Gang Awareness Course
- March 2000 Spanish for Law Enforcement Course
- March 1999 Traffic Law Enforcement Course
- June 1998 Accident Investigation I Course
- February 1997 Tracktest Fifth Wheel Certified
- January 1997 5-Day Breathalyzer® 900/900A Operator Certified
- February 1996 K-55 Radar Operator Certified

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

Ronald M. Morello
 Stratford Boro

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 17-27 OF
 THE LAWS OF NEW JERSEY IN OPERATION OF THE **ALCOTEST 7118 MK IIC**
 A METHOD TO DETERMINE BAC/ALCOHOL
 GIVEN UNDER MY HANDS AT TRENTON NEW JERSEY THIS 1st DAY OF February

TWO THOUSAND AND FIVE

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	1-16-07	CPA	C. P. [Signature]
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

DEPARTMENT OF
Traffic and Public Safety

This is to certify that

Stephen J. McBride
Stratford Boro.

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 143 OF
 THE LAWS OF 1964 IN THE OPERATION OF THE **Alcotest 7110 MKIII-C**

A METHOD TO DETERMINE INTOXICATION
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 14th DAY OF February

TWO THOUSAND AND Five

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course	
	PLACE	INSTRUCTOR
1. 2-2-07	GCPA	WM Koon
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

*Advise subject:

"I want you to take a deep breath and blow into the mouthpiece with one long continuous breath.

Continue to blow until I tell you to stop.

Do you understand these instructions?"

**NO
PORTABLE
RADIOS ON
DURING
BREATH
TESTS**

**NO PORTABLE
RADIOS OR
CELLULAR
TELEPHONES IN
THIS AREA OF THE
PATROL ROOM
DURING BREATH
TESTING
KEEP PORTABLE
RADIOS OFF
DURING TESTING**

*Old sign
1/1/83*

*Advise subject:

"I want you to take a deep breath and blow into the mouthpiece with one long continuous breath."

Continue to blow until I tell you to stop.

Do you understand these instructions?"

1