

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.
Calibration File No.: 00539 Calib. Date: 09/15/2008 Calib. No.: 00009
Certification File No.: 00540 Cert. Date: 09/15/2008 Cert. No.: 00007
Linearity File No.: 00541 Lin. Date: 09/15/2008 Lin. No.: 00006
Solution File No.: 00597 Soln. Date: 12/31/2008 Soln. No.: 00069
Sequential File No.: 00597 File Date: 12/31/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163
Control Solution %: 0.100% Expires: 06/09/2010
Solution Control Lot: 08F053 Bottle No.: 0272

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:42S		
Control 1 EC	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		
Control 2 EC	0.099%	11:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44S		
Control 3 EC	0.100%	11:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  63203

Badge No.: 3200

Date: 12/31/2008

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00539	Calib. Date: 09/15/2008
Certification File No.:	00540	Calib. No.: 00009
Linearity File No.:	00541	Cert. Date: 09/15/2008
Solution File No.:	00579	Cert. No.: 00007
Sequential File No.:	00579	Lin. Date: 09/15/2008
		Lin. No.: 00006
		Soln. Date: 12/06/2008
		Soln. No.: 00068
		File Date: 12/06/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0274

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:30S		
Control 1 EC	0.103%	12:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32S		
Control 2 EC	0.101%	12:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:33S		
Control 3 EC	0.103%	12:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	12:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200

Date: 12/06/2008

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.	
Calibration File No.: 00539	Calib. Date: 09/15/2008
Certification File No.: 00540	Cert. No.: 00009
Linearity File No.: 00541	Cert. Date: 09/15/2008
Solution File No.: 00558	Lin. No.: 00007
Sequential File No.: 00558	Lin. Date: 09/15/2008
	Soln. No.: 00006
	Soln. Date: 11/01/2008
	File Date: 11/01/2008

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 04/24/2009
Solution Control Lot: 07B046		Bottle No.: 0676

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:32S		
Control 1 EC	0.100%	13:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33S		
Control 2 EC	0.100%	13:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34S		
Control 3 EC	0.101%	13:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

63204

Badge No.: 3200

Date: 11/01/2008

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00539	Calib. Date: 09/15/2008
Certification File No.:	00540	Calib. No.: 00009
Linearity File No.:	00541	Cert. Date: 09/15/2008
Solution File No.:	00548	Cert. No.: 00007
Sequential File No.:	00548	Lin. Date: 09/15/2008
		Lin. No.: 00006
		Soln. Date: 10/07/2008
		Soln. No.: 00065
		File Date: 10/07/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	07B046	Expires: 04/24/2009
		Bottle No.: 0689

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:24D		
Control 1 EC	0.101%	15:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25D		
Control 2 EC	0.101%	15:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27D		
Control 3 EC	0.100%	15:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: -

Signature: _____



b720

Badge No.: 3200

Date: 10/07/2008

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00539	Calib. Date: 09/15/2008
Certification File No.:	00540	Calib. No.: 00009
Linearity File No.:	00541	Cert. Date: 09/15/2008
Solution File No.:	00542	Cert. No.: 00007
Sequential File No.:	00542	Lin. Date: 09/15/2008
		Lin. No.: 00006
		Soln. Date: 09/15/2008
		Soln. No.: 00064
		File Date: 09/15/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	07B046	Expires: 04/24/2009
		Bottle No.: 0687

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:30D		
Control 1 EC	0.099%	11:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		
Control 2 EC	0.099%	11:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		
Control 3 EC	0.099%	11:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUJPA-030 NE

Changed By:

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: _____

TPR. I WILLIAM CROSS #5342

Badge No.: 5342

Date: 09/15/2008

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00509	Calib. Date: 07/18/2008
Certification File No.:	00510	Calib. No.: 00008
Linearity File No.:	00511	Cert. Date: 07/18/2008
Solution File No.:	00537	Cert. No.: 00006
Sequential File No.:	00537	Lin. Date: 07/18/2008
		Lin. No.: 00005
		Soln. Date: 09/04/2008
		Soln. No.: 00063
		File Date: 09/04/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0163
Solution Control Lot:	07B046	Expires: 04/24/2009
		Bottle No.: 0690

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:44D		
Control 1 EC	0.100%	17:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:45D		
Control 2 EC	0.100%	17:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:46D		
Control 3 EC	0.099%	17:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:47D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  63200

Badge No.: 3200

Date: 09/04/2008

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00509	Calib. Date:	07/18/2008
Certification File No.:	00510	Cert. Date:	07/18/2008
Linearity File No.:	00511	Lin. Date:	07/18/2008
Solution File No.:	00525	Soln. Date:	08/12/2008
Sequential File No.:	00525	File Date:	08/12/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	07B046	Expires:	04/24/2009
		Bottle No.:	0691

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:17D		
Control 1 EC	0.098%	17:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		
Control 2 EC	0.099%	17:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		
Control 3 EC	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  3200

Badge No.: 3200

Date: 08/12/2008

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00415	Calib. Date: 11/28/2007
Certification File No.:	00416	Cert. No.: 00007
Linearity File No.:	00417	Cert. Date: 11/28/2007
Solution File No.:	00506	Lin. No.: 00004
Sequential File No.:	00506	Lin. Date: 11/28/2007
		Soln. No.: 00059
		Soln. Date: 07/16/2008
		File Date: 07/16/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	07B046	Expires: 04/24/2009
		Bottle No.: 0686

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:28D		
Control 1 EC	----%	15:29D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	----%	15:29D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	----%	--:--		
Control 2 EC	----%	--:--	N/A	
Control 2 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORELO	First Name: RONALD	MI: M
Signature: <u></u>	43200	Badge No.: 3200
		Date: 07/16/2008