

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00698	Calib. Date: 08/25/2009
Certification File No.:	00699	Calib. No.: 00011
Linearity File No.:	00700	Cert. Date: 08/25/2009
Solution File No.:	00755	Cert. No.: 00009
Sequential File No.:	00755	Lin. Date: 08/25/2009
		Lin. No.: 00008
		Soln. Date: 12/07/2009
		Soln. No.: 00088
		File Date: 12/07/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0254

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:25S		
Control 1 EC	0.102%	12:25S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:25S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26S		
Control 2 EC	0.101%	12:27S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:27S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27S		
Control 3 EC	0.101%	12:28S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:28S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 12/07/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00698	Calib. No.: 00011
Certification File No.:	00699	Cert. No.: 00009
Linearity File No.:	00700	Lin. No.: 00008
Solution File No.:	00746	Soln. No.: 00087
Sequential File No.:	00746	File Date: 11/11/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0259

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:13S		
Control 1 EC	0.102%	15:14S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:14S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:15S		
Control 2 EC	0.101%	15:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:16S		
Control 3 EC	0.100%	15:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

Badge No.: 3200

Date: 11/11/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00698	Calib. No.: 00011
Certification File No.:	00699	Cert. No.: 00009
Linearity File No.:	00700	Lin. No.: 00008
Solution File No.:	00729	Soln. No.: 00086
Sequential File No.:	00729	File Date: 10/12/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0257

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:49D		
Control 1 EC	0.100%	14:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:50D		
Control 2 EC	0.100%	14:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	14:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51D		
Control 3 EC	0.100%	14:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.



**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

0320

Badge No.: 3200

Date: 10/12/2009

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00698	Calib. Date:	08/25/2009
Certification File No.:	00699	Cert. No.:	00011
Linearity File No.:	00700	Cert. Date:	08/25/2009
Solution File No.:	00716	Lin. Date:	08/25/2009
Sequential File No.:	00716	Soln. Date:	09/18/2009
		Soln. No.:	00008
		File Date:	09/18/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	08F053	Expires:	06/09/2010
		Bottle No.:	0256

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:02D		
Control 1 EC	0.101%	10:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		
Control 2 EC	0.101%	10:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		
Control 3 EC	0.100%	10:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO	First Name: RONALD	MI: M
Signature: 	63200	Badge No.: 3200
		Date: 09/18/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00625	Calib. Date: 03/09/2009
Certification File No.:	00626	Cert. No.: 00010
Linearity File No.:	00627	Cert. Date: 03/09/2009
Solution File No.:	00696	Lin. No.: 00008
Sequential File No.:	00696	Lin. Date: 03/09/2009
		Soln. No.: 00081
		File Date: 08/22/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0260

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:25D		
Control 1 EC	0.101%	14:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:26D		
Control 2 EC	0.102%	14:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28D		
Control 3 EC	0.101%	14:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 08/22/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00625	Calib. Date:	03/09/2009
Certification File No.:	00626	Cert. Date:	03/09/2009
Linearity File No.:	00627	Lin. Date:	03/09/2009
Solution File No.:	00684	Soln. Date:	07/29/2009
Sequential File No.:	00684	File Date:	07/29/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUK S3-0133
Solution Control Lot:	08F053	Expires:	06/09/2010
		Bottle No.:	0264

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:08D		
Control 1 EC	0.103%	10:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		
Control 2 EC	0.102%	10:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		
Control 3 EC	0.102%	10:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	10:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCBRIDE

First Name: STEPHEN

MI: J

Signature: \_\_\_\_\_

*# 3202*

Badge No.: 3202

Date: 07/29/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00625	Calib. Date: 03/09/2009
Certification File No.:	00626	Cert. No.: 00010
Linearity File No.:	00627	Lin. Date: 03/09/2009
Solution File No.:	00683	Cert. No.: 00008
Sequential File No.:	00683	Lin. No.: 00007
		Soln. Date: 07/27/2009
		Soln. No.: 00079
		File Date: 07/27/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0262

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:18D		
Control 1 EC	0.102%	16:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20D		
Control 2 EC	0.102%	16:21D	34.0°C	AMBIENT AIR CHECK ERR
Control 2 IR	-.---%	16:21D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	0.038%	16:21D		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCBRIDE

First Name: STEPHEN

MI: J

Signature: \_\_\_\_\_

Badge No.: 3202

Date: 07/27/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00625	Calib. Date: 03/09/2009
Certification File No.:	00626	Calib. No.: 00010
Linearity File No.:	00627	Cert. Date: 03/09/2009
Solution File No.:	00674	Cert. No.: 00008
Sequential File No.:	00674	Lin. Date: 03/09/2009
		Lin. No.: 00007
		Soln. Date: 07/02/2009
		Soln. No.: 00078
		File Date: 07/02/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	08F053	Expires: 06/09/2010
		Bottle No.: 0270

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:25D		
Control 1 EC	0.101%	11:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 2 EC	0.101%	11:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 3 EC	0.102%	11:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



63200

Badge No.: 3200

Date: 07/02/2009



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	00625	Calib. Date:	03/09/2009
Certification File No.:	00626	Cert. No.:	00010
Linearity File No.:	00627	Lin. Date:	03/09/2009
Solution File No.:	00658	Soln. No.:	00008
Sequential File No.:	00658	File Date:	05/24/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	08F053	Expires:	06/09/2010
		Bottle No.:	0266

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:43D		
Control 1 EC	0.102%	23:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	23:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:45D		
Control 2 EC	0.103%	23:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	23:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46D		
Control 3 EC	0.102%	23:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	23:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 05/24/2009

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
 Location: STRATFORD POLICE DEPT.  
 Calibration File No.: 00625  
 Certification File No.: 00626  
 Linearity File No.: 00627  
 Solution File No.: 00648  
 Sequential File No.: 00648

Serial No.: ARNK-0039

Calib. Date: 03/09/2009  
 Cert. Date: 03/09/2009  
 Lin. Date: 03/09/2009  
 Soln. Date: 04/24/2009  
 File Date: 04/24/2009

Calib. No.: 00010  
 Cert. No.: 00008  
 Lin. No.: 00007  
 Soln. No.: 00075

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 08F053

Model No.: CU-34

Serial No.: DDUL S3-0163  
 Expires: 06/09/2010  
 Bottle No.: 0269

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:20D		
Control 1 EC	0.101%	10:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21D		
Control 2 EC	0.101%	10:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23D		
Control 3 EC	0.101%	10:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

*[Handwritten Signature]* 3200

Badge No.: 3200

Date: 04/24/2009

# Calibrating Unit New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.		
Calibration File No.: 00625	Calib. Date: 03/09/2009	Calib. No.: 00010
Certification File No.: 00626	Cert. Date: 03/09/2009	Cert. No.: 00008
Linearity File No.: 00627	Lin. Date: 03/09/2009	Lin. No.: 00007
Solution File No.: 00638	Soln. Date: 03/31/2009	Soln. No.: 00074
Sequential File No.: 00638	File Date: 03/31/2009	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 06/09/2010
Solution Control Lot: 08F053		Bottle No.: 0271

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:43S		
Control 1 EC	0.100%	19:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:44S		
Control 2 EC	0.100%	19:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	19:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:45S		
Control 3 EC	0.101%	19:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAGLIARDI

First Name: FRANK

MI: J

Signature: \_\_\_\_\_

Badge No.: 3203

Date: 03/31/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00539	Calib. Date: 09/15/2008
Certification File No.:	00540	Calib. No.: 00009
Linearity File No.:	00541	Cert. Date: 09/15/2008
Solution File No.:	00619	Cert. No.: 00007
Sequential File No.:	00619	Lin. Date: 09/15/2008
		Lin. No.: 00006
		Soln. Date: 03/03/2009
		Soln. No.: 00072
		File Date: 03/03/2009

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 06/09/2010
Solution Control Lot: 08F053		Bottle No.: 0267

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:43S		
Control 1 EC	0.100%	16:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44S		
Control 2 EC	0.099%	16:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45S		
Control 3 EC	0.100%	16:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200

Date: 03/03/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	00539	Calib. Date: 09/15/2008
Certification File No.:	00540	Calib. No.: 00009
Linearity File No.:	00541	Cert. Date: 09/15/2008
Solution File No.:	00607	Cert. No.: 00007
Sequential File No.:	00607	Lin. Date: 09/15/2008
		Lin. No.: 00006
		Soln. Date: 02/04/2009
		Soln. No.: 00071
		File Date: 02/04/2009

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 06/09/2010
Solution Control Lot: 08F053		Bottle No.: 0268

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:39S		
Control 1 EC	0.101%	17:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:40S		
Control 2 EC	0.101%	17:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:42S		
Control 3 EC	0.101%	17:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:43S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: -

Signature: 

63200

Badge No.: 3200

Date: 02/04/2009

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
Location: STRATFORD POLICE DEPT.  
Calibration File No.: 00539 Calib. Date: 09/15/2008 Calib. No.: 00009  
Certification File No.: 00540 Cert. Date: 09/15/2008 Cert. No.: 00007  
Linearity File No.: 00541 Lin. Date: 09/15/2008 Lin. No.: 00006  
Solution File No.: 00604 Soln. Date: 01/17/2009 Soln. No.: 00070  
Sequential File No.: 00604 File Date: 01/17/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163  
Control Solution %: 0.100% Expires: 04/24/2009  
Solution Control Lot: 07B046 Bottle No.: 0694

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:54S		
Control 1 EC	0.099%	12:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55S		
Control 2 EC	0.098%	12:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:57S		
Control 3 EC	0.099%	12:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KELLY

First Name: ROBERT

MI: J

Signature: 

Badge No.: 3204

Date: 01/17/2009