

# Calibrating Unit

## New Standard Solution Report

|                         |                        |                          |
|-------------------------|------------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARNK-0039    |
| Location:               | STRATFORD POLICE DEPT. |                          |
| Calibration File No.:   | 02155                  | Calib. Date: 09/01/2015  |
| Certification File No.: | 02156                  | Calib. No.: 00032        |
| Linearity File No.:     | 02157                  | Cert. Date: 09/01/2015   |
| Solution File No.:      | 02228                  | Cert. No.: 00023         |
| Sequential File No.:    | 02228                  | Lin. Date: 09/01/2015    |
|                         |                        | Lin. No.: 00022          |
|                         |                        | Soln. Date: 12/29/2015   |
|                         |                        | Soln. No.: 00235         |
|                         |                        | File Date: 12/29/2015    |
| <br>                    |                        |                          |
| Calibrating Unit:       | WET                    | Model No.: CU-34         |
| Control Solution %:     | 0.100%                 | Serial No.: DDUK S3-0133 |
| Solution Control Lot:   | 15H140                 | Expires: 08/04/2017      |
|                         |                        | Bottle No.: 0170         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 16:07S |                |                     |
| Control 1 EC      | 0.103% | 16:08S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.103% | 16:08S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:08S |                |                     |
| Control 2 EC      | 0.103% | 16:09S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.103% | 16:09S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:10S |                |                     |
| Control 3 EC      | 0.102% | 16:10S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.101% | 16:10S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:11S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

*[Handwritten Signature]* 63200

Badge No.: 3200

Date: 12/29/2015

# Calibrating Unit

## New Standard Solution Report

|                         |                        |                          |
|-------------------------|------------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARNK-0039    |
| Location:               | STRATFORD POLICE DEPT. |                          |
| Calibration File No.:   | 02155                  | Calib. Date: 09/01/2015  |
| Certification File No.: | 02156                  | Calib. No.: 00032        |
| Linearity File No.:     | 02157                  | Ceri. Date: 09/01/2015   |
| Solution File No.:      | 02210                  | Cert. No.: 00023         |
| Sequential File No.:    | 02210                  | Lin. Date: 09/01/2015    |
|                         |                        | Lin. No.: 00022          |
|                         |                        | Soln. Date: 12/06/2015   |
|                         |                        | Soln. No.: 00234         |
|                         |                        | File Date: 12/06/2015    |
| Calibrating Unit:       | WET                    | Model No.: CU-34         |
| Control Solution %:     | 0.100%                 | Serial No.: DDUK S3-0133 |
| Solution Control Lot:   | 14D129                 | Expires: 04/29/2016      |
|                         |                        | Bottle No.: 1195         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 21:47S |                |                     |
| Control 1 EC      | 0.103% | 21:48S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.105% | 21:48S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:48S |                |                     |
| Control 2 EC      | 0.102% | 21:49S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.103% | 21:49S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:50S |                |                     |
| Control 3 EC      | 0.102% | 21:50S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.104% | 21:50S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:51S |                |                     |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  6320

Badge No.: 3200

Date: 12/06/2015

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
Location: STRATFORD POLICE DEPT.  
Calibration File No.: 02155 Calib. Date: 09/01/2015 Calib. No.: 00032  
Certification File No.: 02156 Cert. Date: 09/01/2015 Cert. No.: 00023  
Linearity File No.: 02157 Lin. Date: 09/01/2015 Lin. No.: 00022  
Solution File No.: 02196 Soln. Date: 11/12/2015 Soln. No.: 00233  
Sequential File No.: 02196 File Date: 11/12/2015


Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133  
Control Solution %: 0.100% Expires: 04/29/2016  
Solution Control Lot: 14D129 Bottle No.: 1199

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 13:16S        |                               |                           |
| Control 1 EC      | 0.103%         | 13:16S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.104%         | 13:16S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 13:17S        |                               |                           |
| Control 2 EC      | 0.102%         | 13:18S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.103%         | 13:18S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 13:19S        |                               |                           |
| Control 3 EC      | 0.102%         | 13:19S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.103%         | 13:19S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 13:20S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MORELLO First Name: RONALD MI: M  
Signature:  Badge No.: 3200  
Date: 11/12/2015

Signature:  Badge No.: 3200  
Date: 11/12/2015

Signature:  Badge No.: 3200  
Date: 11/12/2015

# Calibrating Unit

## New Standard Solution Report

|                          |                        |                          |
|--------------------------|------------------------|--------------------------|
| <b>Equipment</b>         | Alcotest 7110 MKIII-C  | Serial No.: ARNK-0039    |
| Location:                | STRATFORD POLICE DEPT. |                          |
| Calibration File No.:    | 02155                  | Calib. Date: 09/01/2015  |
| Certification File No.:  | 02156                  | Cert. No.: 00032         |
| Linearity File No.:      | 02157                  | Cert. No.: 00023         |
| Solution File No.:       | 02186                  | Lin. No.: 00022          |
| Sequential File No.:     | 02186                  | Soln. No.: 00232         |
|                          |                        | File Date: 10/19/2015    |
| <b>Calibrating Unit:</b> | <b>WET</b>             | <b>Model No.:</b> CU-34  |
| Control Solution %:      | 0.100%                 | Serial No.: DDUK S3-0133 |
| Solution Control Lot:    | 14D129                 | Expires: 04/29/2016      |
|                          |                        | Bottle No.: 1198         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:05D |                |                     |
| Control 1 EC      | 0.102% | 10:06D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.103% | 10:06D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:07D |                |                     |
| Control 2 EC      | 0.101% | 10:07D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.102% | 10:07D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:08D |                |                     |
| Control 3 EC      | 0.101% | 10:09D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.104% | 10:09D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:09D |                |                     |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200

Date: 10/19/2015

# Calibrating Unit

## New Standard Solution Report

|                         |                        |                          |
|-------------------------|------------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARNK-0039    |
| Location:               | STRATFORD POLICE DEPT. |                          |
| Calibration File No.:   | 02155                  | Calib. Date: 09/01/2015  |
| Certification File No.: | 02156                  | Calib. No.: 00032        |
| Linearity File No.:     | 02157                  | Cert. Date: 09/01/2015   |
| Solution File No.:      | 02174                  | Cert. No.: 00023         |
| Sequential File No.:    | 02174                  | Lin. Date: 09/01/2015    |
|                         |                        | Lin. No.: 00022          |
|                         |                        | Soln. Date: 09/25/2015   |
|                         |                        | Soln. No.: 00231         |
|                         |                        | File Date: 09/25/2015    |
| <br>                    |                        |                          |
| Calibrating Unit:       | WET                    | Model No.: CU-34         |
| Control Solution %:     | 0.100%                 | Serial No.: DDUK S3-0133 |
| Solution Control Lot:   | 14D129                 | Expires: 04/29/2016      |
|                         |                        | Bottle No.: 1194         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 09:53D |                |                     |
| Control 1 EC      | 0.103% | 09:54D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.103% | 09:54D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:54D |                |                     |
| Control 2 EC      | 0.102% | 09:55D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.103% | 09:55D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:56D |                |                     |
| Control 3 EC      | 0.101% | 09:56D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.103% | 09:56D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:57D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_

*[Handwritten Signature]* 63200

Badge No.: 3200

Date: 09/25/2015

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |            |
|-------------------------|------------------------|--------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0039  |
| Location:               | STRATFORD POLICE DEPT. |              |            |
| Calibration File No.:   | 02065                  | Calib. Date: | 04/02/2015 |
| Certification File No.: | 02066                  | Cert. Date:  | 04/02/2015 |
| Linearity File No.:     | 02067                  | Lin. Date:   | 04/02/2015 |
| Solution File No.:      | 02143                  | Soln. Date:  | 08/06/2015 |
| Sequential File No.:    | 02143                  | File Date:   | 08/06/2015 |

|                       |        |            |       |             |              |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit:     | WET    | Model No.: | CU-34 | Serial No.: | DDUK S3-0133 |
| Control Solution %:   | 0.100% |            |       | Expires:    | 04/29/2016   |
| Solution Control Lot: | 14D129 |            |       | Bottle No.: | 1178         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:54D |                |                     |
| Control 1 EC      | 0.101% | 10:55D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 10:55D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:55D |                |                     |
| Control 2 EC      | 0.100% | 10:56D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.102% | 10:56D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:57D |                |                     |
| Control 3 EC      | 0.100% | 10:57D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.101% | 10:57D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:58D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  6-32-00

Badge No.: 3200-

Date: 08/06/2015

# Calibrating Unit New Standard Solution Report

|                         |                        |              |              |
|-------------------------|------------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0039    |
| Location:               | STRATFORD POLICE DEPT. |              |              |
| Calibration File No.:   | 02065                  | Calib. Date: | 04/02/2015   |
| Certification File No.: | 02066                  | Cert. No.:   | 00031        |
| Linearity File No.:     | 02067                  | Lin. Date:   | 04/02/2015   |
| Solution File No.:      | 02130                  | Cert. No.:   | 00022        |
| Sequential File No.:    | 02130                  | Lin. No.:    | 00021        |
|                         |                        | Soln. Date:  | 07/13/2015   |
|                         |                        | Soln. No.:   | 00228        |
|                         |                        | File Date:   | 07/13/2015   |
| <br>                    |                        |              |              |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDUK 53-0133 |
| Solution Control Lot:   | 14D129                 | Expires:     | 04/29/2016   |
|                         |                        | Bottle No.:  | 1177         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 12:50D |                |                     |
| Control 1 EC      | 0.100% | 12:51D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 12:51D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:52D |                |                     |
| Control 2 EC      | 0.100% | 12:53D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.101% | 12:53D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:53D |                |                     |
| Control 3 EC      | 0.100% | 12:54D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 12:54D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:55D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAGLIARDI

First Name: FRANK

MI: J

Signature:

*Frank J. Gagliardi* \* 3203

Badge No.: 3203

Date: 07/13/2015

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |            |             |       |
|-------------------------|------------------------|--------------|------------|-------------|-------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0039  |             |       |
| Location:               | STRATFORD POLICE DEPT. |              |            |             |       |
| Calibration File No.:   | 02065                  | Calib. Date: | 04/02/2015 | Calib. No.: | 00031 |
| Certification File No.: | 02066                  | Cert. Date:  | 04/02/2015 | Cert. No.:  | 00022 |
| Linearity File No.:     | 02067                  | Lin. Date:   | 04/02/2015 | Lin. No.:   | 00021 |
| Solution File No.:      | 02123                  | Soln. Date:  | 06/23/2015 | Soln. No.:  | 00227 |
| Sequential File No.:    | 02123                  | File Date:   | 06/23/2015 |             |       |

|                       |        |            |       |             |              |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit:     | WET    | Model No.: | CU-34 | Serial No.: | DDUK S3-0133 |
| Control Solution %:   | 0.100% |            |       | Expires:    | 04/29/2016   |
| Solution Control Lot: | 14D129 |            |       | Bottle No.: | 1184         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 11:49D        |                               |                           |
| Control 1 EC      | 0.101%         | 11:50D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 11:50D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:51D        |                               |                           |
| Control 2 EC      | 0.100%         | 11:51D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 11:51D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:52D        |                               |                           |
| Control 3 EC      | 0.100%         | 11:53D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.101%         | 11:53D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:53D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DL 1191-0015 (K9)

### Changed By:

|                                    |                   |                  |
|------------------------------------|-------------------|------------------|
| Last Name: ALCOTT                  | First Name: KEVIN | MI: W.           |
| Signature: <u>TPA I K W. #6704</u> | Badge No.: 6704   | Date: 06/23/2015 |





**Dräger**

# CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Dräger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: \_\_\_\_\_

Serial Number:

DDUKS3-0133

Certification Date

5-1-15

Technician

BC

Re-Certification Due Date

5-1-16

# Dräger

## ALCOTEST® 7110 TEMPERATURE PROBE

### CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDLPI-0015

Certification date:

4-30-15

Next Certification due

4-30-16

Probe Value

108

Dräger Safety Diagnostics, Inc.  
Technical Services Department

BC

DEPARTMENT OF  
**Water and Public Safety**  
This is to certify that

**KEVIN W. ALCOTT**  
NEW JERSEY STATE POLICE

IS QUALIFIED AND COMPETENT TO CONDUCT BREATH TESTS AND IS SUBJECT TO CHAPTER 142 OF  
THE LAWS OF 1994 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C  
A METHOD TO OBTAINING INTOXICATION  
LEVEL UNDER MY NAME AT TRENTON, NEW JERSEY, ON 18th DAY OF APRIL

TWO THOUSAND AND EIGHT

*[Signature]*  
SUPERVISOR  
NEW JERSEY STATE POLICE

*[Signature]*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| DATE    | Refresher Course PLACE | INSTRUCTOR |
|---------|------------------------|------------|
| 8-25-10 | ALPA                   | A. J. W.   |
| 1-12-12 | BOWEN WOOD             | C. J.      |
| 1-16-14 | SAYREVILLE PD          | R. J. W.   |
| 4.      |                        |            |
| 5.      |                        |            |
| 6.      |                        |            |
| 7.      |                        |            |
| 8.      |                        |            |
| 9.      |                        |            |

S.P. 2848 (Rev. 07/07)

DEPARTMENT OF  
**Water and Public Safety**  
This is to certify that

**Kevin W. Alcott**  
Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT BREATH TESTS AND IS SUBJECT TO CHAPTER 142 OF  
THE LAWS OF 1994 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C  
A METHOD TO OBTAINING INTOXICATION  
LEVEL UNDER MY NAME AT TRENTON, NEW JERSEY, ON 21st DAY OF May

TWO THOUSAND AND Fifteen

*[Signature]*  
SUPERVISOR  
NEW JERSEY STATE POLICE

*[Signature]*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| DATE | Refresher Course PLACE | INSTRUCTOR |
|------|------------------------|------------|
| 1.   |                        |            |
| 2.   |                        |            |
| 3.   |                        |            |
| 4.   |                        |            |
| 5.   |                        |            |
| 6.   |                        |            |
| 7.   |                        |            |
| 8.   |                        |            |
| 9.   |                        |            |

S.P. 2848 (Rev. 08/12)

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |            |
|-------------------------|------------------------|--------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0039  |
| Location:               | STRATFORD POLICE DEPT. |              |            |
| Calibration File No.:   | 02065                  | Calib. Date: | 04/02/2015 |
| Certification File No.: | 02066                  | Cert. Date:  | 04/02/2015 |
| Linearity File No.:     | 02067                  | Lin. Date:   | 04/02/2015 |
| Solution File No.:      | 02111                  | Soln. Date:  | 06/12/2015 |
| Sequential File No.:    | 02111                  | File Date:   | 06/12/2015 |

|                       |        |            |       |             |              |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit:     | WET    | Model No.: | CU-34 | Serial No.: | DDUL S3-0163 |
| Control Solution %:   | 0.100% |            |       | Expires:    | 04/29/2016   |
| Solution Control Lot: | 14D129 |            |       | Bottle No.: | 1193         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:10D |                |                     |
| Control 1 EC      | 0.101% | 10:10D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.102% | 10:10D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:11D |                |                     |
| Control 2 EC      | 0.101% | 10:12D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 10:12D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:12D |                |                     |
| Control 3 EC      | 0.100% | 10:13D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.101% | 10:13D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:14D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 06/12/2015

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |            |
|-------------------------|------------------------|--------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0039  |
| Location:               | STRATFORD POLICE DEPT. |              |            |
| Calibration File No.:   | 02065                  | Calib. Date: | 04/02/2015 |
| Certification File No.: | 02066                  | Cert. Date:  | 04/02/2015 |
| Linearity File No.:     | 02067                  | Lin. Date:   | 04/02/2015 |
| Solution File No.:      | 02094                  | Soln. Date:  | 05/21/2015 |
| Sequential File No.:    | 02094                  | File Date:   | 05/21/2015 |

|                       |        |            |       |             |              |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit:     | WEI    | Model No.: | CU-34 | Serial No.: | DDUL S3-0163 |
| Control Solution %:   | 0.100% |            |       | Expires:    | 04/29/2016   |
| Solution Control Lot: | 14D129 |            |       | Bottle No.: | 1189         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 15:16D |                |                     |
| Control 1 EC      | 0.103% | 15:17D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.103% | 15:17D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:17D |                |                     |
| Control 2 EC      | 0.103% | 15:18D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.104% | 15:18D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:19D |                |                     |
| Control 3 EC      | 0.103% | 15:19D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.104% | 15:19D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:20D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 05/21/2015

# Calibrating Unit New Standard Solution Report

|                         |                        |              |              |
|-------------------------|------------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0039    |
| <b>Location</b>         | STRATFORD POLICE DEPT. |              |              |
| Calibration File No.:   | 02065                  | Calib. Date: | 04/02/2015   |
| Certification File No.: | 02066                  | Cert. No.:   | 00022        |
| Linearity File No.:     | 02067                  | Lin. Date:   | 04/02/2015   |
| Solution File No.:      | 02083                  | Soln. No.:   | 00224        |
| Sequential File No.:    | 02083                  | File Date:   | 04/26/2015   |
| <br>                    |                        |              |              |
| <b>Calibrating Unit</b> | WET                    | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDJL S3-0163 |
| Solution Control Lot:   | 14D129                 | Expires:     | 04/29/2016   |
|                         |                        | Bottle No.:  | 1183         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:07D |                |                     |
| Control 1 EC      | 0.102% | 10:07D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.103% | 10:07D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:08D |                |                     |
| Control 2 EC      | 0.101% | 10:09D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.101% | 10:09D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:10D |                |                     |
| Control 3 EC      | 0.101% | 10:10D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.103% | 10:10D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:11D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: 

Badge No.: 3200

Date: 04/26/2015

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
Location: STRATFORD POLICE DEPT.  
Calibration File No.: 01957 Calib. Date: 10/20/2014 Calib. No.: 00030  
Certification File No.: 01958 Cert. Date: 10/20/2014 Cert. No.: 00021  
Linearity File No.: 01959 Lin. Date: 10/20/2014 Lin. No.: 00020  
Solution File No.: 02061 Soln. Date: 03/28/2015 Soln. No.: 00222  
Sequential File No.: 02061 File Date: 03/28/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0163  
Control Solution %: 0.100% Expires: 04/29/2016  
Solution Control Lot: 14D129 Bottle No.: 1182

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 01:30S        |                               |                           |
| Control 1 EC      | 0.098%         | 01:30S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 01:30S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 01:31S        |                               |                           |
| Control 2 EC      | 0.097%         | 01:32S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 01:32S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 01:33S        |                               |                           |
| Control 3 EC      | 0.097%         | 01:34S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 01:34S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 01:34S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KELLY First Name: ROBERT MI: J

Signature: Sgt Kelly # 3204 Badge No.: 3204  
Date: 03/28/2015

Signature: Sgt Kelly # 3204 Badge No.: 3204  
Date: 03/28/2015

Signature: Sgt Kelly # 3204 Badge No.: 3204  
Date: 03/28/2015

Signature: Sgt Kelly # 3204 Badge No.: 3204  
Date: 03/28/2015

Signature: Sgt Kelly # 3204 Badge No.: 3204  
Date: 03/28/2015

# Calibrating Unit New Standard Solution Report

|                         |                        |                         |
|-------------------------|------------------------|-------------------------|
| Equipment               | Alcotest 7110 MKIII-C  | Serial No.: ARNK-0039   |
| Location:               | STRATFORD POLICE DEPT. |                         |
| Calibration File No.:   | 01957                  | Calib. Date: 10/20/2014 |
| Certification File No.: | 01958                  | Cert. No.: 00030        |
| Linearity File No.:     | 01959                  | Cert. Date: 10/20/2014  |
| Solution File No.:      | 02050                  | Lin. No.: 00021         |
| Sequential File No.:    | 02050                  | Lin. Date: 10/20/2014   |
|                         |                        | Soln. No.: 00221        |
|                         |                        | Soln. Date: 03/17/2015  |
|                         |                        | File Date: 03/17/2015   |

|                       |        |                  |                          |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit:     | WET    | Model No.: CU-34 | Serial No.: DDUL S3-0163 |
| Control Solution %:   | 0.100% |                  | Expires: 04/29/2016      |
| Solution Control Lot: | 14D129 |                  | Bottle No.: 1176         |


| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 23:22S |                |                     |
| Control 1 EC      | 0.100% | 23:23S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 23:23S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:24S |                |                     |
| Control 2 EC      | 0.098% | 23:24S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 23:24S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:25S |                |                     |
| Control 3 EC      | 0.099% | 23:26S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 23:26S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:27S |                |                     |


All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**


Last Name: MCBRIDE      First Name: STEPHEN      MJ: J

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

Signature:  # 3202      Badge No.: 3202  
Date: 03/17/2015

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |            |
|-------------------------|------------------------|--------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0039  |
| Location:               | STRATFORD POLICE DEPT. |              |            |
| Calibration File No.:   | 01957                  | Calib. Date: | 10/20/2014 |
| Certification File No.: | 01958                  | Cert. No.:   | 00030      |
| Lancarity File No.:     | 01959                  | Lin. Date:   | 10/20/2014 |
| Solution File No.:      | 02040                  | Lin. No.:    | 00020      |
| Sequential File No.:    | 02040                  | Soln. Date:  | 02/21/2015 |
|                         |                        | File Date:   | 02/21/2015 |

|                       |        |            |       |             |              |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit:     | WET    | Model No.: | CU-34 | Serial No.: | DDUL 53-0163 |
| Control Solution %:   | 0.100% |            |       | Expires:    | 04/29/2016   |
| Solution Control Lot: | 14D129 |            |       | Bottle No.: | 1192         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:29S |                |                     |
| Control 1 EC      | 0.100% | 10:29S | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.101% | 10:29S | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:30S |                |                     |
| Control 2 EC      | 0.098% | 10:31S | 33.9°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 10:31S | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:32S |                |                     |
| Control 3 EC      | 0.100% | 10:33S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 10:33S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:34S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 02/21/2015



# Calibrating Unit New Standard Solution Report

|                         |                        |                         |
|-------------------------|------------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARNK-0039   |
| Location:               | STRATFORD POLICE DEPT. |                         |
| Calibration File No.:   | 01957                  | Calib. Date: 10/20/2014 |
| Certification File No.: | 01958                  | Calib. No.: 00030       |
| Linearity File No.:     | 01959                  | Cert. Date: 10/20/2014  |
| Solution File No.:      | 02031                  | Cert. No.: 00021        |
| Sequential File No.:    | 02031                  | Lin. Date: 10/20/2014   |
|                         |                        | Lin. No.: 00020         |
|                         |                        | Soln. Date: 02/03/2015  |
|                         |                        | Soln. No.: 00219        |
|                         |                        | File Date: 02/03/2015   |

|                       |        |                  |                          |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit:     | WET    | Model No.: CU-34 | Serial No.: DDUL S3-0163 |
| Control Solution %:   | 0.100% |                  | Expires: 04/29/2016      |
| Solution Control Lot: | 14D129 |                  | Bottle No.: 1187         |


| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:12S |                |                     |
| Control 1 EC      | 0.099% | 10:13S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 10:13S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:14S |                |                     |
| Control 2 EC      | 0.098% | 10:14S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 10:14S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:15S |                |                     |
| Control 3 EC      | 0.098% | 10:16S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 10:16S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:17S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

|                    |                    |       |
|--------------------|--------------------|-------|
| Last Name: MORELLO | First Name: RONALD | MI: M |
|--------------------|--------------------|-------|

|  |                  |
|--|------------------|
| Signature:  63200 | Badge No.: 3200  |
|  | Date: 02/03/2015 |


|                    |                    |       |
|--------------------|--------------------|-------|
| Last Name: MORELLO | First Name: RONALD | MI: M |
|--------------------|--------------------|-------|

|  |                  |
|--|------------------|
| Signature:  63200 | Badge No.: 3200  |
|  | Date: 02/03/2015 |


|                    |                    |       |
|--------------------|--------------------|-------|
| Last Name: MORELLO | First Name: RONALD | MI: M |
|--------------------|--------------------|-------|

|  |                  |
|--|------------------|
| Signature:  63200 | Badge No.: 3200  |
|  | Date: 02/03/2015 |

|                    |                    |       |
|--------------------|--------------------|-------|
| Last Name: MORELLO | First Name: RONALD | MI: M |
|--------------------|--------------------|-------|

|  |                  |
|--|------------------|
| Signature:  63200 | Badge No.: 3200  |
|  | Date: 02/03/2015 |

|                    |                    |       |
|--------------------|--------------------|-------|
| Last Name: MORELLO | First Name: RONALD | MI: M |
|--------------------|--------------------|-------|

|  |                  |
|--|------------------|
| Signature:  63200 | Badge No.: 3200  |
|  | Date: 02/03/2015 |



State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON, NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
 Governor

KIM GUADAGINO  
 Lt. Governor

JOHN J. HOFFMAN  
 Acting Attorney General

COLONEL JOSEPH R. PUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc.

ANALYSIS DATE: 5/06/2014

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 14D129

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1208 to 0.1217 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 29, 2016.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.  
 Research Scientist  
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 6<sup>th</sup> day of May, 2014.

Kim Guadagno  
 Governor

KIM GUADAGINO  
 Lt. Governor

DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON, NJ 08628-0068  
 (609) 882-2000

Acting Attorney General

COLONEL JOSEPH R. PUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**


ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but

# Packing Slip

Customer No 150053901 Order No 10714984 Order date 07/21/2014

Package No. 60875365 Ship date 07/21/2014  
Please refer back to Logistics

Ship to  
150053901  
STRATFORD POLICE DEPARTMENT  
ATTN: SGT MCBRIDE  
CAMDEN COUNTY  
715 UNION AVENUE  
STRATFORD, NJ 08084  
USA



Your Purchase Order Number  
07/21/2014  
072114  
Any questions? Please contact:  
BRENDA BROTHMAN

Customer  
STRATFORD POLICE DEPARTMENT  
CAMDEN COUNTY  
715 UNION AVENUE  
STRATFORD, NJ 08084  
USA

Your sales person  
Don Poallot  
Phone: 973-398-3228

Delivery name  
PA  
WEIGHT ALLOWED

| Item | IN QUAN | Part Description   |
|------|---------|--|
|      |         | Ship via: Fed Ex Ground  |
|      |         | NOTE: All claims for shortages/defects must be reported within 10 Business Days After receipt of order. All returns must have our logo clearly marked on the box. Please call Customer Service at 866-385-5999 |
| 0010 | 1 EA    | 4540013<br>JAR. SIMULATOR<br>Ordered / Back ordered : 1 / 0 EA   |
| 0020 | 20 EA   | 6805700<br>MOUTHPIECE, ALCO 7110/7410/9530 PKG25<br>Ordered / Back ordered : 20 / 0 EA   |

Dräger Safety Diagnostics Inc.  
Accounting Address:  
107 Technology Drive  
Pittsburgh, PA 15225  
Tel: (412) 786-5527  
Fax: (412) 786-5566

Remit Wire Transfers To:  
Citizens Bank  
Acct # 6208426615  
Acct Name: Dräger Safety Diagnostics Inc.  
Transit Number: 036076157  
Federal ID Number: 84-1600159

REMIT TO:  
Dräger Safety Diagnostics Inc.  
P.O. Box 200337  
Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150053903    10714984    07/21/2014

Packing slip no.    Ship date  
80875365    07/21/2014  
Please reference on inquiries

Page    2 / 2

| Item# | sh. Quant. | Part no. Description  |  |  |  |
|-------|------------|---|--|--|--|
| 0030  | 25 EA      | 4407061<br>CERT SOLUTION, .10% NEW JERSEY CERTIFIED<br>Ordered / Back ordered : 25 / 0 EA |  |  |  |
| 0040  | 1 EA       | FREIGHT<br>Freight Charges<br>Ordered / Back ordered : 1 / 0 EA                           |  |  |  |

Page    2 / 2

| Item# | sh. Quant. | Part no. Description  |  |  |  |
|-------|------------|---|--|--|--|
| 0030  | 25 EA      | 4407061<br>CERT SOLUTION, .10% NEW JERSEY CERTIFIED<br>Ordered / Back ordered : 25 / 0 EA |  |  |  |
| 0040  | 1 EA       | FREIGHT<br>Freight Charges<br>Ordered / Back ordered : 1 / 0 EA                           |  |  |  |

# Calibrating Unit New Standard Solution Report

|                        |                      |              |              |
|------------------------|----------------------|--------------|--------------|
| <b>Equipment</b>       | Alcatel 7110 MKIII-C | Serial No.   | ARNE 0039    |
| Location               | STAFFORD POLICE DEPT | Calib. Inst. | 10/30/2014   |
| Calibration File No.   | 01957                | Calib. No.   | 00030        |
| Certification File No. | 01958                | Cal. Date    | 10/29/2014   |
| Laboratory File No.    | 01959                | Lab. No.     | 00020        |
| Solution File No.      | 00018                | Soln. Date   | 01/10/2015   |
| Sequential File No.    | 00018                | File Date    | 01/10/2015   |
| Calibrating Unit       | WB7                  | Model No.    | CU M         |
| Control Solution No.   | 01009                | Serial No.   | UDLL S1-0161 |
| Solution Control Lot   | 142129               | Expires      | 04/29/2016   |
|                        |                      | Route No.    | 1191         |

| Function          | Result | Time    | Temperature | Comment(s)          |
|-------------------|--------|---------|-------------|---------------------|
|                   | %BAC   | min:sec | Sample (°C) | (if Error(s))       |
| Ambient Air Blank | 0.000% | 11:535  |             |                     |
| Control 1 EC      | 0.100% | 11:555  | 34.0°C      | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 11:555  | 34.0°C      | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:565  |             |                     |
| Control 2 EC      | 0.099% | 11:575  | 34.0°C      | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 11:575  | 34.0°C      | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:585  |             |                     |
| Control 3 EC      | 0.100% | 11:585  | 34.0°C      | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 11:585  | 34.0°C      | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:595  |             |                     |

All tests within acceptable tolerance

On this date, I declared the above instrument "NEW SOLUTION" in accordance with Alcatel 7110 rigorous training and procedures established by the (NIST) Chief Forensic Scientist.

**Changed By:**

Last Name: MICHELLO

First Name: RONALD

ML M

Signature: 

Badge No. 1230

Date 01/10/2015