

# Calibrating Unit

## New Standard Solution Report

### Equipment

Alcotest 7110 MKIII-C  
Location: STRATFORD POLICE DEPT.  
Serial No.: ARNK-0039  
Calibration File No.: 02647  
Certification File No.: 02648  
Linearity File No.: 02649  
Solution File No.: 02683  
Sequential File No.: 02683  
Calib. Date: 11/29/2018  
Cert. Date: 11/29/2018  
Lin. Date: 11/29/2018  
Soln. Date: 02/06/2019  
File Date: 02/06/2019  
Calib. No.: 00040  
Cert. No.: 00030  
Lin. No.: 00029  
Soln. No.: 00301

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 18090  
Model No.: CU-34  
Serial No.: DDUL S3-0163  
Expires: 03/13/2020  
Bottle No.: 0682

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:49S		
Control 1 EC	0.102%	13:50S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51S		
Control 2 EC	0.101%	13:52S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:52S		
Control 3 EC	0.100%	13:53S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	13:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KELLY

First Name: ROBERT

MI: J

Signature: \_\_\_\_\_

Badge No.: 3201

Date: 02/06/2019

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
 Location: STRATFORD POLICE DEPT.  
 Calibration File No.: 02647  
 Certification File No.: 02648  
 Linearity File No.: 02649  
 Solution File No.: 02669  
 Sequential File No.: 02669

Serial No.: ARNK-0039

Calib. No.: 00040  
 Cert. No.: 00030  
 Lin. No.: 00029  
 Soln. No.: 00300

Calib. Date: 11/29/2018  
 Cert. Date: 11/29/2018  
 Lin. Date: 11/29/2018  
 Soln. Date: 01/12/2019  
 File Date: 01/12/2019

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 18090

Model No.: CU-34

Serial No.: DDUL S3-0163  
 Expires: 03/13/2020  
 Bottle No.: 0597

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:43S		
Control 1 EC	0.103%	11:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.104%	11:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45S		
Control 2 EC	0.102%	11:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46S		
Control 3 EC	0.102%	11:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature: \_\_\_\_\_



Badge No.: 3200

Date: 01/12/2019