

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02839	Calib. Date: 07/22/2020
Certification File No.:	02840	Calib. No.: 00044
Linearity File No.:	02841	Cert. Date: 07/22/2020
Solution File No.:	02887	Cert. No.: 00034
Sequential File No.:	02887	Lin. Date: 07/22/2020
		Lin. No.: 00033
		Soln. Date: 12/16/2020
		Soln. No.: 00335
		File Date: 12/16/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUK S3-0133
Control Solution %:	0.100%		Expires: 06/10/2021
Solution Control Lot:	19150		Bottle No.: 0042

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:50S		
Control 1 EC	0.099%	05:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:51S		
Control 2 EC	0.100%	05:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:53S		
Control 3 EC	0.099%	05:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	05:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_

Badge No.: 3209

Date: 12/16/2020

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
Location: STRATFORD POLICE DEPT.  
Calibration File No.: 02839 Calib. Date: 07/22/2020 Calib. No.: 00044  
Certification File No.: 02840 Cert. Date: 07/22/2020 Cert. No.: 00034  
Linearity File No.: 02841 Lin. Date: 07/22/2020 Lin. No.: 00033  
Solution File No.: 02877 Soln. Date: 11/19/2020 Soln. No.: 00334  
Sequential File No.: 02877 File Date: 11/19/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133  
Control Solution %: 0.100% Expires: 06/10/2021  
Solution Control Lot: 19150 Bottle No.: 0103

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:10S		
Control 1 EC	0.100%	22:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12S		
Control 2 EC	0.099%	22:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:13S		
Control 3 EC	0.100%	22:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3209

Date: 11/19/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02839	Calib. Date: 07/22/2020
Certification File No.:	02840	Calib. No.: 00044
Linearity File No.:	02841	Cert. Date: 07/22/2020
Solution File No.:	02871	Cert. No.: 00034
Sequential File No.:	02871	Lin. Date: 07/22/2020
		Lin. No.: 00033
		Soln. Date: 10/25/2020
		Soln. No.: 00333
		File Date: 10/25/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUK S3-0133
Control Solution %:	0.100%			Expires:	06/10/2021
Solution Control Lot:	19150			Bottle No.:	0099

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:42S		
Control 1 EC	0.098%	20:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:43S		
Control 2 EC	0.098%	20:44S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:45S		
Control 3 EC	0.098%	20:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_

Badge No.: 3209

Date: 10/25/2020

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0039  
Location: STRATFORD POLICE DEPT.  
Calibration File No.: 02839 Calib. Date: 07/22/2020 Calib. No.: 00044  
Certification File No.: 02840 Cert. Date: 07/22/2020 Cert. No.: 00034  
Linearity File No.: 02841 Lin. Date: 07/22/2020 Lin. No.: 00033  
Solution File No.: 02863 Soln. Date: 10/02/2020 Soln. No.: 00332  
Sequential File No.: 02863 File Date: 10/02/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUK S3-0133  
Control Solution %: 0.100% Expires: 06/10/2021  
Solution Control Lot: 19150 Bottle No.: 0157

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:16D		
Control 1 EC	0.099%	17:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17D		
Control 2 EC	0.099%	17:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		
Control 3 EC	0.099%	17:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: GAGLIARDI

First Name: FRANK

MI: J

Signature: Sgt. Frank Gagliardi #3202

Badge No.: 3202

Date: 10/02/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02839	Calib. Date: 07/22/2020
Certification File No.:	02840	Calib. No.: 00044
Linearity File No.:	02841	Cert. Date: 07/22/2020
Solution File No.:	02856	Cert. No.: 00034
Sequential File No.:	02856	Lin. Date: 07/22/2020
		Lin. No.: 00033
		Soln. Date: 09/06/2020
		Soln. No.: 00331
		File Date: 09/06/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0160

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:28D		
Control 1 EC	0.096%	12:29D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30D		
Control 2 EC	0.097%	12:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31D		
Control 3 EC	0.097%	12:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:33D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature:  5295

Badge No.: 3209

Date: 09/06/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02839	Calib. Date: 07/22/2020
Certification File No.:	02840	Calib. No.: 00044
Linearity File No.:	02841	Cert. Date: 07/22/2020
Solution File No.:	02849	Cert. No.: 00034
Sequential File No.:	02849	Lin. Date: 07/22/2020
		Lin. No.: 00033
		Soln. Date: 08/14/2020
		Soln. No.: 00330
		File Date: 08/14/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUK S3-0133
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0161

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:56D		
Control 1 EC	0.099%	18:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:58D		
Control 2 EC	0.098%	18:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:59D		
Control 3 EC	0.098%	19:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3209

Date: 08/14/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02824	Calib. Date:	03/12/2020
Certification File No.:	02825	Cert. Date:	03/12/2020
Linearity File No.:	02826	Lin. Date:	03/12/2020
Solution File No.:	02838	Soln. Date:	07/09/2020
Sequential File No.:	02838	File Date:	07/09/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0119

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:44D		
Control 1 EC	0.100%	08:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:45D		
Control 2 EC	0.099%	08:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46D		
Control 3 EC	0.099%	08:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3209

Date: 07/09/2020

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location:	STRATFORD POLICE DEPT.	
Calibration File No.:	02824	Calib. No.: 00043
Certification File No.:	02825	Cert. No.: 00033
Linearity File No.:	02826	Lin. No.: 00032
Solution File No.:	02836	Soln. No.: 00327
Sequential File No.:	02836	File Date: 06/15/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %:	0.100%		Expires: 06/10/2021
Solution Control Lot:	19150		Bottle No.: 0117

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:55D		
Control 1 EC	0.102%	05:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	05:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:57D		
Control 2 EC	0.101%	05:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:58D		
Control 3 EC	0.101%	05:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3209

Date: 06/15/2020



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02824	Calib. Date:	03/12/2020
Certification File No.:	02825	Cert. No.:	00043
Linearity File No.:	02826	Cert. Date:	03/12/2020
Solution File No.:	02833	Lin. No.:	00032
Sequential File No.:	02833	Soln. Date:	05/22/2020
		Soln. No.:	00326
		File Date:	05/22/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0100

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:23D		
Control 1 EC	0.100%	16:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:24D		
Control 2 EC	0.099%	16:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:25D		
Control 3 EC	0.100%	16:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MORELLO

First Name: RONALD

MI: M

Signature:  63200

Badge No.: 3200

Date: 05/22/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02824	Calib. Date:	03/12/2020
Certification File No.:	02825	Cert. Date:	03/12/2020
Linearity File No.:	02826	Lin. Date:	03/12/2020
Solution File No.:	02830	Soln. Date:	05/01/2020
Sequential File No.:	02830	File Date:	05/01/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0224

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:22D		
Control 1 EC	0.100%	22:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:23D		
Control 2 EC	0.098%	22:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:25D		
Control 3 EC	0.098%	22:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MJD

Signature: \_\_\_\_\_



Badge No.: 3209

Date: 05/01/2020

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.	Calib. No.:	00043
Calibration File No.:	02824	Cert. Date:	03/12/2020
Certification File No.:	02825	Lin. Date:	03/12/2020
Linearity File No.:	02826	Soln. Date:	04/05/2020
Solution File No.:	02828	File Date:	04/05/2020
Sequential File No.:	02828		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0163
Control Solution %:	0.100%			Expires:	06/10/2021
Solution Control Lot:	19150			Bottle No.:	0210

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:43D		
Control 1 EC	0.100%	10:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44D		
Control 2 EC	0.099%	10:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		
Control 3 EC	0.099%	10:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature:



Badge No.: 3209

Date: 04/05/2020

# Calibrating Unit New Standard Solution Report

**Equipment**      Alcotest 7110 MKIII-C      Serial No.: ARNK-0039  
Location:      STRATFORD POLICE DEPT.  
Calibration File No.: 02772      Calib. Date: 10/07/2019      Calib. No.: 00042  
Certification File No.: 02773      Cert. Date: 10/07/2019      Cert. No.: 00032  
Linearity File No.: 02774      Lin. Date: 10/07/2019      Lin. No.: 00031  
Solution File No.: 02819      Soln. Date: 02/15/2020      Soln. No.: 00322  
Sequential File No.: 02819      File Date: 02/15/2020  
  
Calibrating Unit:      WET      Model No.: CU-34      Serial No.: DDUL S3-0163  
Control Solution %:      0.100%      Expires: 06/10/2021  
Solution Control Lot:      19150      Bottle No.: 0107

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:54S		
Control 1 EC	0.097%	23:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:56S		
Control 2 EC	0.097%	23:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:57S		
Control 3 EC	0.097%	23:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: SCHMIDT


First Name: STEVEN

MI: D

Signature: 

Badge No.: 3209

Date: 02/16/2020

Signature: 

Badge No.: 3209

Date: 02/16/2020

# Calibrating Unit New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C	Serial No.: ARNK-0039
Location: STRATFORD POLICE DEPT.	
Calibration File No.: 02772	Calib. Date: 10/07/2019
Certification File No.: 02773	Cert. No.: 00042
Linearity File No.: 02774	Cert. Date: 10/07/2019
Solution File No.: 02818	Lin. No.: 00031
Sequential File No.: 02818	Soln. No.: 00321
	File Date: 02/15/2020

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0163
Control Solution %: 0.100%		Expires: 06/10/2021
Solution Control Lot: 19150		Bottle No.: 0041


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:29S		
Control 1 EC	----%	22:29S	21.1°C	SIMULATOR TEMP. ERROR
Control 1 IR	----%	22:29S	21.1°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	----%	----		
Control 2 EC	----%	----	N/A	
Control 2 IR	----%	----	N/A	
Ambient Air Blank	----%	----		
Control 3 EC	----%	----	N/A	
Control 3 IR	----%	----	N/A	
Ambient Air Blank	----%	----		


Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT      First Name: STEVEN      MI: D

Signature:       Badge No.: 3209  
Date: 02/15/2020

Signature:       Badge No.: 3209  
Date: 02/15/2020

Signature:       Badge No.: 3209  
Date: 02/15/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02772	Calib. Date:	10/07/2019
Certification File No.:	02773	Cert. Date:	10/07/2019
Linearity File No.:	02774	Lin. Date:	10/07/2019
Solution File No.:	02813	Soln. Date:	02/02/2020
Sequential File No.:	02813	File Date:	02/02/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL 53-0163
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0098

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:14S		
Control 1 EC	0.097%	08:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16S		
Control 2 EC	0.096%	08:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		
Control 3 EC	0.096%	08:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature:



Badge No.: 3209

Date: 02/02/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0039
Location:	STRATFORD POLICE DEPT.		
Calibration File No.:	02772	Calib. Date:	10/07/2019
Certification File No.:	02773	Cert. Date:	10/07/2019
Linearity File No.:	02774	Lin. Date:	10/07/2019
Solution File No.:	02807	Soln. Date:	01/17/2020
Sequential File No.:	02807	File Date:	01/17/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0163
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0164

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:00S		
Control 1 EC	0.099%	23:01S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02S		
Control 2 EC	0.096%	23:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03S		
Control 3 EC	0.099%	23:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHMIDT

First Name: STEVEN

MI: D

Signature: \_\_\_\_\_



Badge No.: 3209

Date: 01/17/2020