

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Stratford Police Department 315 Union Avenue Stratford, NJ 08084	2. Contact: Chief Morello 3. Phone Number: 856-783-8616	4. Date: 01/05/2017 5. County: Camden
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARNK-0039</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
Complaints received regarding error code "(035) EC-SYSTEM" being displayed on the instrument's LCD screen.

10. Comments:
Operational condition of the instrument was assessed by viewing the instrument's LCD screen which displayed "(035) EC-SYSTEM."

Instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 02407

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Trooper II Matthew R. Watson #7078 <small>Name & Badge Number (Print)</small>	 <small>Signature</small>	#7078 1/5/17 <small>Date</small>
--	------------------------------	--

Packing Slip

Customer no Order No Order date
 150053903 10858512 01/19/2017

Packing slip no. Ship date
 81036261 01/19/2017

Please reference on inquiries

Ship to

150053903
 STRATFORD POLICE DEPARTMENT
 ATTN: CHIEF MORELLO
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084-1313
 USA

Payer

Your Purchase Order Number

01/19/2017
 17-00063
 Any questions? Please contact:
 ROSEMARIE HUGHES

Customer

STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084-1313
 USA

Your sales person

CHRISTOPHER FAIRCHILD
 Phone: 866-385-5900
 Fax: 972-929-1260

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110:ARNK-0039 REC'D 1/8/17
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 1 / 0 EA
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.5 / 0.0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address:
 3135 Quarry Road
 Telford, PA 18969 USA
 Toll-free: +1-800-437-2437
 Fax: +1-215-721-5410

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5906

Packing Slip

Customer no Order No Order date
150053903 10858512 01/19/2017

Packing slip no. Ship date
81036261 01/19/2017

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 1 / 0 EA			
0040	2 EA	6808486 PLATES, 7110--9510 Ordered / Back ordered : 2 / 0 EA			
0050	1 EA	MPINS INSURANCE AND HANDLING CHARGE Ordered / Back ordered : 1 / 0 EA			
0060	1 EA	FREIGHT FREIGHT TO DRAEGER Ordered / Back ordered : 1 / 0 EA			
0070	1 EA	FREIGHT FREIGHT TO AGENCY Ordered / Back ordered : 1 / 0 EA			

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Stratford P.D., Stratford, NJ

Date Received: 01/09/2017

Date given to service: 01/09/2017

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR NK-0039

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJH-1556

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 01/18/2017

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Stratford Police Department 315 Union Ave Stratford, NJ 08084	2. Contact: Chief Morello 3. Phone Number: 856-783-8616	4. Date: 01/31/2017 5. County: Camden
--	--	--

6. Alcotest Instrument Serial Number: ARNK-0039	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A
---	---	---

9. Reason for Service:
The Alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:
See Draeger Return and Repair form.

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Matthew R. Watson #7078
 Name & Badge Number (Print)

Tpr. II  #7078
 Signature 01/31/2017
 Date

Packing Slip

Customer no Order No Order date
 150053903 10716582 07/28/2014

Packing slip no. Ship date
 80877314 07/28/2014
 Please reference on inquiries

ship to
 150053903
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

bill to

Your Purchase Order Number
 07/28/2014
 SGT STEPHEN MCBRIDE
 Any questions? Please contact:
 CHARLES ROBINSON II

Customer
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Ground
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		SIM:DDULS3-0163 PROBE:DDUJP2-0303 RECEIVED 07/28/2014
0010	1 EA	MPCERT CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 1 / 0 EA
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 1 / 0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053903 10716582 07/28/2014

Packing slip no. Ship date
80877314 07/28/2014
Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 1 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			
0050	1 EA	4540023 GASKET, SIM JAR, 4"ODX.139"WALL, #240 Ordered / Back ordered : 1 / 0 EA			

Packing Slip

Customer no Order No Order date
 150053903 10625703 05/16/2013

Packing slip no. Ship date
 80774643 05/16/2013

Please reference on inquiries

ship to
 150053903
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

bill to
 [Empty box for bill to information]

Your Purchase Order Number
 05/16/2013
 SGT MCBRIDE
 Any questions? Please contact:
 MARA HERRERA

Customer
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

Your sales person
 Dan Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Ground
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		SIM: DOULS3-10763 PROBE; DOUJP2-030
0010	1 EA	MPCERC CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 1 / 0 EA
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 1 / 0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number 84-1600158

REMIT TO:
 Draeger Safety Diagnostics Inc
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no. Order No. Order date
150053903 10625703 05/16/2013

Packing slip no. Ship date
80774643 05/16/2013
Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 1 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

Packing Slip

Customer no Order No Order date
 150053903 10620010 04/22/2013

Packing slip no. Ship date
 80768292 04/22/2013

Please reference on inquiries

ship to

150053903
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

bill to

Your Purchase Order Number

04/22/2013
 04222013
 Any questions? Please contact:
 BRENDA BROTMAN

Customer
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Ground			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
0010	20 EA	6805700 MOUTHPIECE, ALCO 7110/7410/9510 PKG25 Ordered / Back ordered : 20 / 0 EA			
0020	25 EA	4407061 CERT SOLUTION, .10% NEW JERSEY CERTIFIED Ordered / Back ordered : 25 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Translt Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053903 10620010 04/22/2013

Packing slip no. Ship date
80768292 04/22/2013

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

Packing Slip

Customer no Order No Order date
 150053903 10566164 08/03/2012

Packing slip no. Ship date
 80706759 08/03/2012

Please reference on inquiries

ship to
 150053903
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

bill to
 [Empty box for bill to information]

Your Purchase Order Number
 08/03/2012
 710: ARNK-0039
 Any questions? Please contact:
 MARA HERRERA

Customer
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 PPA
 PREPAY AND ADD

Item#	sh. Quant.	Part no. Description			
		Waybill number: 1001891730480000808400539788182026 Ship via: Fed Ex P1			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARNK-0039			
0010	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 1 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 8209426615
 Accl. Name: Draeger Safety Diagnostics Inc.
 Translt Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no: Order No Order date
150053903 10566164 08/03/2012

Packing slip no. Ship date
80706759 08/03/2012

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant	Part no. Description			
0020	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.5 / 0.0 H			

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: Stratford Borough P.D., Stratford, NJ

Date Received: 08/03/12

Date given to service: 08/03/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR NK-0039

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Unit need recalibration , recalibrated unit.

CAL W/QC AND OPS CHECK

Service Technician: JS

Date: 08-03-12

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Stratford Borough Police Department 315 Union Avenue Stratford NJ, 08084	2. Contact: Chief Morello 3. Phone Number: 856-783-8616	4. Date: 8-2-2012 5. County: Camden
6. Alcohol Instrument Serial Number: <p style="text-align: center;">ARNK-0039</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
Complaint received regarding low IR and EC results.

10. Comments:
While attempting to calibrate the instrument, the instrument displayed "Ctrl Gas Supply" on the LED screen and would not allow a calibration to be completed.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required
12. The above Instrument/Component is placed out of service pending further evaluation
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willfully false, I am subject to punishment."

Tpr. II Michael P. Gibson #6353
 Name & Badge Number (Print)

Michael P. Gibson #6353 8/2/12
 Signature Date

Packing Slip

Customer no Order No Order date
150053903 10553635 06/05/2012

Packing slip no. Ship date
80692066 06/05/2012

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 1 / 0 EA PROBE DDUNP2-316 REPLACED WITH DDLLP1-0015			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Stratford PD 315 Union Ave Stratford, NJ 08084	2. Contact: Chief Morello 3. Phone Number: 856-783-8616	4. Date: 1-19-2011 5. County: Camden
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARNK-0039</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
The Alcotest 7110 came back from outside evaluation from Draeger and is returned to service.

10. Comments:
See attached packing slip for details of Draeger's work conducted on the instrument.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F. Sullivan #5103
 Name & Badge Number (Print)

Tpr I F Sullivan # 5103 1-19-11
 Signature Date

Packing Slip

Customer no Order No Order date
 150053903 10451542 01/12/2011


Packing slip no. Ship date
 80571768 01/12/2011

Please reference on inquiries

ship to

150053903
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

bill to



Your Purchase Order Number

7110 WARRANTY: ARNK-0039
 Any questions? Please contact:
 MARA HERRERA

Customer

STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARNK-0039			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED			
0020	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150053903 10451542 01/12/2011

Packing slip no. Ship date
80571768 02/12/2011

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Stratford Twp. PD 315 Union Ave Stratford, NJ 08084	2. Contact: Chief Morello 3. Phone Number: 856-783-8616	4. Date: 1-5-2011 5. County: Camden
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARNK-0039</p>	7. Simulator Component Serial Number: <p style="text-align: center;">n/a</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">n/a</p>

9. Reason for Service:
Fuel cell would not recovery after control tests.

10. Comments:
Instrument was sent back to Draeger for further evaluation.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F. Sullivan #5103 <hr/> <small>Name & Badge Number (Print)</small>	 <hr/> <small>Signature</small>	1-5-2011 <hr/> <small>Date</small>
---	--	--

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Stratford PD 315 Union Ave Stratford, NJ 08084	2. Contact: Chief Morello 3. Phone Number: 856-783-8616	4. Date: 1-19-2011 5. County: Camden
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARNK-0039</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
The Alcotest 7110 came back from outside evaluation from Draeger and is returned to service.

10. Comments:
See attached packing slip for details of Draeger's work conducted on the instrument.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F. Sullivan #5103
 Name & Badge Number (Print)

Tpr I F Sullivan # 5103 1-19-11
 Signature Date

Packing Slip

Customer no Order No Order date
 150053903 10776459 05/06/2015

Packing slip no. Ship date
 80945255 05/06/2015

Please reference on inquiries

Ship to

150053903
 STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

Payer

Your Purchase Order Number

05/06/2015
 SGT. STEPHEN MCBRIDE
 Any questions? Please contact:
 ROSEMARIE HUGHES

Customer

STRATFORD POLICE DEPARTMENT
 CAMDEN COUNTY
 315 UNION AVENUE
 STRATFORD, NJ 08084
 USA

Your sales person

LINDA SALO
 Phone: 866-385-5900
 Fax: 972-929-1260

Delivery terms

FA
 - FREIGHT ALLOWED.

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Ground			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		SIM: DDUKS3-0133, PROBE DDLLP1-0015 REC'D 4/24/15			
0010	1 EA	MPCERT CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 1 / 0 EA			
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 1 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5906

Packing Slip

Customer no Order No Order date
150053903 10776459 05/06/2015

Packing slip no. Ship date
80945255 05/06/2015

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no, Description			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 1 / 0 EA			
0040	1 EA	4540023 GASKET, SIM JAR, 4"ODX.139"WALL, #240 Ordered / Back ordered : 1 / 0 EA			
0050	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			