

ALCOTEST CHECKLIST

Municipality: WESTAMPTON

Alcotest Ser. #: ARNK-0025

County: BURLINGTON

Date of Calibration: 07-30-2008

1. **Certificate of Accuracy Alcotest 7110 MKIII-C** from Draeger Safety for instrument used in the A.I.R.
2. **Certificate of Accuracy CU34 Unit** on Alcotest Instrument used.
Ser. #: DDUT S3-0117
3. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety for instrument in A.I.R. or equivalent.
Ser. #: DDWA P2-116
4. **Digital Temperature Measuring System Report of Calibration.**
Ser. #: ERTCO-HART A71569
5. A. Alcotest 7110 Calibration Record
 B. Alcotest 7110 Calibration Certificate **Part I - Control Tests.**
 C. Alcotest 7110 Calibration Certificate **Part II - Linearity Tests.**
 D. **Alcotest Card** of operator/coordinator who performed tests.
 E. **Certificate of Accuracy Alcotest 7110 Temperature Probe** from Draeger Safety used in the Calibration Tests ["Black Key" probe of Breath Test Coordinator].
Ser. #: DDWA P2-001
6. **Certificates of Analysis** for each **Simulator Solution** used in Calibration/Linearity Tests:
- | | | | | |
|--|-----------------|-----------|----------|------------|
| <input checked="" type="checkbox"/> A. | 0.04% Solution. | <u>07</u> | <u>A</u> | <u>041</u> |
| <input checked="" type="checkbox"/> B. | 0.08% Solution. | <u>07</u> | <u>A</u> | <u>042</u> |
| <input checked="" type="checkbox"/> C. | 0.10% Solution. | <u>07</u> | <u>B</u> | <u>045</u> |
| <input checked="" type="checkbox"/> D. | 0.16% Solution. | <u>07</u> | <u>A</u> | <u>043</u> |
7. Certificate of Accuracy Alcotest CU34 Simulators from Draeger Safety (when conducting the Calibration/Linearity Tests) for:
- A. 0.04% used in Calibration/Linearity Testing. DDMK S3-0003
- B. 0.08% used in Calibration/Linearity Testing. DDRK S3-0005
- C. 0.10% used in Calibration/Linearity Testing DDUS S3-0117 [Same as instrument's CU34 unit].
- D. 0.16% used in Calibration/Linearity Testing. DDRF S3-0009
8. A. New Standard Solution Report following Calibration.
 B. Calibrating CU34 Unit for same [same as CU34 unit on instrument].
 C. Certificate of Analysis 0.10% solution for same.
Lot #: 07 J 048
 D. Alcotest card of operator/coordinator who completed change.

Draegersafety

Alcotest[®] 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 53 FR 43705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below or sooner according to your State Specifications.

Certification Date:

01 SEPT 04

SERIAL NUMBER:

AP0K - 0025

Draeger Safety Diagnostics, Inc.
Durango, CO





Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDUJS3-0117

Certification Date

Technician

Re-Certification Due Date

07-23-08

AS

07-23-09

Drägersafety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDWAP2-116

Certification date:

07-23-08

Next Certification due:

07-23-09

Probe Value

102

Dräger Safety Diagnostics, Inc.
Technical Service Department

JS

Ertco-Hart Digital Temperature Measuring System

REPORT OF CALIBRATION

This is to certify that the Ertco-Hart Digital Temperature Measuring System has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). Draeger Safety Diagnostics, Inc. (DSDI) recommends accuracy verification of the Ertco-Hart Digital Temperature Measuring System within 12 months of the certification date below, or sooner, according to your state specification.

DSDI equipment used for temperature verification Serial Number: HH41 301316 / HH41 303176 / MERCURY THERMOMETER

Digital Units Serial Number: A71569

Probe Serial Number: 765077

Certification Date: 02-12-2008

Next Certification Due: 02-12-2009

At 34.00 °C digital unit displays 34.05 °C

Draeger Safety Diagnostics, Inc. Technician: 

Alcotest 7110 Calibration Record

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00208 Calib. Date: 07/30/2008 Calib. No.: 00005
Certification File No.: 00151 Cert. Date: 09/04/2007 Cert. No.: 00003
Linearity File No.: 00152 Lin. Date: 09/04/2007 Lin. No.: 00003
Solution File No.: 00205 Soln. Date: 07/04/2008 Soln. No.: 00050
Sequential File No.: 00208 File Date: 07/30/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 02/22/2009
Solution Control Lot: 07B045 Bottle No.: 0960

Coordinator

Last Name: CROSS First Name: WILLIAM MI: M
Signature: TPR. I Will M. Cross #5342 Badge No.: 5342
Date: 07/30/2008

*Black Key Temperature Probe Serial# D11WAP2-001 we
*Ertco-Hart Digital Temperature Measuring System Serial# A71569 we

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00208 Calib. Date: 07/30/2008 Calib. No.: 00005
Certification File No.: 00209 Cert. Date: 07/30/2008 Cert. No.: 00004
Linearity File No.: 00152 Lin. Date: 09/04/2007 Lin. No.: 00003
Solution File No.: 00205 Soln. Date: 07/04/2008 Soln. No.: 00050
Sequential File No.: 00209 File Date: 07/30/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 02/22/2009
Solution Control Lot: 07B045 Bottle No.: 0960

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:14D		
Control 1 EC	0.099%	09:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:16D		
Control 2 EC	0.099%	09:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:17D		
Control 3 EC	0.097%	09:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18D		

All tests within acceptable tolerance.

Coordinator

Last Name: CRQSS

First Name: WILLIAM

MI: M

Signature: _____

TPR J. William M. Criss #5342

Badge No.: 5342

Date: 07/30/2008

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00208 Calib. Date: 07/30/2008 Calib. No.: 00005
Certification File No.: 00209 Cert. Date: 07/30/2008 Cert. No.: 00004
Linearity File No.: 00210 Lin. Date: 07/30/2008 Lin. No.: 00004
Solution File No.: 00205 Soln. Date: 07/04/2008 Soln. No.: 00050
Sequential File No.: 00210 File Date: 07/30/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDMK S3-0003
Control Solution %: 0.040% Expires: 01/26/2009
Solution Control Lot: 07A041 Bottle No.: 0553

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDRK S3-0005
Control Solution %: 0.080% Expires: 01/26/2009
Solution Control Lot: 07A042 Bottle No.: 0444

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDRF S3-0009
Control Solution %: 0.160% Expires: 01/26/2009
Solution Control Lot: 07A043 Bottle No.: 0413

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:34D		
Control 1 EC	0.040%	09:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.040%	09:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36D		
Control 2 EC	0.040%	09:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.040%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 3 EC	0.079%	09:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.079%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 4 EC	0.079%	09:41D	34.0°C	*** TEST PASSED ***
Control 4 IR	0.078%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42D		
Control 5 EC	0.158%	09:43D	34.0°C	*** TEST PASSED ***
Control 5 IR	0.158%	09:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44D		
Control 6 EC	0.157%	09:45D	34.0°C	*** TEST PASSED ***
Control 6 IR	0.159%	09:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47D		

All tests within acceptable tolerance.

Coordinator

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: TPR. I WILLIAM CROSS #5342

Badge No.: 5342

Date: 07/30/2008

DEPARTMENT OF
Water and Public Safety
 This is to certify that

William M. Cross

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 148 OF THE LAWS OF 1986 IN THE OPERATION OF THE Breathalyzer METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 31st DAY OF Jan.
 ONE THOUSAND NINE HUNDRED AND 97

Cal A. Williams
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

Peter Weimer
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

DEPARTMENT OF
Water and Public Safety
 This is to certify that

William M. Cross
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 148 OF THE LAWS OF 1986 IN THE OPERATION OF THE Alcotest 7110 MKIII-C METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF February
 ONE THOUSAND AND 05

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refreshers Course PLACE	INSTRUCTOR
1.	2-11-99	GCDA	<i>[Signature]</i>
2.	10-10-01	BCPA	<i>[Signature]</i>
3.	4-17-03	BCDA	<i>[Signature]</i>
4.	6-22-05	ACTC	<i>[Signature]</i>
5.	7-27-07	OLPA	<i>[Signature]</i>
6.			
7.			
8.			
9.			

SP-293B

ORIGINAL COURSE DATES

	DATE	Refreshers Course PLACE	INSTRUCTOR
1.	12-20-07	O.C.P.A.	<i>[Signature]</i>
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 12/04)



CERTIFICATE

This is to certify that


William M. Cross
Trooper II #5342

has successfully completed the two day Draeger Safety Diagnostics, Inc. Operator Training and Preventive Maintenance Course on the New Jersey specific Alcotest® 7110 MKIII-C and is hereby certified as a qualified

**Operator Trainer and
Maintenance Technician**

Completion of this course qualifies this individual to train and certify Operators in the proper use and operation as well as perform Preventive Maintenance on the New Jersey specific Alcotest® 7110 MKIII-C.

Date: October 28, 2004



Instructor: Hansueli Ryser

Draegersafety

Drägersafety

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDWAP2-001

02-12-08

02-12-09

Probe Value

101

Dräger Safety Diagnostics, Inc.
Technical Service Department





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 West Trenton NJ 08628-0068
 (609) 882-2000

RON B. CORZINE
 Governor

STUART RABNER
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.040 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.045 to 0.051 grams per 100 milliliters of solution.

MANUFACTURER: Dräger Safety, Inc. **ANALYSIS DATE:** 03/06/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07A041

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0483 to 0.0489 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 26, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare

Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 29th day of March, 2007.

Linda L. DeBartis
 Notary

Linda L. DeBartis
 My Commission
 Expires Aug. 17, 2008



New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 West Trenton NJ 08628-0068
 (609) 882-2000

STUART RABNER
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

JOHN S. CORZINE
 Governor

CERTIFICATION OF ANALYSIS
0.080 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.094 to 0.099 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. **ANALYSIS DATE:** 03/06/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07A042

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.0965 to 0.0972 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 26, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungarc

Ajit R. Tungarc
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 29th day of March, 2007.

Linda L. DeSantis
 Notary

Linda L. DeSantis
 My Commission
 Expires Aug. 17, 2009





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

JON S. CORZINE
 Governor

STUART RABNER
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 04/04/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07B045

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1222 to 0.1227 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is February 22, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 20th day of April, 2007.

Notary

Linda L. DeSantis
 My Commission
 Expires Aug. 17, 2009





State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 Post Office Box 7068
 West Trenton NJ 08628-0068
 (609) 882-2000

STUART RABNER
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

RON S. CORZINE
 Governor

CERTIFICATION OF ANALYSIS
0.160 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.188 to 0.199 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc. **ANALYSIS DATE:** 03/06/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07A043

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1932 to 0.1938 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 26, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Linda L. DeSantis
 My Commission
 Expires Aug. 17, 2008

Ajit R. Tungare
 Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 29th day of March, 2007.

Linda L. DeSantis
 Notary



New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable





Dräger safety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST® CU34

Model: MARK IIA

Other: _____

Serial Number:

DDMK 53-0003

Certification Date

02-13-08

Technician

YE

Re-Certification Due Date

02-13-09



Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST[®] CU34

Model: MARK IIA

Other: _____

Serial Number:

DDRK 53-0005

Certification Date

02-13-08

Technician

[Signature]

Re-Certification Due Date

02-13-09



Drägersafety

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.

(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

Model: ALCOTEST[®] CU34

Model: MARK IIA

Other: _____

Serial Number:

DDRF53-0009

Certification Date

02-13-08

Technician

[Signature]

Re-Certification Due Date

02-13-09

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00208	Calib. Date: 07/30/2008
Certification File No.:	00209	Calib. No.: 00005
Linearity File No.:	00210	Cert. Date: 07/30/2008
Solution File No.:	00211	Cert. No.: 00004
Sequential File No.:	00211	Lin. Date: 07/30/2008
		Lin. No.: 00004
		Soln. Date: 07/30/2008
		Soln. No.: 00051
		File Date: 07/30/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 0589

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:57D		
Control 1 EC	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		
Control 2 EC	0.098%	10:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 3 EC	0.098%	11:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWAP2-116 WL

Changed By:

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: _____

TPR. I WILLIAM CROSS #5342

Badge No.: 5342

Date: 07/30/2008



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

JON S. CORZINE
 Governor

ANNE MILGRAM
 Attorney General

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 11/28/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07J048

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1225 to 0.1229 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is October 3, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
 Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 17th day of December, 2007.

Notary

Linda L. DeSantis
 My Commission
 Expires Aug. 17, 2009



New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable

