

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00550 Calib. Date: 08/21/2013 Calib. No.: 00018
Certification File No.: 00551 Cert. Date: 08/21/2013 Cert. No.: 00015
Linearity File No.: 00552 Lin. Date: 08/21/2013 Lin. No.: 00016
Solution File No.: 00576 Soln. Date: 12/31/2013 Soln. No.: 00142
Sequential File No.: 00576 File Date: 12/31/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 1264

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:43S		
Control 1 EC	0.099%	11:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45S		
Control 2 EC	0.098%	11:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46S		
Control 3 EC	0.098%	11:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BREWER

First Name: ANDREW

MI: J

Signature: Sgt. A. Brewer

Badge No.: 2739

Date: 12/31/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00550 Calib. Date: 08/21/2013 Calib. No.: 00018
Certification File No.: 00551 Cert. Date: 08/21/2013 Cert. No.: 00015
Linearity File No.: 00552 Lin. Date: 08/21/2013 Lin. No.: 00016
Solution File No.: 00568 Soln. Date: 12/15/2013 Soln. No.: 00141
Sequential File No.: 00568 File Date: 12/15/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 1254

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:56S		
Control 1 EC	0.098%	03:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:57S		
Control 2 EC	0.097%	03:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:59S		
Control 3 EC	0.097%	03:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FERGUSON

First Name: BRIAN

MI: C

Signature: _____



#2736

Badge No.: 2736

Date: 12/15/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00550 **Calib. Date:** 08/21/2013 **Calib. No.:** 00018
Certification File No.: 00551 **Cert. Date:** 08/21/2013 **Cert. No.:** 00015
Linearity File No.: 00552 **Lin. Date:** 08/21/2013 **Lin. No.:** 00016
Solution File No.: 00564 **Soln. Date:** 11/26/2013 **Soln. No.:** 00140
Sequential File No.: 00564 **File Date:** 11/26/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0116
Control Solution %: 0.100% **Expires:** 11/05/2014
Solution Control Lot: 12K108 **Bottle No.:** 1251

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:16S		
Control 1 EC	0.099%	10:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17S		
Control 2 EC	0.097%	10:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19S		
Control 3 EC	0.097%	10:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 11/26/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00550	Calib. Date: 08/21/2013
Certification File No.:	00551	Cert. No.: 00018
Linearity File No.:	00552	Cert. Date: 08/21/2013
Solution File No.:	00560	Lin. No.: 00015
Sequential File No.:	00560	Lin. Date: 08/21/2013
		Soln. No.: 00139
		File Date: 11/01/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1269

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:01S		
Control 1 EC	0.100%	09:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03S		
Control 2 EC	0.099%	09:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04S		
Control 3 EC	0.097%	09:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT	First Name: STEPHEN	MI: P
Signature: <u>Sen. Stephen [Signature]</u> #2730	Badge No.: 2730	Date: 11/01/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00550 Calib. Date: 08/21/2013 Calib. No.: 00018
Certification File No.: 00551 Cert. Date: 08/21/2013 Cert. No.: 00015
Linearity File No.: 00552 Lin. Date: 08/21/2013 Lin. No.: 00016
Solution File No.: 00559 Soln. Date: 10/04/2013 Soln. No.: 00138
Sequential File No.: 00559 File Date: 10/04/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 1258

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:59D		
Control 1 EC	0.100%	09:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		
Control 2 EC	0.098%	10:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 3 EC	0.099%	10:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

M1: P

Signature: _____

Sgt. *Step B* → 2730

Badge No.: 2730

Date: 10/04/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00550	Calib. Date: 08/21/2013
Certification File No.:	00551	Calib. No.: 00018
Linearity File No.:	00552	Cert. Date: 08/21/2013
Solution File No.:	00557	Cert. No.: 00015
Sequential File No.:	00557	Lin. Date: 08/21/2013
		Lin. No.: 00016
		Soln. Date: 09/07/2013
		Soln. No.: 00137
		File Date: 09/07/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1255

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:05D		
Control 1 EC	0.098%	08:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07D		
Control 2 EC	0.097%	08:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08D		
Control 3 EC	0.097%	08:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 09/07/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00520	Calib. Date: 03/13/2013
Certification File No.:	00521	Calib. No.: 00017
Linearity File No.:	00522	Cert. Date: 03/13/2013
Solution File No.:	00548	Cert. No.: 00014
Sequential File No.:	00548	Lin. Date: 03/13/2013
		Lin. No.: 00015
		Soln. Date: 08/19/2013
		Soln. No.: 00135
		File Date: 08/19/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1265

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:00D		
Control 1 EC	0.098%	21:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:01D		
Control 2 EC	0.097%	21:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		
Control 3 EC	0.097%	21:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

M1: P

Signature: _____



Badge No.: 2730

Date: 08/19/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00520	Calib. Date: 03/13/2013
Certification File No.:	00521	Calib. No.: 00017
Linearity File No.:	00522	Cert. Date: 03/13/2013
Solution File No.:	00544	Cert. No.: 00014
Sequential File No.:	00544	Lin. Date: 03/13/2013
		Lin. No.: 00015
		Soln. Date: 08/11/2013
		Soln. No.: 00134
		File Date: 08/11/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1266

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:33D		
Control 1 EC	0.098%	14:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:34D		
Control 2 EC	0.097%	14:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:35D		
Control 3 EC	0.097%	14:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 08/11/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00520	Calib. Date:	03/13/2013
Certification File No.:	00521	Cert. Date:	03/13/2013
Linearity File No.:	00522	Lin. Date:	03/13/2013
Solution File No.:	00541	Soln. Date:	07/15/2013
Sequential File No.:	00541	File Date:	07/15/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0116
Solution Control Lot:	12K108	Expires:	11/05/2014
		Bottle No.:	1263

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:15D		
Control 1 EC	0.099%	08:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		
Control 2 EC	0.097%	08:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17D		
Control 3 EC	0.097%	08:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

M1: P

Signature: 

Badge No.: 2730

Date: 07/15/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00520	Calib. Date: 03/13/2013
Certification File No.:	00521	Calib. No.: 00017
Linearity File No.:	00522	Cert. Date: 03/13/2013
Solution File No.:	00538	Cert. No.: 00014
Sequential File No.:	00538	Lin. Date: 03/13/2013
		Lin. No.: 00015
		Soln. Date: 06/19/2013
		Soln. No.: 00132
		File Date: 06/19/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1261

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:18D		
Control 1 EC	0.097%	20:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:19D		
Control 2 EC	0.097%	20:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:21D		
Control 3 EC	0.097%	20:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

M1: P

Signature: _____



Badge No.: 2730

Date: 06/19/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00520 Calib. Date: 03/13/2013 Calib. No.: 00017
Certification File No.: 00521 Cert. Date: 03/13/2013 Cert. No.: 00014
Linearity File No.: 00522 Lin. Date: 03/13/2013 Lin. No.: 00015
Solution File No.: 00535 Soln. Date: 05/28/2013 Soln. No.: 00131
Sequential File No.: 00535 File Date: 05/28/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 1275

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:07D		
Control 1 EC	0.098%	02:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:08D		
Control 2 EC	0.097%	02:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:10D		
Control 3 EC	0.097%	02:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 05/28/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00520 Calib. Date: 03/13/2013 Calib. No.: 00017
Certification File No.: 00521 Cert. Date: 03/13/2013 Cert. No.: 00014
Linearity File No.: 00522 Lin. Date: 03/13/2013 Lin. No.: 00015
Solution File No.: 00531 Soln. Date: 05/08/2013 Soln. No.: 00130
Sequential File No.: 00531 File Date: 05/08/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 1252 Bottle No.: 1252

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:55D		
Control 1 EC	0.099%	20:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57D		
Control 2 EC	0.097%	20:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:58D		
Control 3 EC	0.099%	20:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 05/08/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00520	Calib. Date: 03/13/2013
Certification File No.:	00521	Calib. No.: 00017
Linearity File No.:	00522	Cert. Date: 03/13/2013
Solution File No.:	00528	Cert. No.: 00014
Sequential File No.:	00528	Lin. Date: 03/13/2013
		Lin. No.: 00015
		Soln. Date: 04/10/2013
		Soln. No.: 00129
		File Date: 04/10/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1273

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:58D		
Control 1 EC	0.099%	08:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00D		
Control 2 EC	0.096%	09:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01D		
Control 3 EC	0.097%	09:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 04/10/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00492 Calib. Date: 11/15/2012 Calib. No.: 00016
Certification File No.: 00493 Cert. Date: 11/15/2012 Cert. No.: 00013
Linearity File No.: 00494 Lin. Date: 11/15/2012 Lin. No.: 00014
Solution File No.: 00512 Soln. Date: 02/08/2013 Soln. No.: 00126
Sequential File No.: 00512 File Date: 02/08/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 1253

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:52S		
Control 1 EC	0.099%	19:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54S		
Control 2 EC	0.098%	19:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55S		
Control 3 EC	0.098%	19:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 02/08/2013

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00492	Calib. Date:	11/05/2012
Certification File No.:	00493	Cert. No.:	00016
Linearity File No.:	00494	Calib. Date:	11/15/2012
Solution File No.:	00509	Cert. No.:	00013
Sequential File No.:	00509	Lin. Date:	11/15/2012
		Lin. No.:	00014
		Soln. Date:	01/11/2013
		Soln. No.:	00125
		File Date:	01/11/2013
Calibrating Unit:	WE1	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0117
Solution Control Lot:	12K108	Expires:	11/05/2014
		Bottle No.:	1250

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:39S		
Control 1 EC	0.100%	11:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41S		
Control 2 EC	0.097%	11:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42S		
Control 3 EC	0.098%	11:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MU: P

Signature: _____



Badge No.: 2730

Date: 01/11/2013