

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00627	Calib. Date: 10/06/2014
Certification File No.:	00628	Calib. No.: 00021
Linearity File No.:	00629	Cert. Date: 10/06/2014
Solution File No.:	00640	Cert. No.: 00017
Sequential File No.:	00640	Lin. Date: 10/06/2014
		Lin. No.: 00018
		Soln. Date: 12/08/2014
		Soln. No.: 00154
		File Date: 12/08/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1318

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:05S		
Control 1 EC	0.101%	17:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:06S		
Control 2 EC	0.100%	17:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:07S		
Control 3 EC	0.100%	17:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_

*Stephen Ent* 2730

Badge No.: 2730

Date: 12/08/2014

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00627	Calib. Date: 10/06/2014
Certification File No.:	00628	Calib. No.: 00021
Linearity File No.:	00629	Cert. Date: 10/06/2014
Solution File No.:	00637	Cert. No.: 00017
Sequential File No.:	00637	Lin. Date: 10/06/2014
		Lin. No.: 00018
		Soln. Date: 11/03/2014
		Soln. No.: 00153
		File Date: 11/03/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1312

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:55S		
Control 1 EC	0.100%	15:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:57S		
Control 2 EC	0.099%	15:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58S		
Control 3 EC	0.100%	15:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ENT	First Name: STEPHEN	MI: P
Signature: 		Badge No.: 2730
		Date: 11/03/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00627	Calib. Date: 10/06/2014
Certification File No.:	00628	Calib. No.: 00021
Linearity File No.:	00629	Cert. Date: 10/06/2014
Solution File No.:	00630	Cert. No.: 00017
Sequential File No.:	00630	Lin. Date: 10/06/2014
		Lin. No.: 00018
		Soln. Date: 10/06/2014
		Soln. No.: 00152
		File Date: 10/06/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1321

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:44D		
Control 1 EC	0.100%	11:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 2 EC	0.100%	11:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		
Control 3 EC	0.100%	11:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWAP 2-116 (ME)

**Changed By:**

Last Name: GONCALVES                      First Name: MICHELLE                      MI: L

Signature: Sgt. Michelle Gonalves #6040                      Badge No.: 6040

Date: 10/06/2014

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0025  
Location: WESTAMPTON POLICE  
Calibration File No.: 00589 Calib. Date: 04/14/2014 Calib. No.: 00020  
Certification File No.: 00590 Cert. Date: 04/14/2014 Cert. No.: 00016  
Linearity File No.: 00591 Lin. Date: 04/14/2014 Lin. No.: 00017  
Solution File No.: 00624 Soln. Date: 09/08/2014 Soln. No.: 00151  
Sequential File No.: 00624 File Date: 09/08/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117  
Control Solution %: 0.100% Expires: 11/12/2015  
Solution Control Lot: 1307 Bottle No.: 1307

Function	Result	Time	Temperature	Comment(s) or Error(s)
	%BAC	HH:MM	Simulator (°C)	
Ambient Air Blank	0.000%	11:34D		
Control 1 EC	0.098%	11:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 2 EC	0.098%	11:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37D		
Control 3 EC	0.098%	11:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ENT

First Name: STEPHEN

MJ: P

Signature: \_\_\_\_\_

 2730

Badge No.: 2730

Date: 09/08/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00589	Calib. Date: 04/14/2014
Certification File No.:	00590	Calib. No.: 00020
Linearity File No.:	00591	Cert. Date: 04/14/2014
Solution File No.:	00619	Cert. No.: 00016
Sequential File No.:	00619	Lin. Date: 04/14/2014
		Lin. No.: 00017
		Soln. Date: 08/12/2014
		Soln. No.: 00150
		File Date: 08/12/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 1271

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:12D		
Control 1 EC	0.097%	10:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		
Control 2 EC	0.097%	10:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		
Control 3 EC	0.097%	10:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_

Sgt. Stephen #2730

Badge No.: 2730

Date: 08/12/2014

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0025  
**Location:** WESTAMPTON POLICE  
**Calibration File No.:** 00589 **Calib. Date:** 04/14/2014 **Calib. No.:** 00020  
**Certification File No.:** 00590 **Cert. Date:** 04/14/2014 **Cert. No.:** 00016  
**Linearity File No.:** 00591 **Lin. Date:** 04/14/2014 **Lin. No.:** 00017  
**Solution File No.:** 00612 **Soln. Date:** 07/14/2014 **Soln. No.:** 00149  
**Sequential File No.:** 00612 **File Date:** 07/14/2014

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0117  
**Control Solution %:** 0.100% **Expires:** 11/05/2014  
**Solution Control Lot:** 12K108 **Bottle No.:** 1259

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:32D		
Control 1 EC	0.095%	10:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	10:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33D		
Control 2 EC	0.096%	10:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	10:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35D		
Control 3 EC	0.096%	10:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	10:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_

Sgt. Stephen P. # 2730

Badge No.: 2730

Date: 07/14/2014

# Calibrating Unit

## New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00589	Calib. Date:	04/14/2014
Certification File No.:	00590	Cert. Date:	04/14/2014
Linearity File No.:	00591	Lin. Date:	04/14/2014
Solution File No.:	00608	Soln. Date:	06/25/2014
Sequential File No.:	00608	File Date:	06/25/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUI S3-0117
Solution Control Lot:	12K108	Expires:	11/05/2014
		Bottle No.:	1267

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:51D		
Control 1 EC	0.096%	14:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	14:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:53D		
Control 2 EC	0.096%	14:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	14:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		
Control 3 EC	0.096%	14:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	14:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_



Badge No.: 2730

Date: 06/25/2014

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0025  
Location: WESTAMPTON POLICE  
Calibration File No.: 00589 Calib. Date: 04/14/2014 Calib. No.: 00020  
Certification File No.: 00590 Cert. Date: 04/14/2014 Cert. No.: 00016  
Linearity File No.: 00591 Lin. Date: 04/14/2014 Lin. No.: 00017  
Solution File No.: 00602 Soln. Date: 05/29/2014 Soln. No.: 00147  
Sequential File No.: 00602 File Date: 05/29/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117  
Control Solution %: 0.100% Expires: 11/05/2014  
Solution Control Lot: 12K108 Bottle No.: 1274

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:26D		
Control 1 EC	0.097%	05:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:27D		
Control 2 EC	0.097%	05:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	05:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28D		
Control 3 EC	0.097%	05:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_

*Steph. ENT-2730*

Badge No.: 2730

Date: 05/29/2014



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00589	Calib. Date:	04/14/2014
Certification File No.:	00590	Cert. No.:	00020
Linearity File No.:	00591	Calib. No.:	00016
Solution File No.:	00594	Lin. Date:	04/14/2014
Sequential File No.:	00594	Lin. No.:	00017
		Soln. Date:	05/09/2014
		Soln. No.:	00146
		File Date:	05/09/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0117
Solution Control Lot:	12K108	Expires:	11/05/2014
		Bottle No.:	1257

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:07D		
Control 1 EC	0.096%	20:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	20:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:08D		
Control 2 EC	0.096%	20:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	20:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09D		
Control 3 EC	0.096%	20:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	20:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_



Badge No.: 2730

Date: 05/09/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment:</b> Alcotest 7110 MKIII-C	Serial No.: ARNK-0027	
<b>Location:</b> WESTAMPTON POLICE		
<b>Calibration File No.:</b> 00550	<b>Calib. Date:</b> 08/21/2013	<b>Calib. No.:</b> 00018
<b>Certification File No.:</b> 00551	<b>Cert. Date:</b> 08/21/2013	<b>Cert. No.:</b> 00015
<b>Linearity File No.:</b> 00552	<b>Lin. Date:</b> 08/21/2013	<b>Lin. No.:</b> 00016
<b>Solution File No.:</b> 00579	<b>Soln. Date:</b> 01/21/2014	<b>Soln. No.:</b> 00140
<b>Sequential File No.:</b> 00579	<b>File Date:</b> 01/21/2014	
<b>Calibrating Unit:</b> WET	<b>Model No.:</b> CU-04	<b>Serial No.:</b> DDU/30-0116
<b>Control Solution %:</b> 0.100%		<b>Expires:</b> 11/05/2014
<b>Solution Control Lot:</b> 12K108		<b>Batch No.:</b> 1270

Function	Result %DAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:44S		
Control 1 EC	0.099%	11:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45S		
Control 2 EC	0.096%	11:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:47S		
Control 3 EC	0.097%	11:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48S		

All tests within acceptable tolerance.

On 01/21/14, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_



Badge No.: ENT

Date: 01/21/2014