

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARNK-0025 |
| Location: | WESTAMPTON POLICE | |
| Calibration File No.: | 00919 | Calib. Date: 08/07/2019 |
| Certification File No.: | 00920 | Calib. No.: 00032 |
| Linearity File No.: | 00921 | Cert. Date: 08/07/2019 |
| Solution File No.: | 00942 | Cert. No.: 00028 |
| Sequential File No.: | 00942 | Lin. Date: 08/07/2019 |
| | | Lin. No.: 00029 |
| | | Soln. Date: 12/07/2019 |
| | | Soln. No.: 00229 |
| | | File Date: 12/07/2019 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUJ S3-0117 |
| Solution Control Lot: | 19150 | Expires: 06/10/2021 |
| | | Bottle No.: 0746 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:04S | | |
| Control 1 EC | 0.099% | 10:05S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 10:05S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:06S | | |
| Control 2 EC | 0.098% | 10:06S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 10:06S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:07S | | |
| Control 3 EC | 0.098% | 10:08S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 10:08S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:09S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LLOYD

First Name: MATTHEW

MI: T

Signature: _____



Badge No.: 2760

Date: 12/07/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00919 Calib. Date: 08/07/2019 Calib. No.: 00032
Certification File No.: 00920 Cert. Date: 08/07/2019 Cert. No.: 00028
Linearity File No.: 00921 Lin. Date: 08/07/2019 Lin. No.: 00029
Solution File No.: 00938 Soln. Date: 11/12/2019 Soln. No.: 00228
Sequential File No.: 00938 File Date: 11/12/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0697

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 01:43S | | |
| Control 1 EC | 0.099% | 01:44S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 01:44S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:44S | | |
| Control 2 EC | 0.098% | 01:45S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 01:45S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:46S | | |
| Control 3 EC | 0.098% | 01:47S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 01:47S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:47S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____



Badge No.: 2732

Date: 11/12/2019

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARNK-0025 |
| Location: | WESTAMPTON POLICE | |
| Calibration File No.: | 00919 | Calib. Date: 08/07/2019 |
| Certification File No.: | 00920 | Calib. No.: 00032 |
| Linearity File No.: | 00921 | Cert. Date: 08/07/2019 |
| Solution File No.: | 00935 | Cert. No.: 00028 |
| Sequential File No.: | 00935 | Lin. Date: 08/07/2019 |
| | | Lin. No.: 00029 |
| | | Soln. Date: 10/14/2019 |
| | | Soln. No.: 00227 |
| | | File Date: 10/14/2019 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUJ S3-0117 |
| Solution Control Lot: | 19150 | Expires: 06/10/2021 |
| | | Bottle No.: 0953 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 13:22D | | |
| Control 1 EC | 0.100% | 13:22D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 13:22D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:23D | | |
| Control 2 EC | 0.099% | 13:24D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 13:24D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:24D | | |
| Control 3 EC | 0.099% | 13:25D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 13:25D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:26D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____



Badge No.: 2732

Date: 10/14/2019

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARNK-0025 |
| Location: | WESTAMPTON POLICE | |
| Calibration File No.: | 00919 | Calib. Date: 08/07/2019 |
| Certification File No.: | 00920 | Calib. No.: 00032 |
| Linearity File No.: | 00921 | Cert. Date: 08/07/2019 |
| Solution File No.: | 00934 | Cert. No.: 00028 |
| Sequential File No.: | 00934 | Lin. Date: 08/07/2019 |
| | | Lin. No.: 00029 |
| | | Soln. Date: 10/14/2019 |
| | | Soln. No.: 00226 |
| | | File Date: 10/14/2019 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUJ S3-0117 |
| Solution Control Lot: | 19150 | Expires: 06/10/2021 |
| | | Bottle No.: 1010 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 08:14D | | |
| Control 1 EC | ----% | 08:14D | 21.8°C | SIMULATOR TEMP. ERROR |
| Control 1 IR | ----% | 08:14D | 21.8°C | SIMULATOR TEMP. ERROR |
| Ambient Air Blank | ----% | --:-- | | |
| Control 2 EC | ----% | --:-- | N/A | |
| Control 2 IR | ----% | --:-- | N/A | |
| Ambient Air Blank | ----% | --:-- | | |
| Control 3 EC | ----% | --:-- | N/A | |
| Control 3 IR | ----% | --:-- | N/A | |
| Ambient Air Blank | ----% | --:-- | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____



Badge No.: 2732

Date: 10/14/2019

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARNK-0025 |
| Location: | WESTAMPTON POLICE | |
| Calibration File No.: | 00919 | Calib. Date: 08/07/2019 |
| Certification File No.: | 00920 | Calib. No.: 00032 |
| Linearity File No.: | 00921 | Cert. Date: 08/07/2019 |
| Solution File No.: | 00930 | Cert. No.: 00028 |
| Sequential File No.: | 00930 | Lin. Date: 08/07/2019 |
| | | Lin. No.: 00029 |
| | | Soln. Date: 09/15/2019 |
| | | Soln. No.: 00225 |
| | | File Date: 09/15/2019 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUJ S3-0117 |
| Solution Control Lot: | 19150 | Expires: 06/10/2021 |
| | | Bottle No.: 0949 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 21:11D | | |
| Control 1 EC | 0.099% | 21:12D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 21:12D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:12D | | |
| Control 2 EC | 0.098% | 21:13D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 21:13D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:14D | | |
| Control 3 EC | 0.098% | 21:14D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 21:14D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:15D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____

Badge No.: 2732

Date: 09/15/2019

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00919 Calib. Date: 08/07/2019 Calib. No.: 00032
Certification File No.: 00920 Cert. Date: 08/07/2019 Cert. No.: 00028
Linearity File No.: 00921 Lin. Date: 08/07/2019 Lin. No.: 00029
Solution File No.: 00927 Soln. Date: 09/02/2019 Soln. No.: 00224
Sequential File No.: 00927 File Date: 09/02/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0017

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 08:45D | | |
| Control 1 EC | 0.098% | 08:46D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 08:46D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:46D | | |
| Control 2 EC | 0.098% | 08:47D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 08:47D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:48D | | |
| Control 3 EC | 0.099% | 08:48D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 08:48D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:49D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI:

Signature: _____



Badge No.: 2732

Date: 09/02/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00890 Calib. Date: 02/28/2019 Calib. No.: 00031
Certification File No.: 00891 Cert. Date: 02/28/2019 Cert. No.: 00027
Linearity File No.: 00892 Lin. Date: 02/28/2019 Lin. No.: 00028
Solution File No.: 00915 Soln. Date: 07/17/2019 Soln. No.: 00222
Sequential File No.: 00915 File Date: 07/17/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0457

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 16:34D | | |
| Control 1 EC | 0.099% | 16:34D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 16:34D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:35D | | |
| Control 2 EC | 0.099% | 16:36D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 16:36D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:36D | | |
| Control 3 EC | 0.099% | 16:37D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 16:37D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:38D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROWBOTTOM

First Name: PTL

MI: J

Signature: _____

Badge No.: 2751

Date: 07/17/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00890 Calib. Date: 02/28/2019 Calib. No.: 00031
Certification File No.: 00891 Cert. Date: 02/28/2019 Cert. No.: 00027
Linearity File No.: 00892 Lin. Date: 02/28/2019 Lin. No.: 00028
Solution File No.: 00912 Soln. Date: 06/10/2019 Soln. No.: 00221
Sequential File No.: 00912 File Date: 06/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0850

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 20:29D | | |
| Control 1 EC | 0.099% | 20:30D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 20:30D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:31D | | |
| Control 2 EC | 0.098% | 20:31D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 20:31D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:32D | | |
| Control 3 EC | 0.098% | 20:33D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 20:33D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:33D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____



Badge No.: 2732

Date: 06/10/2019

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00890 Calib. Date: 02/28/2019 Calib. No.: 00031
Certification File No.: 00891 Cert. Date: 02/28/2019 Cert. No.: 00027
Linearity File No.: 00892 Lin. Date: 02/28/2019 Lin. No.: 00028
Solution File No.: 00901 Soln. Date: 05/10/2019 Soln. No.: 00220
Sequential File No.: 00901 File Date: 05/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0243

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 22:31D | | |
| Control 1 EC | 0.099% | 22:32D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 22:32D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:33D | | |
| Control 2 EC | 0.098% | 22:33D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 22:33D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:34D | | |
| Control 3 EC | 0.099% | 22:35D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 22:35D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:35D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____



Badge No.: 2732

Date: 05/10/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00890 Calib. Date: 02/28/2019 Calib. No.: 00031
Certification File No.: 00891 Cert. Date: 02/28/2019 Cert. No.: 00027
Linearity File No.: 00892 Lin. Date: 02/28/2019 Lin. No.: 00028
Solution File No.: 00899 Soln. Date: 03/30/2019 Soln. No.: 00219
Sequential File No.: 00899 File Date: 03/30/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0449

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 19:51S | | |
| Control 1 EC | 0.101% | 19:51S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 19:51S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:52S | | |
| Control 2 EC | 0.100% | 19:53S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 19:53S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:53S | | |
| Control 3 EC | 0.100% | 19:54S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 19:54S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:55S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROWBOTTOM

First Name: PTL.

MI: J

Signature: _____

Badge No.: 2751

Date: 03/30/2019

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00864 Calib. Date: 09/12/2018 Calib. No.: 00030
Certification File No.: 00865 Cert. Date: 09/12/2018 Cert. No.: 00026
Linearity File No.: 00866 Lin. Date: 09/12/2018 Lin. No.: 00027
Solution File No.: 00885 Soln. Date: 02/05/2019 Soln. No.: 00217
Sequential File No.: 00885 File Date: 02/05/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 1010

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 03:12S | | |
| Control 1 EC | 0.101% | 03:13S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 03:13S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:14S | | |
| Control 2 EC | 0.099% | 03:14S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 03:14S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:15S | | |
| Control 3 EC | 0.100% | 03:16S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 03:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:17S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____



Badge No.: 2732

Date: 02/05/2019

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARNK-0025 |
| Location: | WESTAMPTON POLICE | |
| Calibration File No.: | 00864 | Calib. Date: 09/12/2018 |
| Certification File No.: | 00865 | Calib. No.: 00030 |
| Linearity File No.: | 00866 | Cert. Date: 09/12/2018 |
| Solution File No.: | 00882 | Cert. No.: 00026 |
| Sequential File No.: | 00882 | Lin. Date: 09/12/2018 |
| | | Lin. No.: 00027 |
| | | Soln. Date: 01/07/2019 |
| | | Soln. No.: 00216 |
| | | File Date: 01/07/2019 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUJ S3-0116 |
| Solution Control Lot: | 18290 | Expires: 09/17/2020 |
| | | Bottle No.: 0898 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:45S | | |
| Control 1 EC | 0.101% | 08:46S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 08:46S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:46S | | |
| Control 2 EC | 0.099% | 08:47S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 08:47S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:48S | | |
| Control 3 EC | 0.099% | 08:48S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 08:48S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:49S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____



2732

Badge No.: 2732

Date: 01/07/2019