

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00919	Calib. Date:	08/07/2019
Certification File No.:	00920	Cert. Date:	08/07/2019
Linearity File No.:	00921	Lin. Date:	08/07/2019
Solution File No.:	00945	Soln. Date:	01/06/2020
Sequential File No.:	00945	File Date:	01/06/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0117
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	1006

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:53S		
Control 1 EC	0.097%	00:53S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:54S		
Control 2 EC	0.096%	00:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:56S		
Control 3 EC	0.097%	00:56S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: ROWBOTTOM

First Name: JOSH

MI: T

Signature: \_\_\_\_\_



Badge No.: 2751

Date: 01/06/2020