

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00291 **Calib. Date:** 07/30/2009 **Calib. No.:** 00009
Certification File No.: 00292 **Cert. Date:** 07/30/2009 **Cert. No.:** 00006
Linearity File No.: 00294 **Lin. Date:** 07/30/2009 **Lin. No.:** 00007
Solution File No.: 00313 **Soln. Date:** 12/04/2009 **Soln. No.:** 00072
Sequential File No.: 00313 **File Date:** 12/04/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0116
Control Solution %: 0.100% **Expires:** 12/01/2010
Solution Control Lot: 08L061 **Bottle No.:** 1145

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:31S		
Control 1 EC	0.101%	15:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	15:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32S		
Control 2 EC	0.101%	15:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34S		
Control 3 EC	0.101%	15:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 12/04/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00291 **Calib. Date:** 07/30/2009 **Calib. No.:** 00009
Certification File No.: 00292 **Cert. Date:** 07/30/2009 **Cert. No.:** 00006
Linearity File No.: 00294 **Lin. Date:** 07/30/2009 **Lin. No.:** 00007
Solution File No.: 00310 **Soln. Date:** 11/06/2009 **Soln. No.:** 00071
Sequential File No.: 00310 **File Date:** 11/06/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0116
Control Solution %: 0.100% **Expires:** 12/01/2010
Solution Control Lot: 08L061 **Bottle No.:** 1132

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:15S		
Control 1 EC	0.101%	16:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17S		
Control 2 EC	0.101%	16:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	16:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18S		
Control 3 EC	0.101%	16:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____

Sgt. [Signature] 2730

Badge No.: 2730

Date: 11/06/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00291 **Calib. Date:** 07/30/2009 **Calib. No.:** 00009
Certification File No.: 00292 **Cert. Date:** 07/30/2009 **Cert. No.:** 00006
Linearity File No.: 00294 **Lin. Date:** 07/30/2009 **Lin. No.:** 00007
Solution File No.: 00308 **Soln. Date:** 10/09/2009 **Soln. No.:** 00070
Sequential File No.: 00308 **File Date:** 10/09/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0116
Control Solution %: 0.100% **Expires:** 12/01/2010
Solution Control Lot: 08L061 **Bottle No.:** 1126

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:32D		
Control 1 EC	0.100%	21:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:34D		
Control 2 EC	0.099%	21:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:35D		
Control 3 EC	0.099%	21:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:36D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: 

Badge No.: 2730

Date: 10/09/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00291 Calib. Date: 07/30/2009 Calib. No.: 00009
Certification File No.: 00292 Cert. Date: 07/30/2009 Cert. No.: 00006
Linearity File No.: 00294 Lin. Date: 07/30/2009 Lin. No.: 00007
Solution File No.: 00305 Soln. Date: 09/11/2009 Soln. No.: 00069
Sequential File No.: 00305 File Date: 09/11/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 12/01/2010
Solution Control Lot: 08L061 Bottle No.: 1129

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:50D		
Control 1 EC	0.100%	17:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51D		
Control 2 EC	0.100%	17:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:53D		
Control 3 EC	0.100%	17:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____

Sgt. Stephen P. ... 2730

Badge No.: 2730

Date: 09/11/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00291 Calib. Date: 07/30/2009 Calib. No.: 00009
Certification File No.: 00292 Cert. Date: 07/30/2009 Cert. No.: 00006
Linearity File No.: 00294 Lin. Date: 07/30/2009 Lin. No.: 00007
Solution File No.: 00299 Soln. Date: 08/14/2009 Soln. No.: 00068
Sequential File No.: 00299 File Date: 08/14/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 12/01/2010
Solution Control Lot: 08L061 Bottle No.: 1128

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:54D		
Control 1 EC	0.099%	09:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55D		
Control 2 EC	0.099%	09:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 3 EC	0.100%	09:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOUGH

First Name: SEAN

MI: D

Signature: _____



Badge No.: 418

Date: 08/14/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00249 **Calib. Date:** 01/27/2009 **Calib. No.:** 00006
Certification File No.: 00250 **Cert. Date:** 01/27/2009 **Cert. No.:** 00005
Linearity File No.: 00251 **Lin. Date:** 01/27/2009 **Lin. No.:** 00005
Solution File No.: 00286 **Soln. Date:** 06/22/2009 **Soln. No.:** 00065
Sequential File No.: 00286 **File Date:** 06/22/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0117
Control Solution %: 0.100% **Expires:** 12/01/2010
Solution Control Lot: 08L061 **Bottle No.:** 1127

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:11D		
Control 1 EC	0.099%	18:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:12D		
Control 2 EC	0.096%	18:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:14D		
Control 3 EC	0.097%	18:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FERGUSON

First Name: BRIAN

MI: C

Signature: 

#2736

Badge No.: 2736

Date: 06/22/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00249	Calib. Date: 01/27/2009
Certification File No.:	00250	Calib. No.: 00006
Linearity File No.:	00251	Cert. Date: 01/27/2009
Solution File No.:	00284	Cert. No.: 00005
Sequential File No.:	00284	Lin. Date: 01/27/2009
		Lin. No.: 00005
		Soln. Date: 06/07/2009
		Soln. No.: 00064
		File Date: 06/07/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	08L061	Expires: 12/01/2010
		Bottle No.: 1134

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:53D		
Control 1 EC	0.096%	07:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54D		
Control 2 EC	0.095%	07:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	07:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:56D		
Control 3 EC	0.095%	07:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	07:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 06/07/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00249	Calib. Date:	01/27/2009
Certification File No.:	00250	Cert. No.:	00006
Linearity File No.:	00251	Cert. Date:	01/27/2009
Solution File No.:	00281	Lin. No.:	00005
Sequential File No.:	00281	Soln. Date:	05/26/2009
		Soln. No.:	00063
		File Date:	05/26/2009

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUJ S3-0117
Control Solution %:	0.100%			Expires:	10/03/2009
Solution Control Lot:	07J048			Bottle No.:	0598

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:27D		
Control 1 EC	0.096%	09:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28D		
Control 2 EC	0.095%	09:29D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		
Control 3 EC	0.096%	09:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	09:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LANDELL

First Name: PHILIP

MI: M

Signature: _____

Sgt Philip Landell

Badge No.: 421

Date: 05/26/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00249	Calib. Date:	01/27/2009
Certification File No.:	00250	Cert. No.:	00006
Linearity File No.:	00251	Cert. Date:	01/27/2009
Solution File No.:	00276	Lin. No.:	00005
Sequential File No.:	00276	Soln. Date:	05/11/2009
		Soln. No.:	00062
		File Date:	05/11/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0117
Solution Control Lot:	07J048	Expires:	10/03/2009
		Bottle No.:	0599

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:40D		
Control 1 EC	0.100%	14:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41D		
Control 2 EC	0.097%	14:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43D		
Control 3 EC	0.097%	14:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LANDELL

First Name: PHILIP

M1: M

Signature: _____



Badge No.: 421

Date: 05/11/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00249 **Calib. Date:** 01/27/2009 **Calib. No.:** 00006
Certification File No.: 00250 **Cert. Date:** 01/27/2009 **Cert. No.:** 00005
Linearity File No.: 00251 **Lin. Date:** 01/27/2009 **Lin. No.:** 00005
Solution File No.: 00269 **Soln. Date:** 04/16/2009 **Soln. No.:** 00061
Sequential File No.: 00269 **File Date:** 04/16/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0117
Control Solution %: 0.100% **Expires:** 10/03/2009
Solution Control Lot: 07J048 **Bottle No.:** 0586

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:30D		
Control 1 EC	0.100%	16:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:32D		
Control 2 EC	0.097%	16:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:33D		
Control 3 EC	0.097%	16:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

M1: P

Signature: 

Badge No.: 2730

Date: 04/16/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00249	Calib. Date:	01/27/2009
Certification File No.:	00250	Cert. Date:	01/27/2009
Linearity File No.:	00251	Lin. Date:	01/27/2009
Solution File No.:	00263	Soln. Date:	03/18/2009
Sequential File No.:	00263	File Date:	03/18/2009

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUJ S3-0117
Control Solution %:	0.100%			Expires:	10/03/2009
Solution Control Lot:	07J048			Bottle No.:	0585

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:39S		
Control 1 EC	0.101%	07:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		
Control 2 EC	0.098%	07:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		
Control 3 EC	0.099%	07:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 03/18/2009

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00249 **Calib. Date:** 01/27/2009 **Calib. No.:** 00006
Certification File No.: 00250 **Cert. Date:** 01/27/2009 **Cert. No.:** 00005
Linearity File No.: 00251 **Lin. Date:** 01/27/2009 **Lin. No.:** 00005
Solution File No.: 00257 **Soln. Date:** 02/21/2009 **Soln. No.:** 00059
Sequential File No.: 00257 **File Date:** 02/21/2009

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0117
Control Solution %: 0.100% **Expires:** 10/03/2009
Solution Control Lot: 07J048 **Bottle No.:** 0592

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:11S		
Control 1 EC	0.099%	05:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:12S		
Control 2 EC	0.097%	05:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:14S		
Control 3 EC	0.097%	05:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LANDELL

First Name: PHILIP

MI: M

Signature: _____



Badge No.: 421

Date: 02/21/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00208	Calib. Date: 07/30/2008
Certification File No.:	00209	Calib. No.: 00005
Linearity File No.:	00210	Cert. Date: 07/30/2008
Solution File No.:	00248	Cert. No.: 00004
Sequential File No.:	00248	Lin. Date: 07/30/2008
		Lin. No.: 00004
		Soln. Date: 01/18/2009
		Soln. No.: 00057
		File Date: 01/18/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 0591

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:19S		
Control 1 EC	0.101%	20:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:21S		
Control 2 EC	0.099%	20:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:22S		
Control 3 EC	0.098%	20:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: Sgt. Stephen ENT

Badge No.: 2730

Date: 01/18/2009

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00208	Calib. Date: 07/30/2008
Certification File No.:	00209	Calib. No.: 00005
Linearity File No.:	00210	Cert. Date: 07/30/2008
Solution File No.:	00238	Cert. No.: 00004
Sequential File No.:	00238	Lin. Date: 07/30/2008
		Lin. No.: 00004
		Soln. Date: 12/19/2008
		Soln. No.: 00056
		File Date: 12/19/2008

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUJ S3-0117
Control Solution %:	0.100%		Expires: 10/03/2009
Solution Control Lot:	07J048		Bottle No.: 0597

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:06S		
Control 1 EC	0.100%	10:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07S		
Control 2 EC	0.098%	10:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09S		
Control 3 EC	0.097%	10:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 12/19/2008

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00208 Calib. Date: 07/30/2008 Calib. No.: 00005
Certification File No.: 00209 Cert. Date: 07/30/2008 Cert. No.: 00004
Linearity File No.: 00210 Lin. Date: 07/30/2008 Lin. No.: 00004
Solution File No.: 00234 Soln. Date: 11/21/2008 Soln. No.: 00055
Sequential File No.: 00234 File Date: 11/21/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 10/03/2009
Solution Control Lot: 07J048 Bottle No.: 0590

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:36S		
Control 1 EC	0.099%	19:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:37S		
Control 2 EC	0.098%	19:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:38S		
Control 3 EC	0.097%	19:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____

Sgt. AEB 2730

Badge No.: 2730

Date: 11/21/2008

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00208	Calib. Date: 07/30/2008
Certification File No.:	00209	Calib. No.: 00005
Linearity File No.:	00210	Cert. Date: 07/30/2008
Solution File No.:	00227	Cert. No.: 00004
Sequential File No.:	00227	Lin. Date: 07/30/2008
		Lin. No.: 00004
		Soln. Date: 10/21/2008
		Soln. No.: 00054
		File Date: 10/21/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 0600

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:36D		
Control 1 EC	0.100%	18:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:37D		
Control 2 EC	0.099%	18:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:39D		
Control 3 EC	0.097%	18:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

M1: P

Signature: _____



Badge No.: 2730

Date: 10/21/2008

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00208	Calib. Date: 07/30/2008
Certification File No.:	00209	Calib. No.: 00005
Linearity File No.:	00210	Cert. Date: 07/30/2008
Solution File No.:	00218	Cert. No.: 00004
Sequential File No.:	00218	Lin. Date: 07/30/2008
		Lin. No.: 00004
		Soln. Date: 09/24/2008
		Soln. No.: 00053
		File Date: 09/24/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0117
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 0588

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:15D		
Control 1 EC	0.101%	05:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:16D		
Control 2 EC	0.100%	05:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:17D		
Control 3 EC	0.100%	05:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____



Badge No.: 2730

Date: 09/24/2008

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00208 Calib. Date: 07/30/2008 Calib. No.: 00005
Certification File No.: 00209 Cert. Date: 07/30/2008 Cert. No.: 00004
Linearity File No.: 00210 Lin. Date: 07/30/2008 Lin. No.: 00004
Solution File No.: 00215 Soln. Date: 08/27/2008 Soln. No.: 00052
Sequential File No.: 00215 File Date: 08/27/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0117
Control Solution %: 0.100% Expires: 10/03/2009
Solution Control Lot: 07J048 Bottle No.: 0579

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:58D		
Control 1 EC	0.098%	10:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 2 EC	0.097%	11:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00D		
Control 3 EC	0.096%	11:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

M1: P

Signature: _____



Badge No.: 2730

Date: 08/27/2008